

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd
72, Dongtancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do, Republic of Korea
Phone : 82-31-379-5114, Fax : 82-31-379-5115 e-mail : sictadmin@sict.co.kr

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018(1/208)

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102. Linear dimension			10236	Coating thickness testers	Y	10413	Straight rules	N
10201	Balls	N	10237	Torque arms	N	10415	Test bars	N
10203	Electrical /Mechanical comparators	Y	103. Angle			105. Complex geometry		
10206	Dial/cylinder gauge testers	N	10302	Angle gauge blocks	N	10501	Base gauges for electric bulb	N
10207	Doctor blades	N	10303	Autocollimators	N	10502	Baench centers	Y
10208	Distance meters;electrooptic/laser/ultrasonic	N	10304	Bevel protractors	Y	10503	Contact coordinate measuring machines	Y
10209	End bars	N	10306	Clinometers	N	10504	Non-contact coordinate measuring machines	Y
10210	Extensometers, lineardisplacement transducers	Y	10308	Fine angle generators, level comparators	N	10505	Gauge block accessories	N
10211	Filler gauges	Y	10310	Indexing tables	N	10511	Measuring microscopes, Profile projectors	Y
10212	Film applicators	N	10311	Plate/Square/Electric levels	N	10512	Microscopes, micro measuring	Y
10213	Gap gauges	N	10314	Penta-prisms	N	10514	Taper plug gauges	N
10214	Gauge blocks, by comparison	N	10315	Polygons	N	10515	Taper ring gauges	N
10216	Height gauges/measuring machines	Y	10316	Rotary tables	Y	10517	Stylus type roughness testers	Y
10219	Linear scales	N	10317	Sine bars, Plates, Tables, Centers	N	10518	Socket gauges for electric bulb	N
10220	Measuring machines, standard	Y	10318	Squareness testers, Right angle testers	Y	10519	Roughness standard /comparison specimens	N
10221	Micro scales/Standard scales	N	10319	Cylindrical squares	N	10525	Thread plug gauges	N
10223	Electronic micrometers	Y	10320	Precision squares	N	10526	Taper thread plug gauges	N
10224	Height micrometers, Riser blocks	N	10322	Angular displacement transducers	N	10527	Thread ring gauges	N
10225	Laser scan micrometers	Y	104. Form			10528	Taper thread ring gauges	N
10227	Standard taper rules, Peripheral gauges	N	10401	Form testers	Y	10529	V-blocks, Box blocks	N
10228	Cylindrical plug/pin gauges, Thread measuring wire gauges	Y	10404	Optical flats	N	106. Various dimensional		
10229	Radius gauges	N	10405	Optical parallels	N	10601	Inside/Outside/Gear tooth calipers, Caliper gauges	Y
10230	Cylindrical ring gauges	N	10406	Parallel blocks	Y	10603	Cylinder/bore gauges	Y
10232	Step gauges	N	10407	Precision surface plates	Y	10604	Depth gauges, Depth micrometers	Y
10233	Thickness gauges, taper	N	10408	Profile gauges	N	10605	Dial/digital gauges	Y
10234	Ultrasonic thickness gauges	Y	10409	Roundness measurement instruments	Y	10608	Grind gauges	N
10235	Ultrasonic/coating thickness specimens	N	10411	Roundness standard/ Roundness magnification standard specimens	N	10609	Micro indicators, Test indicators	Y
			10412	Straight edges	Y			

Accreditation No : KC01-018(2/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site		
10610	Micrometer heads	Y	20503	Ionization gauges	N	20916	Gas flowmeters;turbine	N		
10611	3-points, Micrometers	Y				20917	Liquid flowmeters;turbine	N		
10612	Inside micrometers	Y	20504	Thermal conductivity gauge; Pirani, thermocouple, convection etc.	N	20918	Gas flowmeters;ultrasonic	N		
10613	Outside micrometers	Y				20919	Liquid flowmeters;ultrasonic	N		
10615	Particle counters	N				20920	Gas flowmeters;variable area	N		
10617	Standard sieves	N	20505	Standard leaks, Helium leak dectors	Y	20921	Liquid flowmeters;variable area	N		
10620	Welding gauges	N				20922	Gas flowmeters;vortex	N		
201. Mass			206. Volume			20923	Liquid flowmeters;vortex	N		
20102	Auto-hopper scale balances	Y	20601	Volumetric glasswares	N	20925	Anemometers; vane, etc	N		
20103	Auto-packer scale balances	Y	20602	Pycnometers	N	20999	Others ;	N		
20104	Axle weigher balances	N	20604	Standard volume vessels	Y		Anemometers; ultrasonic waves			
20106	Dial platform scale balances	Y	20605	Concrete air content meters	N	210. Hardness				
20107	Dial swing scale balances	Y	20606	Piston type volume meters	N	21001	Brinell hardness testers	Y		
20109	Electric balances	Y	207. Density			21002	Rockwell hardness testers	Y		
20111	Manual swing scale balances	Y	20702	Liquid density meters	N	21003	Shore hardness testers	Y		
20112	Platform scale balances	Y	20704	Salinity meters	N	21004	Vickers hardness testers	Y		
20113	Spring scale balances	Y	20705	Sucrose meters	N	21005	Durometer hardness testers	N		
20116	Weights	Y	20706	Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil salinity, LPG, etc.	N	21006	Leeb hardness testers	Y		
202. Force						301. Time/frequency				
20202	Force measuring devices	N				30102	Frequency standards	N		
20203	Tension/compression testing machines	Y				30103	General frequency sources	Y		
20204	push-pull gauge	Y	20707	Chloride meters	N	30104	Frequency meters/counters	Y		
203. Torque			20799 Others ; Solid density			N	30105	Time interval sources	Y	
20302 Torque measuring devices			Y	20801 Kinematic viscometers; capillary, etc.			N	30106	Time interval meters /Stop watches/Timers	Y
20303 Torque wrenches/drivers			Y	20802 Dynamic viscometers; rotaional, etc			Y	302. Velocity & revolution		
20401	Altimeters	Y	20901 Anemometers; hot-wire			N	30201	Standard RPM generators	Y	
20402	Manometers	Y	20902 Anemometers; pitot tube, etc.			N	30202	Contact type tachometers	Y	
20403	Pneumatic pressure ballances	N	20908	Gas flowmeters;differential pressure	N	30203	Photo tachometers /stroboscopes	Y		
20404	Hydraulic pressure ballances	N				30204	Speed meters	Y		
20405	Air data test systems	Y	20909	Liquid flowmeters;differentia pressure	N	30205	Wow-flutter generators	N		
20406	Absolutepressuregauges	Y				30206	Wow-flutter meters	Y		
20407	Blood pressure gauges	Y	401. DC Voltage & current							
20408	Compound pressure gauges	Y	20910	Liquid flowmeters; electromagnetic	N	40101	DC ammeters	Y		
20409	Differential pressure gauges	Y				40102	Transconductance amplifiers	Y		
20411	Gauge pressure gauges	Y	20911	Gas flowmeters;thermal mass, etc.	N	40103	DC voltage/current calibrators	Y		
20412	Pressure transducers/ transmitters	Y				40104	Electricaltemperature calibrators	Y		
20413	Dial type vacuum gauges	Y	20912	Liquid flowmeters; Coriolis, etc.	N	40105	DC current shunts	Y		
205. Vacuum						40106	Galvanometers /null detectors	Y		
20501	Capacitance diaphragm gauges	N	20914	Gas flowmeters;open channel, etc.	N	40107	Potentiometers	Y		
20502	Spinning rotor gauges	N								
			20915	Liquid flowmeters; positive displacement	N					

Accreditation No : KC01-018(3/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
40108	DC power supplies	Y	40316	Current transformer	N	405. Low frequency electric & magnetic fields		
40110	DC voltage dividers	N	40318	AC voltmeters	Y			
40111	DC voltage standards	N	40319	Watt hour meters	N	40503	Flux meters	Y
40112	DC voltmeters	Y	40321	Ratio transformers	N	40504	Flux sources	N
40113	Static/Ionicvoltmeters	N	404. Other DC & LF Measurements			40508	Magnetometers	Y
402. Resistance, Capacitance and Inductance			40401	LF amplifiers	Y	40510	Reference/standard magnets	N
			40402	DC/LF attenuators	Y	406. Radio frequency measurements		
40201	Capacitance bridges /indicators	Y	40403	Multimeter calibrators	N	40601	RF amplifiers	Y
40202	Decade capacitors	Y	40404	Oscilloscope calibrators	N	40602	Coaxial attenuators	Y
40204	Standard capacitors	Y	40405	CD/DVD meters/analyzers	Y	40605	Burst pulse generators	Y
40205	Earth testers	Y	40406	Video signal generators	Y	40606	Attenuator calibrators	N
40206	Inductance bridges /indicators	Y	40407	Audio distortion analyzers /meters	Y	40607	RF power meter calibrators	Y
40208	Inductors	Y	40408	LF filters	Y	40608	EMC transducers ; current probes, absorbing absorbing clamps, etc.	Y
40210	Insulation testers	Y	40409	LF/Audiosignalanalyzers	Y	40610	Coaxial directional couplers /splitters	Y
40211	Q-meters	Y	40410	Line frequency meters	Y			
40213	Resistance bridges & similar instruments	Y	40411	Function generators	Y	40613	Electrostatic discharge generators	Y
40214	Resistance meters	Y	40412	Genescopes	Y	40614	EMC receivers	Y
40215	Resistors	Y	40413	AC/DC high voltages volt meters	Y			
40217	Impedance bridges/LCR meters	Y	40415	Jitter meters	Y	40615	RF filters	Y
403. AC voltage, current & power			40416	Leakage current testers	Y	40616	RF impedance meters	N
40301	AC ammeters	Y	40417	Electronic AC/DC loads	Y	40617	RF impulse generators	Y
40302	Clamp ammeters/voltmeters	Y	40418	Modulation meters	Y	40618	Line impedance stabilization networks ; LISN, CDN, ISN, etc.	Y
40303	AC voltage/current calibrators	Y	40419	Analogue/Digital multimeters	Y			
40304	Wattmeter calibrators	N	40420	Noise meters	Y	40619	Coaxialstandardmismatches	Y
40305	AC current shunts	Y	40421	Oscilloscopes	Y	40621	Mobile communication test sets	Y
40306	Phase angle generators, synchro resolve generators	N	40422	LF phase meters	Y			
40307	Voltage/current phase angle meters/synchro resolve meters	N	40424	Volt/Current recorders	Y	40622	Modulation meters	Y
			40425	Relay test sets	Y	40623	Network analyzers	Y
40308	Potential transformer test sets	Y	40426	LF signal generators	Y	40624	Noise figure meters	Y
			40427	LF spectrum analyzers	Y	40625	Noise generators	N
40309	Potential transformer	N	40428	Spot generators	Y	40626	Noise impulse simulators	Y
40310	Power factor meters	Y	40429	Sweep generators	Y	40627	RF phase noise meters	N
40311	AC power meters	Y	40430	Signal transducers	Y	40628	Coaxial noise sources	N
40312	AC power supplies	Y	40432	Transistor curve tracers	Y	40635	RF power meters	Y
40313	Puncture/safety testers	Y	40434	AC/DC high voltage generators	Y	40636	Diode power sensors	Y
40314	Power recorders	Y	40435	AC/DC high voltage probes	Y	40637	Thermocouple power sensors	Y
40315	Current transformer test sets	Y	40436	Logic analyzers	Y	40638	Pulse generators	Y
			40437	Telephone testers	Y	40639	Radar test sets	Y
			40438	Video signal analyzers	Y	40640	RF signal generators	Y
			40438	Video signal analyzers	Y	40641	RF spectrum analyzers	Y
						40643	Surge generators	Y

Accreditation No : KC01-018(4/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site						
40644	SWR meters	N	50303	Psychrometers; assmann ventilated, PRT type, etc.	N	70213	Display color analyzers; luminance, chromaticity, white balance, etc.	Y						
40645	RF terminations	Y												
40646	Coaxial thermistor mounts	Y												
40650	RF voltmeters	Y	50304	Temperature humidity recorders ; Hygrothermograph, etc	N	70214	Luminous intensity standard lamps	N						
40651	Vector voltmeters	Y												
40652	Field strength meters	Y												
40653	AM/FM test sources	Y	50305	Transducers; dew-point /relative humidity	N	70215	Spectral irradiance standard lamps	N						
40654	Dip simulators	Y												
407. Field strength & antennas									50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y	70216	Total spectral radiand flux standard lampa	N
40704	Loop antennas	N												
40705	Monopole antennas	N												
501. Contact thermometry			70217	Luminance standard sources	N									
50101	Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y				70218	Spectral radiance standard lamps	N						
									70219	UV irradiance meters	N			
												70220	Spectral irradiance meters	Y
504. Moisture														
50401	Cereal moisture meters	Y												
50102	Temperature indicators /recorders/controllers, temperature calibrators	Y	50402	Wood moisture meters	N	70222	Spectral radiance meters	Y						
50403			Paper moisture meters	N										
601. Sound in air			70301	Colorimeters; material color	Y									
50103	Glass thermometers; liquid-in-glass, Beckmann	N	60102	Sound calibrator	N	70304	Color standard tiles	N						
			60104	Microphones	N	70306	Gloss meters	Y						
50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y	60106	Sound level meters	Y	70307	Gloss standard plates	Y						
			603. Vibration			70308	Haze meters	Y						
50105	Thermal expansion thermometers ; bimetal, gas or liquid type	Y	60301	Vibration calibrators	N	70312	Lens testers	Y						
			60302	Vibration transducers	N	70315	Optical densitometers	Y						
			60303	Vibration measuring instruments	N	70319	Reflectance meters	Y						
70321	Refractometers	Y												
50106	Thermomecouples: noble metal, base metal, pure metal, special type, etc.	Y	701. Photometry			70323	Transmittance meters	Y						
			70101	Illuminance meters	N	70325	Spectrophotometers including FT-IR spectrophotometers	Y						
50107	Temperature transducers	Y	70102	Luminance meters	N									
50108	Primary fixed-point cells and apparatus	N	70103	Total luminous flux meters	Y				70326	Wavelength reference material; absorption cell, bandpass filter, etc.	N			
			70104	Luminous intensity meters	Y									
502. non contact thermometry			702. Properties of detectos & sources											
50203	Optical pyrometers	N	70202	Color temperature meters	Y									
50204	Standard radiation	N	70203	Color temperature standard lamps	N									
50205	Thermal image apparatus	N												
50206	Blackbody furnaces	Y	70204	Colorimeters; source color	Y									
50207	Others ; ear thermometers, etc.	N	70207	Laser power meters	N									
			503. Humidity			70208	Standard LED light sources	N						
50301	Dew-point hygrometers; chilled mirror, alumina thinfilm, etc.	N	70209	Total luminous flux standard lamps	N									
			70210	Optical detectors	N									
50302	Relative humidity hygrometers polimer thinfilm, hair, etc.	Y	70211	Pyranometers and pyrheliometers	N									

Accreditation No : KC01-018(5/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
704. Fiber optics			901. Chemical analysis					
70402	Broadband light sources	Y	90101	Breath alcohol analyzer	N			
70410	Optical attenuators	Y	90102	Environmental air monitoring instruments	Y			
70412	Fiber-optic power meters	Y						
70413	Optical loss testers	Y	90103	Gas analyzers	Y			
70415	Optical multimeters	Y	90104	Exhaust gas test inastruments	Y			
70416	Optical network analyzers	Y	90199	Others ; pH meter, Electrical conductivity meter	Y			
70417	Optical spectrum analyzers	Y						
70418	Optical time domain reflectometers, OTDR	Y						
70423	Return loss meters	Y						
70429	Frequency stabilized lasers and LDs	Y						
70430	ASE light sources	Y						
70433	Optical power stabilized lasers and LDs	Y						

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95%, which usually requires the use of a coverage factor of $k=2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls	10201	(0.3 ~ 100) mm	$\sqrt{0.38^2 + (0.004\ 6 \times l_0)^2}$ μm	Mesuring Machine, Standard/ SICT-CP-10201
Electrical/Mechanical comparators	10203	(0 ~ 5) mm	0.14 μm	Gauge Block/ SICT-CP-10203
Dial/cylinder gauge testers	10206	(0 ~ 25) mm	$\sqrt{0.12^2 + (0.002\ 8 \times l_0)^2}$ μm	Laser Measurement Machine/ SICT-CP-10206
Doctor Blades	10207	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10207
Distance meters; electrooptic/laser/ultrasonic	10208	(0 ~ 40) m	$\sqrt{0.28^2 + (0.005 \times l_0)^2}$ mm	Laser interferometer/ SICT-CP-10208
End bars	10209	(25 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{0.74^2 + (0.003 \times l_0)^2}$ μm $\sqrt{0.80^2 + (0.003 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10209
Extensometers, lineardisplacementtransducers	10210	(0 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{1.9^2 + (0.004\ 2 \times l_0)^2}$ μm $\sqrt{7.8^2 + (0.058 \times l_0)^2}$ μm	Gauge Block, Laser Measurement Machine/ SICT-CP-10210
Filler gauges	10211	(0 ~ 10) mm	1.2 μm	Mesuring Machine, Standard/ SICT-CP-10211
Film Applicators	10212	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10212
Gap gauges	10213	(1 ~ 500) mm	$\sqrt{0.72^2 + (0.004\ 8 \times l_0)^2}$ μm	Gauge Block, contact coordinate measuring machines/ SICT-CP-10213
Gauge blocks, by comparison	10214	(0.5 ~ 100) mm	$\sqrt{80^2 + (1.3 \times l_0)^2}$ nm	Gauge Block Comparator/ SICT-CP-10214
Height gauges/measuring machines	10216	(0 ~ 1 500) mm	$\sqrt{0.68^2 + (0.003\ 5 \times l_0)^2}$ μm	Gauge Block, Step gauge/ SICT-CP-10216
	10219	(0 ~ 40) m	$\sqrt{0.03^2 + (0.002\ 7 \times l_0)^2}$ mm	Laser interferometer/ SICT-CP-10219
Measuring machines, standard	10220	(0 ~ 500) mm	$\sqrt{0.38^2 + (0.002 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10220
Micro scales/Standard scales Micro scales Standard scales	10221	(0 ~ 1) mm (0 ~ 600) mm	0.4 μm $\sqrt{0.51^2 + (0.002\ 1 \times l_0)^2}$ μm	Non-contact coordinate measuring machines, standard scale/ SICT-CP-10221
Electronic micrometers	10223	(0 ~ 5) mm	0.14 μm	Gage Block/ SICT-CP-10223
Height micrometers,Riser blocks Block Head	10224	(0 ~ 600) mm (0 ~ 25) mm	$\sqrt{1.1^2 + (0.001\ 9 \times l_0)^2}$ μm 1.2 μm	Gauge Block Electronic Micrometer/ SICT-CP-10224
Laser scan micrometers	10225	(0.5 ~ 85) mm	$\sqrt{0.46^2 + (0.003 \times l_0)^2}$ μm	Cylindrical plug/pin gauge/ SICT-CP-10225
Standard tape rules,Peripheral gauges	10227	(0 ~ 40) m (40 ~ 80) m (80 ~ 100) m	$\sqrt{0.22^2 + (0.004\ 6 \times l_0)^2}$ mm $\sqrt{0.25^2 + (0.004\ 6 \times l_0)^2}$ mm $\sqrt{0.34^2 + (0.004\ 6 \times l_0)^2}$ mm	Laser Measurement Machine/ SICT-CP-10227
Cylindrical plug/pin gauges, Thread measuring wire gauges Cylindrical plug/pin gauges Thread measuring wire gauges	10228	(0.01 ~ 200) mm (0.1 ~ 10) mm	$\sqrt{0.42^2 + (0.003 \times l_0)^2}$ μm 0.41 μm	Mesuring Machine, Standard/ SICT-CP-10228
Radius gauges	10229	(0.1 ~ 100) mm	2.8 μm	non-contact coordinate measuring machines, standard scale/ SICT-CP-10229
Cylindrical ring gauges	10230	(1.0 ~ 200) mm	$\sqrt{0.69^2 + (0.003\ 4 \times l_0)^2}$ μm	Mesuring Machine, Standard/ SICT-CP-10230

Note 1. l_0 unit : mm (10208,10227 unit : m)

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Step gauges	10232	(0 ~ 1 500) mm	$\sqrt{0.76^2 + (0.0025 \times l_0)^2} \text{ } \mu\text{m}$	Gauge Block, Step gauge/ SICT-CP-10232
Taper thickness gauges	10233	(0 ~ 60) mm	4.3 μm	Profile Projector/ SICT-CP-10223
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Ultrasonic/coating thickness specimens coating thickness specimens Ultrasonic thickness specimens	10235	(0.01 ~ 2) mm (0.5 ~ 500) mm	0.7 μm $\sqrt{0.64^2 + (0.006 \times l_0)^2} \text{ } \mu\text{m}$	Gauge Block, Measuring Machine, Standard/ SICT-CP-10235
Coating thickness testers	10236	(0 ~ 2) mm	1.1 μm	Thickness specimens/ SICT-CP-10236
Torque arms Torque arms Wires	10237	(1 ~ 1 500) mm (0 ~ 5) mm	$\sqrt{0.60^2 + (0.0061 \times l_0)^2} \text{ } \mu\text{m}$ 1.2 μm	Gauge Block, contact coordinate measuring machines/ SICT-CP-10237

Note 1. l_0 unit : mm (10208,10227 unit : m)

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angle gauge blocks Angle	10302	(0 ~ 360)°	0.6"	Indexing tables/ SICT-CP-10302
Autocollimators Angle	10303	± 500"	0.6"	Fine angle generators/ SICT-CP-10303
Bevel protractors Angle Accuracy Accessory Angle	10304	(0 ~ 360)° (0 ~ 90)°	0.9' 0.7'	Angle Gauge Block, Coordinate Measur Machine/ SICT-CP-10304
Clinometers Angle	10306	(0 ~ 360)°	3.3"	Rotary tables/ SICT-CP-10306
Fine angle generators, level comparators Angle	10308	± 1 000"	0.4"	Autocollimators/ SICT-CP-10308
Indexing tables Angle	10310	(0 ~ 360)°	0.5"	Polygons/ SICT-CP-10310
Plate/Square/Electric levels Angle Inclino meter Squareness	10311	(0 ~ 516)" (516 ~ 1 000)" (0 ~ 90)° (0 ~ 400) mm	0.4" 1.2" 0.05' 1.8 μm	Fine angle generators, Rotary tables/ SICT-CP-10311
Penta-prisms Angle	10314	90°	0.6"	Autocollimators/ SICT-CP-10314
Polygons Angle	10315	(0 ~ 360)°	0.4"	Indexing tables/ SICT-CP-10315
Rotary tables Angle	10316	(0 ~ 360)°	0.5"	Polygons/ SICT-CP-10316
Sinebars, Plates, Tables, Centers (Sinebars) distance, between two roller center parallelism, between two roller parallelism, between flat-two roller (Plates) Center length Flatness Parallelism	10317	(100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm	$\sqrt{0.36^2 + (0.002 \times L_0)^2}$ μm 0.5 μm 0.6 μm $\sqrt{0.12^2 + (0.028 \times L_0)^2}$ μm 1.0 μm 1.2 μm	Mesuring Machine, Standard/ SICT-CP-10317
Squareness testers, Right angle testers	10318	(0 ~ 600) mm	2.0 μm	Cylindrical Square, Precision Square/ SICT-CP-10318

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cylindrical squares	10319	(0 ~ 300) mm (300 ~ 600) mm	1.6 μ m 1.9 μ m	Cylindrical Square/ SICT-CP-10319
Precision squares Squareness Parallelism	10320	(0 ~ 600) mm (0 ~ 600) mm	2.9 μ m 2.0 μ m	contact coordinate measuring machines/ SICT-CP-10320
Angular displacement transducers Angle	10322	(0 ~ 360)°	2.9"	Rotary tables/ SICT-CP-10322

Note 1. l_0 unit : mm

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers Height length Width Angle	10401	(0 ~ 50) mm (0 ~ 50) mm 15° ~ 45°	1.0 μm 1.4 μm 2'	Form Standard Specimens Gage Block, Angle Gage Block/ SICT-CP-10401
Optical flats Flatness	10404	∅ (0 ~ 60) mm ∅ (60 ~ 100) mm	0.05 μm 0.08 μm	Optical Flat/ SICT-CP-10404
Optical parallels Flatness Parallelism	10405	∅ (0 ~ 30) mm ∅ (0 ~ 30) mm	0.05 μm 0.07 μm	Optical Flat, Gauge block comparator/ SICT-CP-10405
Parallel blocks Parallelism Flatness Length Difference	10406	(0 ~ 1 000) mm (0 ~ 1 000) mm (0 ~ 1 000) mm	1.5 μm 1.5 μm 2.2 μm	Electronic Micrometer/ SICT-CP-10406
Precision surface plates Flatness	10407	(2 000 × 2 000) mm (5 000 × 5 000) mm	2.0 μm 4.8 μm	Electronic Level/ SICT-CP-10407
Profile gauges	10408	(0 ~ 5) mm	0.7 μm	Gage Block/ SICT-CP-10408
Roundness measurement instruments Detector accuracy Rotational accuracy of spindle Rotational accuracy of axis	10409	(0 ~ 20) μm 360° 360°	0.51 μm 0.03 μm 0.04 μm	Roundness Standard Ball/ SICT-CP-10409
Roundness standard/Roundness magnification standard specimens Standard specimens Standard ball	10411	(0 ~ 300) μm 360°	0.52 μm 0.08 μm	Roundness Tester/ SICT-CP-10411
Straight edges Straightness Parallelism	10412	(0 ~ 2 000) mm (0 ~ 2 000) mm	1.8 μm 1.8 μm	Electronic levels/ SICT-CP-10412
Straight rules Length	10413	(0 ~ 2 000) mm	0.10 mm	LASER INTERFEROMETER/ SICT-CP-10413
Test bars Roundness Cylindricity Run-out	10415	(0 ~ 400) mm (0 ~ 400) mm (0 ~ 400) mm	0.6 μm 0.6 μm 1.1 μm	Roundness Tester, Electronic Micrometer/ SICT-CP-10415

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Base gauges for electric bulb Inside diameter of pass/stop and screw Pitch	10501	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.47^2 + (0.0028 \times l_0)^2} \mu\text{m}$ 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10501
Bench centers Difference of both center Flatness of both bed Center height difference	10502	(0 ~ 200) mm (200 ~ 500) mm (0 ~ 500) mm (0 ~ 200) mm (200 ~ 500) mm	1.8 μm 3.4 μm 1.5 μm 1.8 μm 3.4 μm	Test Bar/ SICT-CP-10502
Contact coordinate measuring machines	10503	(0 ~ 1 500) mm (0 ~ 600) mm (0 ~ 600) mm	$\sqrt{0.56^2 + (0.0044 \times l_0)^2} \mu\text{m}$ 3.2 μm 1.2 μm	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 1 000) mm	$\sqrt{0.43^2 + (0.0034 \times l_0)^2} \mu\text{m}$	Standard Scale/ SICT-CP-10504
Gauge block accessories Round the ministry of Justice A he ministry of Justice,Base block Center point Plane figure Parallelism(triangle edge) Parallelism(equilibrium tide)	10505	(0 ~ 50) mm (0 ~ 50) mm (0 ~ 20) mm (0 ~ 50) mm (0 ~ 300) mm (0 ~ 150) mm	$\sqrt{0.32^2 + (0.0044 \times l_0)^2} \mu\text{m}$ $\sqrt{0.26^2 + (0.0044 \times l_0)^2} \mu\text{m}$ 1.1 μm 0.04 μm 0.4 μm 0.4 μm	Gauge Block/ SICT-CP-10505
Measuring microscopes,Profile projectors Length Magnification Angle	10511	(0 ~ 500) mm (5 ~ 100) mm (0 ~ 360) °	$\sqrt{0.43^2 + (0.0034 \times l_0)^2} \mu\text{m}$ 0.04 % 0.9'	Standard Scale/ SICT-CP-10511
Microscopes, micro measuring	10512	(0 ~ 1) mm (1 ~ 50) mm	0.7 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Taper plug gauges Small end diameter Big end diameter Plane angle Gage height	10514	(2 ~ 200) mm (2 ~ 200) mm (0 ~ 90) ° (2 ~ 200) mm	$\sqrt{1.3^2 + (0.0041 \times l_0)^2} \mu\text{m}$ $\sqrt{1.4^2 + (0.0041 \times l_0)^2} \mu\text{m}$ 5.9" $\sqrt{1.2^2 + (0.0044 \times l_0)^2} \mu\text{m}$	Mesuring Machine, Standard/ SICT-CP-10514
Taper ring gauges Small end diameter Big end diameter Plane angle	10515	(5 ~ 200) mm (5 ~ 200) mm (0 ~ 90) °	2.5 μm 2.5 μm 0.006°	contact coordinate measuring machines/ SICT-CP-10515
Stylus type roughness testers Rounghness parameter(Ra) Rounghness parameter(Rz) Mean width(RSm) H,D	10517	(0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm (0 ~ 300) μm (0 ~ 6) μm (6 ~ 20) μm	9 nm 24 nm 77 nm 0.27 μm 1.3 μm 63 nm 97 nm	Roughness Specimen/ SICT-CP-10517
Socket gauges for electric bulb Outside diameter of pass/stop and screw Pitch	10518	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.44^2 + (0.0028 \times l_0)^2} \mu\text{m}$ 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10518

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Roughness standard/comparison specimens	10519			Roughness Tester/ SICT-CP-10519
Depth(H)		(0 ~ 6) μm	$\sqrt{(9.6 \times R)^2 + 12^2}$ nm	
		(6 ~ 20) μm	$\sqrt{(8.94 \times R)^2 + 15^2}$ nm	
Mean width(RSm)		(0 ~ 300) μm	$\sqrt{(0.01 \times R)^2 + 0.51^2}$ μm	
Roughness parameter(Ra)		(0 ~ 2) μm	$\sqrt{(10 \times R)^2 + 2.6^2}$ nm	
		(2 ~ 10) μm	$\sqrt{(9.2 \times R)^2 + 7.4^2}$ nm	
Roughness parameter(Rz)		(0 ~ 7) μm	$\sqrt{(29.2 \times R)^2 + 7.4^2}$ nm	
		(7 ~ 30) μm	$\sqrt{(0.025 \times R)^2 + 0.096^2}$ μm	
Thread plug gauges	10525			Measuring Machine, Standard/ SICT-CP-10525
Outside diameter		(1 ~ 205) mm	1.7 μm	
Effective diameter		(1 ~ 210) mm	1.1 μm	
Pitch		(0.3 ~ 10) mm	1.2 μm	
Half angle		(0.5 ~ 45) °	2'	
Taper thread plug gauges	10526			Measuring Machine, Standard/ SICT-CP-10526
Half angle		(0 ~ 45) °	2'	
Pitch		(0.3 ~ 6) mm	1.3 μm	
Gage length		(2 ~ 50) mm	2.6 μm	
Notch length		(2 ~ 50) mm	3.6 μm	
Small outside diameter		(2 ~ 200) mm	2.3 μm	
Big outside diameter		(2 ~ 200) mm	4.8 μm	
Small effective diameter		(2 ~ 200) mm	2.9 μm	
Big effective diameter		(2 ~ 200) mm	5.1 μm	
Thread ring gauges	10527			Measuring Machine, Standard/ SICT-CP-10527
Minor diameter		(3 ~ 200) mm	1.5 μm	
Effective diameter		(3 ~ 200) mm	2.3 μm	
Pitch		(0.3 ~ 10) mm	1.6 μm	
Taper thread ring gauges	10528			Measuring Machine, Standard/ SICT-CP-10528
Alternation		± 3 mm	2.7 μm	
Thickness		(0 ~ 100) mm	2.3 μm	
Notch length		(0 ~ 100) mm	3.2 μm	
V-blocks, Boxblocks	10529			contact coordinate measuring machines/ SICT-CP-10529
Plane figure		(5 ~ 300) mm	1.7 μm	
Parallelism		(5 ~ 300) mm	2.0 μm	
Difference of both part		(5 ~ 300) mm	2.8 μm	

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/Outside/Geartooth calipers, Caliper gauges	10601	Inside/Outside calipers (0 ~ 2 000) mm Caliper gauges (0 ~ 300) mm	$\sqrt{8.2^2 + (0.007 \times l_0)^2}$ μm $\sqrt{3.7^2 + (0.0032 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10601
Cylinder/bore gauges Cylinder gauges Hole gauges	10603	(0 ~ 1 000) mm (0.1 ~ 25) mm	1.0 μm 3.3 μm	Dial Gauge Tester/ SICT-CP-10603
Depth gauges, Depth micrometers Depth micrometers Depth gauges	10604	(0 ~ 300) mm (0 ~ 1 000) mm	$\sqrt{0.86^2 + (0.0034 \times l_0)^2}$ μm $\sqrt{5.9^2 + (0.0048 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10604
Dial/digital gauges	10605	(0 ~ 100) mm (0 ~ 25) mm	$\sqrt{0.21^2 + (0.0082 \times l_0)^2}$ μm $\sqrt{0.59^2 + (0.004 \times l_0)^2}$ μm	Dial Gauge Tester/
Grind gauges Depth Straightness	10608	(0 ~ 1) mm (0 ~ 150) mm	1.8 μm 2.5 μm	Electronic micrometer/ SICT-CP-10608
Micro indicators, Test indicators	10609	(0 ~ 5) mm	0.22 μm	Dial Gauge Tester/ SICT-CP-10609
Micrometer heads	10610	(0 ~ 50) mm	0.8 μm	Gauge Block/ SICT-CP-10610
3-points, Micrometers	10611	(2 ~ 200) mm (200 ~ 300) mm	$\sqrt{1.3^2 + (0.0034 \times l_0)^2}$ μm 3 μm	Ring Gauge/ SICT-CP-10611
Inside micrometers Length Accuracy of scale Extension rod	10612	(5 ~ 300) mm (25 ~ 500) mm (13 ~ 500) mm	$\sqrt{1.1^2 + (0.0042 \times l_0)^2}$ μm $\sqrt{1.1^2 + (0.0042 \times l_0)^2}$ μm $\sqrt{1.2^2 + (0.0048 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10612
Outside micrometers Outside micrometers V-anvil micrometers	10613	(0 ~ 25) mm (25 ~ 1 000) mm (1 ~ 85) mm	$\sqrt{0.2^2 + (0.003 \times l_0)^2}$ μm $\sqrt{0.83^2 + (0.003 \times l_0)^2}$ μm 0.8 μm	Gauge Block, cylindrical plug gauges/ SICT-CP-10613
Particle counters (Air) Flow Threshold voltage Counting efficiency (Liquid) Flow Threshold voltage	10615	(0.1 ~ 1) μm (0 ~ 100) L/min (0 ~ 10) V (0 ~ 110) % (0.05 ~ 25) μm (0 ~ 100) mL/min (0 ~ 10) V	0.09 L/min 0.42 mV 4.1 % 1.4 mL/min 0.42 mV	Particle calibration system/ SICT-CP-10615
Standard sieves Sieve opening Wire rod diameter	10617	(0.004 ~ 10) mm (0.004 ~ 130) mm	1.5 μm 2.4 μm	Non-contact coordinate measuring machines/ SICT-CP-10617
Welding gauges Height or depth Rule Angle	10620	(0 ~ 100) mm (0 ~ 100) mm (0 ~ 90)°	8.2 μm 6.0 μm 0.7'	Non-contact coordinate measuring machine, Gauge Block/ SICT-CP-10620

Note 1. l_0 unit : mm

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Auto-hopper scale balances	20102	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	1.2 g 14 g 21 g 49 g 0.10 kg	Hopper Scale Weight/ SICT-CP-20102
Auto-packer scale balances	20103	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg	0.8 g 7.7 g 16 g	Weight/ SICT-CP-20103
Axle weigher balances Portable	20104	(100 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg (5 000 ~ 10 000) kg (10 000 ~ 30 000) kg	0.2 kg 0.4 kg 1 kg 5 kg 10 kg 24 kg	Force Calibration Machine/ SICT-CP-20104
Dial platform scale balances	20106	(0 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg	42 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20106
Dial swing scale balances	20107	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 20) kg (20 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	0.96 g 9.6 g 20 g 48 g 96 g 0.23 kg 0.48 kg 0.96 kg 1.9 kg 4.6 kg	Weight/ SICT-CP-20107
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (500 ~ 1 000) g (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 25) kg (25 ~ 40) kg (40 ~ 60) kg (60 ~ 150) kg (150 ~ 600) kg (600 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	7.0 µg 9.3 µg 14 µg 19 µg 29 µg 0.05 mg 0.10 mg 0.20 mg 0.5 mg 1.0 mg 3 mg 6 mg 16 mg 24 mg 0.30 g 1.2 g 2.0 g 38 g 0.10 kg	Weight/ SICT-CP-20109

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manual swing scale balances	20111	(0 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	38 g 0.19 kg 0.38 kg 0.94 kg	Weight/ SICT-CP-20111
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.09 kg 0.46 kg	Weight/ SICT-CP-20112
Spring scale balances	20113	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg	21 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20113
Weights less than class E2	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	1.0 µg 1.0 µg 1.0 µg 1.0 µg 1.2 µg 1.4 µg 1.7 µg 2.3 µg 2.9 µg 3.5 µg 4.6 µg 5.8 µg 7.0 µg 9.3 µg 12 µg 18 µg 36 µg 95 µg 0.12 mg 0.36 mg 0.94 mg 1.8 mg 3.7 mg	Weights, Mass Comparator/ SICT-CP-20116
less than class F2		(20 ~ 100) kg 50 kg 100 kg	0.12 g 0.21 g	
less than class M1		(100 ~ 200) kg 200 kg	1.0 g	
less than class F2		(200 ~ 1 000) kg 500 kg 1 000 kg	1.3 g 2.1 g	

202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Force measuring devices Force	20202	(0.4 ~ 20) N (20 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 200) kN (200 ~ 500) kN (500 ~ 1 000) kN	6.0×10^{-4} 8.4×10^{-5} 8.5×10^{-5} 7.3×10^{-5} 6.9×10^{-5} 9.2×10^{-5} 9.0×10^{-5} 8.4×10^{-5} 8.7×10^{-5} 3.6×10^{-4} 4.1×10^{-4} 4.8×10^{-4} 4.5×10^{-4} 4.3×10^{-4} 4.5×10^{-4}	Force Calibration Machine/ SICT-CP-20202
Tension/compression testing machines tensile compression	20203	0.1 N ~ 2 kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (0.1 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (300 ~ 500) kN (500 ~ 1 000) kN (1 000 ~ 3 000) kN	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.2×10^{-3} 1.5×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.6×10^{-3}	Weights, Force Measuring Device/ SICT-CP-20203
push-pull gauge Force	20204	(0.02 ~ 0.2) N 0.2 N ~ 2 kN (2 ~ 5) kN	1.4×10^{-2} 1.3×10^{-3} 8.4×10^{-4}	Weights, Force Calibration Machine/ SICT-CP-20204

203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices Torque	20302	(0.001 ~ 1) N·m (1 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	3.3×10^{-3} 4.1×10^{-4} 7.5×10^{-4} 2.3×10^{-4} 3.4×10^{-4} 2.1×10^{-4} 2.2×10^{-4} 1.5×10^{-4} 1.6×10^{-4}	Torque Calibration Machine/ SICT-CP-20302
Torque wrenches/drivers Torque	20303	(0.02 ~ 0.1) N·m (0.1 ~ 0.5) N·m (0.5 ~ 1) N·m (1 ~ 2) N·m (2 ~ 5) N·m (5 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	1.4×10^{-2} 9.5×10^{-3} 7.8×10^{-3} 6.2×10^{-3} 4.6×10^{-3} 4.5×10^{-3} 4.7×10^{-3} 4.5×10^{-3} 4.9×10^{-3} 3.8×10^{-3} 3.7×10^{-3} 3.8×10^{-3} 2.8×10^{-3}	Torque Measuring Device/ SICT-CP-20303

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Altimeters	20401	(-4 000 ~ 20 000) m (20 000 ~ 47 000) m	12 m 15 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20401
Manometers	20402	(0 ~ 200) kPa	2.9×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20402
Pneumatic pressure ballances	20403	(5 ~ 7 000) kPa	6.1×10^{-5}	분동식 압력계/ SICT-CP-20403
Hydraulic pressure ballances	20404	(2 ~ 200) MPa	1.2×10^{-4}	분동식 압력계/ SICT-CP-20404
Air data test systems	20405	Static pressure (-2 500 ~ 20 000) m (20 000 ~ 30 500) m Dynamic pressure (0 ~ 342) km/hr (342 ~ 2 122) km/hr	0.8 m 7 m 0.1 km/hr 0.3 km/hr	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20405
Absolute pressure gauges	20406	Dail, digital (5 ~ 7 000) kPa abs (7 ~ 200) MPa abs	8.4×10^{-5} 1.2×10^{-4}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20406
Blood pressure gauges	20407	(0 ~ 40) kPa	0.01 kPa	Digital Manometer/ SICT-CP-20407
Compound pressure gauges	20408	(-95 ~ 7 000) kPa	7.9×10^{-5}	Air Dead Weight Tester/ SICT-CP-20408
Differential pressure gauges	20409	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 7 000) kPa	0.13 Pa 1.5 Pa 2.2×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20409
Gauge pressure gauges	20411	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 7 000) kPa (7 ~ 200) MPa (200 ~ 500) MPa	0.13 Pa 1.3 Pa 7.9×10^{-5} 1.2×10^{-4} 3.5×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20411
Pressure transducers/transmitters	20412	Absolute pressure 5 kPa abs ~ 200 MPa abs Gauge pressure (0 ~ 500) Pa (500 ~ 5 000) Pa 5 kPa ~ 500 MPa	6.7×10^{-4} 0.13 Pa 1.2 Pa 6.7×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20412
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	0.059 kPa	Air Dead Weight Tester, SICT-CP-20413

205. Vacuum

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance diaphragm gauges Vacuum	20501	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 12 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20501
Spinning rotor gauges Vacuum	20502	0.15 mPa abs ~ 0.01 Pa abs	3.4×10^{-2}	Baratron gauge, SRG / SICT-CP-20502
Ionization gauges Vacuum	20503	0.093 mPa abs ~ 0.15 mPa abs 0.15 mPa abs ~ 0.01 Pa abs	6.0×10^{-2} 3.5×10^{-2}	Baratron gauge, SRG, Ion / SICT-CP-20503
Thermal conductivity gauges; pirani, thermocouple, convectron etc. Vacuum	20504	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 13 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20504
Standard leaks, Helium leak detectors Leak	20505	22.0 $\mu\text{Pa m}^3/\text{s}$ 1.60 $\mu\text{Pa m}^3/\text{s}$ 0.51 $\mu\text{Pa m}^3/\text{s}$ 15.0 nPa m^3/s 6.4 nPa m^3/s 0.24 nPa m^3/s	4.8 $\mu\text{Pa m}^3/\text{s}$ 0.38 $\mu\text{Pa m}^3/\text{s}$ 0.098 $\mu\text{Pa m}^3/\text{s}$ 3.2 nPa m^3/s 1.3 nPa m^3/s 0.049 nPa m^3/s	Standard leaks, Helium leak detectors / SICT-CP-20505

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0 ~ 0.1) ml (0.1 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (2 ~ 10) ml (10 ~ 25) ml (25 ~ 50) ml (50 ~ 100) ml (100 ~ 200) ml (200 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml (2 000 ~ 5 000) ml (5 000 ~ 10 000) ml	0.31 μ l 0.40 μ l 0.43 μ l 1.1 μ l 1.7 μ l 2.5 μ l 4.3 μ l 5.1 μ l 7.9 μ l 13 μ l 43 μ l 69 μ l 92 μ l 0.17 ml 0.49 ml 0.87 ml	Weight, balances / SICT-CP-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 250) ml (250 ~ 500) ml	2.4 μ l 4.4 μ l 10 μ l 21 μ l	Weight, balances / SICT-CP-20602
Standard volume vessels	20604	(0 ~ 20) L (20 ~ 200) L (200 ~ 10 000) L	9.0×10^{-5} 1.3×10^{-4} 1.1×10^{-3}	Balances, Master Meter, Standard volume vessel/ SICT-CP-20604
Concrete air content meters	20605	(0 ~ 10) %	0.032 %	Weight, balances / SICT-CP-20605
Piston type volume meters	20606	(0 ~ 1) μ l (1 ~ 2) μ l (2 ~ 5) μ l (5 ~ 10) μ l (0.01 ~ 0.02) ml (0.02 ~ 0.05) ml (0.05 ~ 0.1) ml (0.1 ~ 0.2) ml (0.2 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml (20 ~ 50) ml (50 ~ 100) ml	0.004 μ l 0.005 μ l 0.006 μ l 0.008 μ l 0.018 μ l 0.035 μ l 0.047 μ l 0.14 μ l 0.35 μ l 0.65 μ l 1.6 μ l 1.9 μ l 2.4 μ l 5.0 μ l 14 μ l 64 μ l	Weight, balances / SICT-CP-20606

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid density meters	20702	(0.7 ~ 1.4) g/cm ³	0.000 078 g/cm ³	Density standard materials/ SICT-CP-20702
Salinity meters	20704	(0 ~ 10) % (10 ~ 26) %	0.004 % 0.007 %	NaCl/ SICT-CP-20704
Sucrose meters	20705	(0 ~ 20) % (20 ~ 60) % (60 ~ 80) %	0.014 % 0.016 % 0.025 %	Sucrose/ SICT-CP-20705
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706	density (0.600 ~ 0.700) g/cm ³ (0.700 ~ 0.800) g/cm ³ (0.800 ~ 0.900) g/cm ³ (0.900 ~ 1.000) g/cm ³ (1.000 ~ 1.100) g/cm ³ (1.100 ~ 1.200) g/cm ³ (1.200 ~ 1.300) g/cm ³ (1.300 ~ 1.400) g/cm ³ (1.400 ~ 1.500) g/cm ³ (1.500 ~ 1.600) g/cm ³ (1.600 ~ 1.700) g/cm ³ (1.700 ~ 1.800) g/cm ³ (1.800 ~ 1.900) g/cm ³ (1.900 ~ 2.000) g/cm ³ (2.000 ~ 2.200) g/cm ³ (2.200 ~ 3.000) g/cm ³ (3.000 ~ 3.600) g/cm ³ (3.600 ~ 4.000) g/cm ³ specific gravity 0.590 ~ 0.700 0.700 ~ 0.800 0.800 ~ 0.900 0.900 ~ 1.000 1.000 ~ 1.100 1.100 ~ 1.200 1.200 ~ 1.300 1.300 ~ 1.400 1.400 ~ 1.500 1.500 ~ 1.600 1.600 ~ 1.800 1.800 ~ 2.000 2.000 ~ 2.020 2.020 ~ 2.500 2.500 ~ 3.000	0.000 035 g/cm ³ 0.000 038 g/cm ³ 0.000 042 g/cm ³ 0.000 046 g/cm ³ 0.000 050 g/cm ³ 0.000 055 g/cm ³ 0.000 059 g/cm ³ 0.000 066 g/cm ³ 0.000 071 g/cm ³ 0.000 075 g/cm ³ 0.000 079 g/cm ³ 0.000 084 g/cm ³ 0.000 088 g/cm ³ 0.000 093 g/cm ³ 0.000 25 g/cm ³ 0.000 28 g/cm ³ 0.000 30 g/cm ³ 0.000 32 g/cm ³ 0.000 068 0.000 069 0.000 072 0.000 075 0.000 078 0.000 082 0.000 086 0.000 091 0.000 096 0.000 10 0.000 11 0.000 12 0.000 26 0.000 60 0.000 61	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-1 Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-2

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
alcohol (Volum)		(0 ~ 10) % (10 ~ 30) % (30 ~ 40) % (40 ~ 50) % (50 ~ 60) % (60 ~ 70) % (70 ~ 80) % (80 ~ 90) % (90 ~ 100) %	0.039 % 0.043 % 0.038 % 0.030 % 0.025 % 0.023 % 0.020 % 0.019 % 0.017 %	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-3
API		-1 ~ 51 51 ~ 101	0.013 0.014	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-4
Baumé-light		10 ~ 30 30 ~ 40 40 ~ 60 60 ~ 70 70 ~ 100	0.015 0.016 0.018 0.019 0.12	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
Baumé - heavy		0 ~ 40 40 ~ 75	0.014 0.013	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
sugar		(0 ~ 10) % (10 ~ 90) %	0.018 % 0.017 %	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-6
milk		(15 ~ 20) (20 ~ 40)	0.081 0.082	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-7
Bouyoucos		(-5.0 ~ 60.0) g/L	0.14 g/L	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-8
salinity		(0 ~ 26.4) %	0.025 %	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-9
LPG		(0.50 ~ 0.55) g/cm ³ (0.55 ~ 0.60) g/cm ³ (0.60 ~ 0.65) g/cm ³	0.000 065 g/cm ³ 0.000 066 g/cm ³ 0.000 068 g/cm ³	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-10

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc. Twaddell	20706	0 ~ 12 12 ~ 74 74 ~ 102 102 ~ 170 170 ~ 200	0.016 0.059 0.060 0.061 0.062	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
Chloride meters	20707	(0 ~ 0.1) % (0.1 ~ 2.0) %	0.000 2 % 0.001 0 %	Chlorine standard liquid/ SICT-CP-20707
Solid density Stainless steel Glass	20799	(1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 500) g (1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 500) g	0.004 6 g/cm ³ 0.002 4 g/cm ³ 0.001 0 g/cm ³ 0.000 59 g/cm ³ 0.000 43 g/cm ³ 0.000 37 g/cm ³ 0.000 36 g/cm ³ 0.000 46 g/cm ³ 0.000 25 g/cm ³ 0.000 14 g/cm ³ 0.000 12 g/cm ³ 0.000 11 g/cm ³	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-11

* 20704, 20705, 20706, 20707 unit % is weight percent.

208. Viscosity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Kinematic viscometers; capillary,etc	20801	(2.5 ~ 100 000) mPa·s	1.3×10^{-2}	Viscosity CRM/ SICT-CP-20801
Dynamic viscometers; rotaional, etc Viscosity	20802	(2.5 ~ 200 000) mPa·s	1.4×10^{-2}	Viscosity CRM/ SICT-CP-20802

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Anemometers; hot-wire	20901	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20901
Anemometers; pitot tube, etc.	20902	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20902
Gas flowmeters;differential pressure	20908	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m^3/h (300 ~ 4 000) m^3/h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters;differential pressure	20909	(0.01 ~ 50) m^3/h (0.000 12 ~ 0.01) m^3/h (0.01 ~ 50) m^3/h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Liquid flowmeters; electromagnetic	20910	(0.01 ~ 50) m^3/h (0.000 12 ~ 0.01) m^3/h (0.01 ~ 50) m^3/h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;thermal mass, etc.	20911	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m^3/h (300 ~ 4 000) m^3/h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; Coriolis, etc.	20912	(0.01 ~ 50) m^3/h (0.000 12 ~ 0.01) m^3/h (0.01 ~ 50) m^3/h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;open channel, etc.	20914	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m^3/h (300 ~ 4 000) m^3/h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; positive displacement	20915	(0.01 ~ 50) m^3/h (0.000 12 ~ 0.01) m^3/h (0.01 ~ 50) m^3/h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;turbine	20916	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m^3/h (300 ~ 4 000) m^3/h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters;turbine	20917	(0.01 ~ 50) m^3/h (0.000 12 ~ 0.01) m^3/h (0.01 ~ 50) m^3/h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;ultrasonic	20918	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m^3/h (300 ~ 4 000) m^3/h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid flowmeters;ultrasonic	20919	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;variable area	20920	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters;variable area	20921	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;vortex	20922	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters;vortex	20923	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Anemometers; vane, etc	20925	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20925
Others; Ultrasonic current meter	20999	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20999

210. Hardness

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Brinell hardness testers	21001	(100 ~ 250) HBW 10/3 000 (250 ~ 450) HBW 10/3 000 (450 ~ 650) HBW 10/3 000	3.1 HBW 10/3 000 4.9 HBW 10/3 000 8.2 HBW 10/3 000	Brinell Hardness CRM/ SICT-CP-21001
Rockwell hardness testers	21002	(20 ~ 70) HRC (20 ~ 100) HRBW (42 ~ 86) HR30N (29 ~ 82) HR30TW	0.45 HRC 0.80 HRBW 0.70 HR30N 1.1 HR30TW	Rockwell Hardness CRM/ SICT-CP-21002
Shore hardness testers	21003	(20 ~ 100) HS	0.9 HS	Shore Hardness CRM/ SICT-CP-21003
Vickers hardness testers	21004	(50 ~ 300) HV 0.2 (300 ~ 600) HV 0.2 (600 ~ 850) HV 0.2 (50 ~ 300) HV 0.3 (300 ~ 600) HV 0.3 (600 ~ 850) HV 0.5 (50 ~ 300) HV 0.5 (300 ~ 600) HV 0.5 (600 ~ 850) HV 1 (50 ~ 300) HV 10 (300 ~ 600) HV 10 (600 ~ 850) HV 10 (300 ~ 600) HV 30 (600 ~ 850) HV 30	6.0 HV 0.2 18 HV 0.2 27 HV 0.2 5.0 HV 0.3 14 HV 0.3 26 HV 0.5 6.0 HV 0.5 15 HV 0.5 20 HV 1 3.0 HV 10 8.0 HV 10 11 HV 10 8.0 HV 30 11 HV 30	Vickers Hardness CRM/ SICT-CP-21004
Durometer hardness testers	21005	(0 ~ 100) HDA (0 ~ 100) HDAM (0 ~ 100) HDA0 (0 ~ 100) HDB (0 ~ 100) HDC (0 ~ 100) HDC2 (0 ~ 100) HDCS (0 ~ 100) HDD (0 ~ 100) HDD0 (0 ~ 100) HDE (0 ~ 100) HDE2 (0 ~ 100) HDF (0 ~ 100) HDF0 (0 ~ 100) HDM (0 ~ 100) HD0 (0 ~ 100) HD00 (0 ~ 100) HD000 (0 ~ 100) HD000-S	0.4 HDA 0.8 HDAM 0.4 HDA0 0.4 HDB 0.3 HDC 0.6 HDC2 0.3 HDCS 0.3 HDD 0.3 HDD0 0.4 HDE 0.6 HDE2 0.6 HDF 0.6 HDF0 0.8 HDM 0.3 HD0 0.4 HD00 0.4 HD000 0.3 HD000-S	Durometer Calibration device/ SICT-CP-21005
Leeb hardness testers	21006	(400 ~ 1 000) HLD	4.6 HLD	Leeb Hardness CRM/ SICT-CP-21006

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards Time Base Frequency	30102	(1 ~ 100) MHz	4.3×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30102
General frequency sources Time Base Frequency	30103	(10 ~ 100) kHz (0.1 ~ 100) MHz	1.3×10^{-11} 2.2×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30103
Frequency meters/counters Time Base Frequency Input Frequency	30104	(1 ~ 10) MHz 0.1 Hz 0.1 Hz ~ 40 GHz	4.3×10^{-12} 6.4×10^{-10} 6.4×10^{-11}	GPS Receiver, Universal Counter/ SICT-CP-30104
Time interval sources Period Time interval	30105	1 ns ~ 10 s (1 ~ 100) ns 100 ns ~ 1 ms 1 ms ~ 10 s	6.1×10^{-9} 0.15 ns 1.3 ns 2.1 ns	GPS Receiver, Universal Counter/ SICT-CP-30105
Time interval meters /Stop watches/Timers Trigger Voltage Period Reference Frequency Relative Time Difference Time rate Timer Count	30106	(-5 ~ 5) V (5 ~ 100) ns (1 ~ 10) MHz day month (-9.95 ~ 9.95) s / day (-300 ~ 300) s / month (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s ≥ 1	1.2×10^{-4} 6.2×10^{-5} ns 6.2×10^{-11} 1.1×10^{-7} 3.6×10^{-7} 6.1 ms 6.2 ms 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5} 0.58	Stop Watch Calibrator/ SICT-CP-30106

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators Revolution Velocity Measurement Revolution Velocity Measurement (Centrifuge)	30201	(1 ~ 10 000) min ⁻¹ (30 ~ 5 000) min ⁻¹ (5 000 ~ 8 500) min ⁻¹ (8 500 ~ 50 000) min ⁻¹ (50 000 ~ 80 000) min ⁻¹ (80 000 ~ 99 000) min ⁻¹	0.004 0 min ⁻¹ 0.059 min ⁻¹ 0.099 min ⁻¹ 0.59 min ⁻¹ 0.93 min ⁻¹ 1.1 min ⁻¹	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30201
Contact type tachometers Revolution Velocity Measurement	30202	(1 ~ 10) min ⁻¹ (10 ~ 1 000) min ⁻¹ (1 000 ~ 5 000) min ⁻¹	0.10 min ⁻¹ 0.016 min ⁻¹ 0.063 min ⁻¹	GPS Receiver, Tachometer Cal System/ SICT-CP-30202
Photo tachometers/stroboscopes Revolution Velocity Measurement (Photo-tachometer) Revolution Velocity Measurement (Stroboscope)	30203	(1 ~ 999.99) min ⁻¹ (1 000.0 ~ 99 999.9) min ⁻¹ (100 000 ~ 600 000) min ⁻¹ (30 ~ 9 000) min ⁻¹ (9 000 ~ 90 000) min ⁻¹ (90 000 ~ 500 000) min ⁻¹	0.006 1 min ⁻¹ 0.061 min ⁻¹ 0.61 min ⁻¹ 0.005 8 min ⁻¹ 0.058 min ⁻¹ 0.58 min ⁻¹	GPS Receiver, Photo Signal Detector/ SICT-CP-30203
Speed meters Speed Test	30204	(0 ~ 400) km/h	6.1 × 10 ⁻³ km/h	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30204
Wow-flutter generators Carrier Frequency Function Frequency Wow/Flutter Deviation Output Level CCIR Pulse	30205	10 Hz ~ 99.99 kHz 1 Hz ~ 10 kHz (10 ~ 30) kHz (1 Hz ~ 100 Hz) (0 ~ 3) % (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 6) V 10 ms 30 ms 60 ms 100 ms	6.2 × 10 ⁻⁶ 6.2 × 10 ⁻⁶ 2.1 × 10 ⁻⁶ 0.025 % 5.8 × 10 ⁻⁴ 1.7 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 1.0 × 10 ⁻² ms 3.0 × 10 ⁻² ms 6.0 × 10 ⁻² ms 1.0 × 10 ⁻¹ ms	GPS Receiver, Universal Counter/ SICT-CP-30205
Wow-flutter meters Wow/Flutter Deviation Carrier Frequency CCIR Pulse Output Voltage	30206	(0.1 ~ 0.3) % (0.3 ~ 3) % 3 kHz 3.15 kHz (10 ~ 100) ms (1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	0.019 % 0.020 % 6.2 × 10 ⁻⁵ kHz 6.2 × 10 ⁻⁵ kHz 0.59 % 6.8 μV 9.8 μV 76 μV	GPS Receiver, Wow Flutter Calibrator/ SICT-CP-30206

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters DC Current	40101	(±) 0 nA (0 ~ 1) nA (1 ~ 10) nA (10 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 200) mA (0.2 ~ 1) A (1 ~ 10) A (10 ~ 100) A	6 pA 6.6×10^{-3} 9.5×10^{-4} 5.1×10^{-4} 6.1×10^{-4} 2.9×10^{-4} 3.6×10^{-5} 2.9×10^{-5} 2.6×10^{-5} 1.9×10^{-5} 1.3×10^{-5} 2.1×10^{-5} 1.4×10^{-4} 1.5×10^{-4}	Calibrator/ SICT-CP-40101
Transconductance amplifiers DC Current AC Current	40102	(±) 10 μA ~ 10 A (10 ~ 50) A (50 ~ 100) A (10 μA) 10 Hz ~ 10 kHz (10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 μA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.3×10^{-5} 4.3×10^{-5} 4.4×10^{-5} 2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5} 8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	AC-DC Active Current Shunt/ SICT-CP-40102

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transconductance amplifiers AC Current	40102	(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 120) A 60 Hz	 8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5} 9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5} 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-5} 4.6×10^{-4}	AC-DC Active Current Shunt/ SICT-CP-40102
DC voltage/current calibrators DC Voltage DC Current	40103	($\pm$) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V ($\pm$) 0 nA (0 ~ 1) nA (1 ~ 100) nA 100 nA ~ 1 A (1 ~ 10) A (10 ~ 100) A	 0.24 μ V 4.0×10^{-4} 2.0×10^{-4} 8.1×10^{-5} 4.1×10^{-5} 1.0×10^{-5} 6.3×10^{-6} 2.6×10^{-6} 7.0 pA 7.0×10^{-3} 4.7×10^{-3} 1.3×10^{-5} 1.3×10^{-4} 4.4×10^{-5}	Reference Multimeter/ SICT-CP-40103
Electrical temperature calibrators TEMPERATURE(SOURCE) T/C RTD	40104	(-9.835 ~ 0.000) mV 0.000 mV (0.000 ~ 13.421) mV (13.421 ~ 37.006) mV (37.006 ~ 61.017) mV (61.017 ~ 76.373) mV 0.999 Ω (0.999 ~ 2.499) Ω (2.499 ~ 4.322) Ω (4.322 ~ 100.000) Ω (100.000 ~ 177.155) Ω (177.155 ~ 313.708) Ω (313.708 ~ 627.422) Ω (627.422 ~ 3 233.3) Ω	 0.42 μ V 0.24 μ V 0.42 μ V 0.48 μ V 0.53 μ V 0.57 μ V 0.063 m Ω 3.0×10^{-5} 1.9×10^{-5} 9.9×10^{-6} 8.8×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 1.1×10^{-5}	디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators DC Voltage(SOURCE)	40104	(±) 0 mV (1 ~ 2) mV (2 ~ 3) mV (3 ~ 4) mV (4 ~ 5) mV (5 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 0.2) V (0.2 ~ 0.3) V (0.3 ~ 0.6) V (0.6 ~ 1) V (1 ~ 6) V (6 ~ 10) V (10 ~ 70) V (70 ~ 100) V	0.24 μ V 4.0×10^{-4} 2.0×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 8.2×10^{-5} 4.2×10^{-5} 6.3×10^{-6} 6.2×10^{-5} 3.1×10^{-5} 3.1×10^{-5} 9.5×10^{-6} 3.1×10^{-5} 9.3×10^{-6} 6.2×10^{-5} 9.2×10^{-6}	디지털 멀티미터/ SICT-CP-40104
DC Current(SOURCE)		(±) 0 mA (0 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 7) mA (7 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 100) mA	0.064 μ A 9.2×10^{-5} 6.2×10^{-5} 3.5×10^{-5} 2.3×10^{-5} 1.9×10^{-5} 3.3×10^{-5} 8.2×10^{-5} 7.0×10^{-5} 6.3×10^{-5}	
Resistance(SOURCE)		0 Ω (0 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 10) Ω (10 ~ 20) Ω (20 ~ 30) Ω (30 ~ 50) Ω (50 ~ 70) Ω (70 ~ 100) Ω (0.1 ~ 0.2) k Ω (0.2 ~ 0.3) k Ω (0.3 ~ 0.5) k Ω (0.5 ~ 0.8) k Ω (0.8 ~ 1) k Ω (1 ~ 2) k Ω (2 ~ 3) k Ω (3 ~ 5) k Ω (5 ~ 8) k Ω (8 ~ 10) k Ω (10 ~ 20) k Ω (20 ~ 30) k Ω (30 ~ 40) k Ω (40 ~ 50) k Ω (50 ~ 100) k Ω	0.061 m Ω 6.1×10^{-4} 8.9×10^{-5} 6.7×10^{-5} 3.3×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.2×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 2.4×10^{-5} 1.9×10^{-5} 1.6×10^{-5} 1.1×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators	40104			디지털 멀티미터/ SICT-CP-40104
TEMPERATURE(MEASURE)				
T/C		(-9.835 ~ 0.000) mV	0.59 μ V	
		0.000 mV	0.50 μ V	
		(0.000 ~ 13.422) mV	0.62 μ V	
		(13.422 ~ 28.947) mV	0.75 μ V	
		(28.947 ~ 45.094) mV	0.88 μ V	
		(45.094 ~ 53.113) mV	0.95 μ V	
		(53.113 ~ 76.374) mV	1.1 μ V	
RTD		0.998 Ω	0.24 m Ω	
		(0.998 ~ 2.496) Ω	1.0×10^{-4}	
		(2.496 ~ 4.315) Ω	7.1×10^{-5}	
		(4.315 ~ 16.994) Ω	3.9×10^{-5}	
		(16.994 ~ 249.580) Ω	3.5×10^{-5}	
		(249.580 ~ 317.988) Ω	4.3×10^{-5}	
		(317.988 ~ 390.474) Ω	4.0×10^{-5}	
		(390.474 ~ 3 233.2) Ω	3.5×10^{-5}	
DC Voltage(MEASURE)		($\pm$)		
		0 mV	0.50 μ V	
		(1 ~ 5) mV	5.2×10^{-4}	
		(5 ~ 10) mV	9.3×10^{-5}	
		(10 ~ 100) mV	5.9×10^{-5}	
		(0.1 ~ 0.5) V	6.3×10^{-5}	
		(0.5 ~ 0.8) V	1.3×10^{-5}	
		(0.8 ~ 1) V	1.6×10^{-5}	
		(1 ~ 10) V	6.6×10^{-6}	
		(10 ~ 20) V	9.1×10^{-6}	
		(20 ~ 40) V	7.9×10^{-6}	
		(40 ~ 70) V	6.9×10^{-6}	
		(70 ~ 100) V	6.4×10^{-6}	
		(100 ~ 200) V	7.8×10^{-6}	
		(200 ~ 300) V	2.2×10^{-5}	
DC Current(MEASURE)		($\pm$)		
		0 mA	0.062 μ A	
		(0 ~ 0.1) mA	6.4×10^{-4}	
		(0.1 ~ 0.2) mA	3.2×10^{-4}	
		(0.2 ~ 0.3) mA	2.2×10^{-4}	
		(0.3 ~ 0.7) mA	1.7×10^{-4}	
		(0.7 ~ 1) mA	9.3×10^{-5}	
		(1 ~ 2) mA	9.9×10^{-5}	
		(2 ~ 5) mA	7.6×10^{-5}	
		(5 ~ 10) mA	5.8×10^{-5}	
		(10 ~ 20) mA	9.9×10^{-5}	
		(20 ~ 30) mA	8.2×10^{-5}	
		(30 ~ 40) mA	7.4×10^{-5}	
		(40 ~ 50) mA	7.0×10^{-5}	
		(50 ~ 100) mA	6.7×10^{-5}	
		(100 ~ 130) mA	8.7×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators Resistance(MEASURE)	40104	$0\ \Omega$ $(0 \sim 1)\ \Omega$ $(1 \sim 10)\ \Omega$ $(10 \sim 100)\ \Omega$ $(0.1 \sim 1)\ k\Omega$ $(1 \sim 10)\ k\Omega$ $(10 \sim 20)\ k\Omega$ $(20 \sim 30)\ k\Omega$ $(30 \sim 40)\ k\Omega$ $(40 \sim 50)\ k\Omega$ $(50 \sim 100)\ k\Omega$	$0.098\ m\Omega$ 6.4×10^{-5} 1.1×10^{-5} 9.6×10^{-6} 6.5×10^{-5} 6.1×10^{-5} 4.7×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5} 3.4×10^{-5}	디지털 멀티미터/ SICT-CP-40104
DC current shunts Resistance	40105	$1\ \mu\Omega$ $(0.001 \sim 0.01)\ m\Omega$ $(0.01 \sim 0.2)\ m\Omega$ $(0.2 \sim 1)\ m\Omega$ $(1 \sim 10)\ m\Omega$ $(10 \sim 100)\ m\Omega$ $(0.1 \sim 1)\ \Omega$ $(1 \sim 10)\ \Omega$ $(10 \sim 100)\ \Omega$ $(0.1 \sim 1)\ k\Omega$ $(1 \sim 10)\ k\Omega$ $(10 \sim 100)\ k\Omega$ $(0.1 \sim 1)\ M\Omega$ $(1 \sim 10)\ M\Omega$ $(10 \sim 100)\ M\Omega$	$0.32\ n\Omega$ 2.8×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	Trans Conductance Amplifier/ SICT-CP-40105
Galvanometers/null detectors DC Voltage	40106	$(\pm)$ $(100 \sim 300)\ \mu V$ $(0.3 \sim 1)\ mV$ $1\ mV \sim 1\ 000\ V$	1.4×10^{-2} 1.2×10^{-2} 6.8×10^{-3}	Calibrator/ SICT-CP-40106
Potentiometers DC Voltage	40107	$(100 \sim 300)\ \mu V$ $(0.3 \sim 1)\ mV$ $(1 \sim 3)\ mV$ $3\ mV \sim 1\ 000\ V$	5.7×10^{-3} 2.2×10^{-3} 6.0×10^{-4} 3.0×10^{-4}	Calibrator/ SICT-CP-40107

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC power supplies DC Voltage	40108	($\pm$) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μ V 5.8×10^{-4} 5.8×10^{-5} 7.5×10^{-6} 3.3×10^{-6} 7.7×10^{-6} 1.3×10^{-5} 6.6×10^{-5}	DC Electronics Load/ SICT-CP-40108
DC Current		(1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 300) A (300 ~ 500) A (500 ~ 1 000) A (1 000 ~ 3 000) A	5.8×10^{-3} 5.9×10^{-4} 2.4×10^{-4} 3.1×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 4.7×10^{-5} 5.1×10^{-4}	
Load regulation		(0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.16 mV 7.8×10^{-2} 8.2×10^{-3}	
Ripple		(0.1 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	3.8×10^{-1} 1.1×10^{-1} 7.3×10^{-2} 4.4×10^{-2} 7.1×10^{-2}	
DC voltage dividers DC Voltage Ratio	40110	($\pm$) (0.01 ~ 1 000) V (1 ~ 50) kV (50 ~ 100) kV	4.5×10^{-6} 8.8×10^{-5} 8.4×10^{-5}	Calibrator/ SICT-CP-40110
DC voltage standards DC Voltage	40111	1 V 1.018 V 10 V	1.6 μ V 0.8 μ V 3.1 μ V	Null Detector/ SICT-CP-40111
DC voltmeters DC Voltage	40112	($\pm$) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 80) mV (80 ~ 100) mV (0.1 ~ 1 000) V	0.51 μ V 5.0×10^{-4} 2.5×10^{-4} 1.0×10^{-4} 6.2×10^{-5} 5.0×10^{-5} 2.5×10^{-5} 1.0×10^{-5} 6.2×10^{-6} 8.0×10^{-6}	Calibrator/ SICT-CP-40112
Static/Ionic voltmeters DC Voltage	40113	($\pm$) 0 V 0 V ~ 50 kV	68 mV 1.3×10^{-2}	DC Power Supply/ SICT-CP-40113

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators	40201	50 Hz ~ 100 MHz	7.0×10^{-8}	Standard Capacitance Set/ SICT-CP-40201
Frequency		(1 pF)		
		50 Hz ~ 1 kHz	3.5×10^{-4}	
		1 kHz ~ 1 MHz	3.6×10^{-4}	
		2 MHz	4.2×10^{-4}	
		3 MHz	5.4×10^{-4}	
		4 MHz	7.2×10^{-4}	
		5 MHz	9.4×10^{-4}	
		10 MHz	2.5×10^{-3}	
		13 MHz	3.6×10^{-3}	
Capacitance		(10 pF)		
		50 Hz ~ 5 MHz	3.5×10^{-4}	
		10 MHz	3.7×10^{-4}	
		13 MHz	3.8×10^{-4}	
		(100 pF)		
		50 Hz ~ 3 MHz	3.5×10^{-4}	
		4 MHz	3.6×10^{-4}	
		5 MHz	3.7×10^{-4}	
		10 MHz	4.8×10^{-4}	
		13 MHz	6.0×10^{-4}	
		(1 000 pF)		
		50 Hz ~ 1 MHz	3.5×10^{-4}	
		2 MHz	3.8×10^{-4}	
		3 MHz	4.5×10^{-4}	
		4 MHz	5.6×10^{-4}	
		5 MHz	7.2×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	2.9×10^{-3}	
		(10 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-5}	
		(100 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-4}	
		(1 μF)		
		(50 ~ 100) Hz	5.1×10^{-4}	
		100 Hz ~ 10 kHz	8.1×10^{-5}	
		(10 ~ 100) kHz	1.0×10^{-4}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators Capacitance	40201	(10 μ F) 120 Hz ~ 1 kHz (100 μ F) 120 Hz (1 mF) 120 Hz (3 mF) 120 Hz (10 mF) 120 Hz (30 mF) 120 Hz	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 2.9×10^{-3}	Standard Capacitance Set/ SICT-CP-40201
Decade capacitors Capacitance	40202	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F (1 kHz) 1 pF (1 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F	5.7×10^{-5} 4.6×10^{-5} 3.8×10^{-5} 4.6×10^{-5} 2.9×10^{-4} 5.1×10^{-4} 2.5×10^{-5} 2.4×10^{-5} 5.5×10^{-5} 9.3×10^{-5}	Standard Capacitance Set/ SICT-CP-40202
Standard capacitors Capacitance	40204	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F (1 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F (1 pF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	5.2×10^{-5} 4.0×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 2.9×10^{-4} 5.1×10^{-4} 9.1×10^{-6} 7.1×10^{-6} 6.1×10^{-6} 7.1×10^{-6} 5.0×10^{-5} 9.0×10^{-5} 2.4×10^{-4} 2.5×10^{-4} 3.3×10^{-4} 4.7×10^{-4} 6.7×10^{-4} 9.1×10^{-4} 2.5×10^{-3} 3.7×10^{-3}	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors	40204			Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204
Capacitance		(1 ~ 10) pF		
		1 kHz ~ 3 MHz	2.3×10^{-4}	
		(3 ~ 5) MHz	2.4×10^{-4}	
		(5 ~ 10) MHz	2.6×10^{-4}	
		(10 ~ 13) MHz	2.8×10^{-4}	
		(10 ~ 100) pF		
		1 kHz ~ 1 MHz	2.3×10^{-4}	
		(1 ~ 3) MHz	2.4×10^{-4}	
		(3 ~ 4) MHz	2.5×10^{-4}	
		(4 ~ 5) MHz	2.7×10^{-4}	
		(5 ~ 10) MHz	4.0×10^{-4}	
		(10 ~ 13) MHz	5.4×10^{-4}	
		(100 pF ~ 1 nF)		
		1 kHz	2.3×10^{-4}	
		1 kHz ~ 1 MHz	2.4×10^{-4}	
		(1 ~ 2) MHz	2.8×10^{-4}	
		(2 ~ 3) MHz	3.6×10^{-4}	
		(3 ~ 4) MHz	5.0×10^{-4}	
		(4 ~ 5) MHz	6.6×10^{-4}	
		(5 ~ 10) MHz	1.9×10^{-3}	
		(10 ~ 13) MHz	2.8×10^{-3}	
		(1 ~ 100) nF		
		120 Hz ~ 100 kHz	2.3×10^{-4}	
		(100 nF ~ 1 μF)		
		120 Hz	2.4×10^{-4}	
		120 Hz ~ 10 kHz	2.3×10^{-4}	
		(10 ~ 100) kHz	2.4×10^{-4}	
		(1 ~ 10) μF		
		120 Hz ~ 1 kHz	1.2×10^{-3}	
		(30 μF)		
		120 Hz	1.3×10^{-3}	
		(100 μF)		
		120 Hz	1.3×10^{-3}	
		(300 μF)		
		120 Hz	1.5×10^{-3}	
		(1 mF)		
		120 Hz	1.4×10^{-3}	
		(3 mF)		
		120 Hz	1.5×10^{-3}	
		(10 mF)		
		120 Hz	1.4×10^{-3}	
		(30 mF)		
		120 Hz	2.9×10^{-3}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	40205			Decade Resistor/ SICT-CP-40205
Test Voltage		1 V (1 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	6.4×10^{-3} 6.4×10^{-4} 1.3×10^{-3} 6.4×10^{-4} 1.6×10^{-4} 6.4×10^{-4}	
Resistance		1 m Ω (1 ~ 10) m Ω 10 m Ω ~ 100 k Ω	8.6×10^{-4} 7.2×10^{-4} 6.8×10^{-4}	
AC Current out		1 A (1 ~ 3) A (3 ~ 20) A (20 ~ 30) A (30 ~ 60) A (60 ~ 100) A (100 ~ 150) A (150 ~ 200) A	1.2×10^{-3} 1.5×10^{-3} 9.7×10^{-4} 1.0×10^{-3} 8.4×10^{-4} 1.0×10^{-3} 4.6×10^{-3} 3.7×10^{-3}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8×10^{-6} 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Inductance bridges/indicators	40206			Standard Inductor/ SICT-CP-40206
Frequency		50 Hz ~ 100 MHz	7.0×10^{-8}	
Inductance		(1 kHz) 100 μ H 1 mH 10 mH 100 mH 1 H 10 H	1.9×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4}	
Inductors	40208			Standard Inductor/ SICT-CP-40208
Standard Inductance		(1 kHz) 100 μ H 1 mH 10 mH 100 mH 1 H 10 H	28 nH 0.24 μ H 2.4 μ H 24 μ H 0.24 mH 2.5 mH	
Decade Inductance		(1 kHz) 100 μ H ~ 10 H	3.5×10^{-3}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Insulation testers	40210			High Resistance Decade/ SICT-CP-40210
AC Voltage		1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	8.8×10^{-5} 9.0×10^{-5} 1.0×10^{-4} 1.1×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω 10 T Ω	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	
Q-meters	40211			Universal Counter/ SICT-CP-40211
Frequency Test Quality Factor		60 Hz ~ 100 MHz 0 ~ 1 000	7.0×10^{-8} 9.7×10^{-4}	
Resistance bridges & similar instruments Resistance(Rheostat Arm)	40213			Standard Resistance Set/ SICT-CP-40213
		1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	1.5×10^{-6} 7.3×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.4×10^{-7} 2.2×10^{-7} 5.4×10^{-7} 2.0×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	
Resistance(Ratio Arm)		1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	1.5×10^{-6} 7.3×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.4×10^{-7} 2.2×10^{-7} 5.4×10^{-7} 2.0×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance meters	40214			Standard Resistance Set/ SICT-CP-40214
DC Resistance		10 $\mu\Omega$ 100 $\mu\Omega$ 1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω 10 T Ω	3.3×10^{-4} 4.7×10^{-4} 1.3×10^{-6} 8.0×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.3×10^{-7} 2.2×10^{-7} 5.3×10^{-7} 1.9×10^{-6} 1.2×10^{-6} 5.2×10^{-6} 9.8×10^{-6} 3.2×10^{-5} 6.2×10^{-5} 1.3×10^{-4} 2.7×10^{-4} 6.4×10^{-4}	
Frequency		1 kHz	7.0×10^{-8}	
AC Voltage		10 mV (10 ~ 100) mV (0.1 ~ 10) V	1.6×10^{-4} 7.9×10^{-5} 8.2×10^{-5}	
AC Resistance		(1 kHz) 1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω 100 m Ω ~ 10 k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω	5.0×10^{-3} 5.2×10^{-4} 3.3×10^{-4} 1.3×10^{-4} 1.4×10^{-4} 3.0×10^{-4} 2.9×10^{-3}	
Resistors	40215			Standard Resistance Set/ SICT-CP-40215
DC Resistance		1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω (1 ~ 10) T Ω (10 ~ 100) T Ω	1.6×10^{-6} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.2×10^{-6} 9.7×10^{-6} 2.3×10^{-4} 6.9×10^{-4} 9.3×10^{-4} 1.4×10^{-3} 4.1×10^{-3} 7.6×10^{-3}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors AC Resistance	40215	(50 Hz ~ 1 kHz) 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 100) Ω (10 Ω) 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 ~ 100) Ω 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (100 Ω ~ 1 kΩ) 1 kHz 100 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (1 ~ 10) kΩ 1 kHz 100 kHz 1 MHz (10 ~ 100) kΩ 1 kHz 100 kHz 1 MHz (100 kΩ ~ 1 MΩ) 1 kHz (1 ~ 10) MΩ 1 kHz	 1.0×10^{-3} 5.9×10^{-4} 3.9×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 5.6×10^{-4} 6.5×10^{-4} 7.5×10^{-4} 1.0×10^{-3} 4.0×10^{-3} 6.0×10^{-3} 2.4×10^{-4} 4.0×10^{-4} 4.8×10^{-4} 5.6×10^{-4} 5.6×10^{-4} 5.6×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 2.4×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.8×10^{-4} 5.6×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 2.4×10^{-4} 3.3×10^{-4} 4.0×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 3.8×10^{-4} 3.0×10^{-3}	Standard Resistance Set/ SICT-CP-40215

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors Decade Resistance	40215	$0\ \Omega$ $(0 \sim 10)\ m\Omega$ $(10 \sim 100)\ m\Omega$ $(0.1 \sim 1)\ \Omega$ $(1 \sim 7)\ \Omega$ $(7 \sim 10)\ \Omega$ $(10 \sim 70)\ \Omega$ $(70 \sim 100)\ \Omega$ $(100 \sim 700)\ \Omega$ $(0.7 \sim 1)\ k\Omega$ $(1 \sim 7)\ k\Omega$ $(7 \sim 10)\ k\Omega$ $(10 \sim 70)\ k\Omega$ $(70 \sim 100)\ k\Omega$ $(100 \sim 600)\ k\Omega$ $(0.6 \sim 1)\ M\Omega$ $(1 \sim 7)\ M\Omega$ $(7 \sim 10)\ M\Omega$ $(10 \sim 100)\ M\Omega$ $(0.1 \sim 1)\ G\Omega$ $(1 \sim 10)\ G\Omega$ $(10 \sim 100)\ G\Omega$ $(0.1 \sim 1)\ T\Omega$ $(1 \sim 10)\ T\Omega$	$64\ \mu\Omega$ $6.5\ \mu\Omega$ $64\ \mu\Omega$ $66\ \mu\Omega$ 3.9×10^{-5} 1.3×10^{-5} 2.0×10^{-5} 9.8×10^{-6} 1.9×10^{-5} 9.6×10^{-6} 3.5×10^{-5} 1.2×10^{-5} 2.0×10^{-5} 9.8×10^{-6} 2.9×10^{-5} 1.2×10^{-5} 8.0×10^{-5} 2.7×10^{-5} 2.0×10^{-4} 2.5×10^{-4} 7.0×10^{-4} 1.0×10^{-3} 1.5×10^{-3} 4.3×10^{-3}	Standard Resistance Set/ SICT-CP-40215
Impedance bridges/LCR meters Frequency AC Voltage Capacitance	40217	$50\ Hz \sim 100\ MHz$ $1\ mV$ $(1 \sim 10)\ mV$ $(10 \sim 100)\ mV$ $(0.1 \sim 10)\ V$ $(10 \sim 100)\ V$ $(1\ pF)$ $50\ Hz \sim 1\ kHz$ $1\ kHz \sim 1\ MHz$ $2\ MHz$ $3\ MHz$ $4\ MHz$ $5\ MHz$ $10\ MHz$ $13\ MHz$ $(10\ pF)$ $50\ Hz \sim 5\ MHz$ $10\ MHz$ $13\ MHz$	7.0×10^{-8} 2.1×10^{-3} 4.4×10^{-4} 8.8×10^{-5} 8.2×10^{-5} 8.9×10^{-5} 3.5×10^{-4} 3.6×10^{-4} 4.2×10^{-4} 5.4×10^{-4} 7.2×10^{-4} 9.4×10^{-4} 2.5×10^{-3} 3.6×10^{-3} 3.5×10^{-4} 3.7×10^{-4} 3.8×10^{-4}	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters Capacitance	40217	(100 pF) 50 Hz ~ 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (1 000 pF) 50 Hz ~ 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 nF) (50 ~ 100) Hz 100 Hz ~ 100 kHz (100 nF) (50 ~ 100) Hz 100 Hz ~ 100 kHz (1 μF) (50 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (10 μF) 120 Hz ~ 1 kHz (100 μF) 120 Hz (1 mF) 120 Hz (3 mF) 120 Hz (10 mF) 120 Hz (30 mF) 120 Hz (1 pF) 1 kHz ~ 1 MHz 1 MHz ~ 5 MHz 5 MHz ~ 13 MHz	 3.5×10^{-4} 3.6×10^{-4} 3.7×10^{-4} 4.8×10^{-4} 6.0×10^{-4} 3.5×10^{-4} 3.8×10^{-4} 4.5×10^{-4} 5.6×10^{-4} 7.2×10^{-4} 2.0×10^{-3} 2.9×10^{-3} 3.0×10^{-4} 8.1×10^{-5} 3.0×10^{-4} 8.1×10^{-5} 5.1×10^{-4} 8.1×10^{-5} 1.0×10^{-4} 1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 2.9×10^{-3} 0.000 12 0.000 23 0.000 84	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Capacitance		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz	0.000 13	
		5 MHz ~ 13 MHz	0.000 27	
Dissipation Factor		(1 pF) 1 kHz ~ 1 MHz	0.000 12	
		(1 ~ 5) MHz	0.000 23	
		(5 ~ 13) MHz	0.000 84	
		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz	0.000 13	
		(5 ~ 13) MHz	0.000 27	
		(1 000 pF) 1 kHz ~ 1 MHz	0.000 12	
		(1 ~ 5) MHz	0.000 24	
		(5 ~ 13) MHz	0.000 86	
		(10 nF) 120 Hz ~ 100 kHz	0.000 46	
		(100 nF) 120 Hz ~ 100 kHz	0.000 58	
		(1 μF) 120 Hz ~ 100 kHz	0.000 81	
Resistance		(1 mΩ) 50 Hz	6.0×10^{-3}	
		50 Hz ~ 1 kHz	5.0×10^{-3}	
		(10 mΩ) 50 Hz	1.0×10^{-3}	
		50 Hz ~ 1 kHz	5.2×10^{-4}	
		(100 mΩ) 50 Hz	7.1×10^{-4}	
		50 Hz ~ 1 kHz	3.3×10^{-4}	
		(1 Ω) 50 Hz	6.8×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Resistance		(10 Ω)		
		50 Hz	9.1×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	5.2×10^{-4}	
		(2 ~ 3) MHz	6.1×10^{-4}	
		(3 ~ 4) MHz	7.1×10^{-4}	
		(4 ~ 5) MHz	1.0×10^{-3}	
		(5 ~ 10) MHz	4.0×10^{-3}	
		(10 ~ 13) MHz	6.0×10^{-3}	
		(100 Ω)		
		50 Hz	6.2×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	4.2×10^{-4}	
		(2 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(1 kΩ)		
		1 kHz	1.3×10^{-4}	
		1 kHz ~ 3 MHz	3.3×10^{-4}	
		(3 ~ 4) MHz	4.2×10^{-4}	
		(4 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(10 kΩ)		
		1 kHz	1.3×10^{-4}	
		(1 ~ 100) kHz	2.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(100 kΩ)		
		1 kHz	1.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 MΩ)		
		1 kHz	3.0×10^{-4}	
		(10 MΩ)		
		1 kHz	2.9×10^{-3}	
Inductance		(1 kHz)		
		100 μH	1.9×10^{-4}	
		1 mH	1.3×10^{-4}	
		10 mH	1.3×10^{-4}	
		100 mH	1.3×10^{-4}	
		1 H	1.3×10^{-4}	
		10 H	1.3×10^{-4}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	($\pm$)		Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
DC Bias		0 μ V	0.68 μ V	
		0 μ V ~ 100 mV	1.1×10^{-5}	
		(0.1 ~ 1) V	7.5×10^{-6}	
		(1 ~ 10) V	7.2×10^{-6}	
		(10 ~ 100) V	8.2×10^{-6}	
DC Current		0 μ A	5.8 μ A	
		0 μ A ~ 200 mA	4.1×10^{-5}	
		(0.2 ~ 2) A	3.6×10^{-5}	
		(2 ~ 20) A	1.9×10^{-4}	
		(20 ~ 100) A	1.4×10^{-4}	

[illegible]

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Current	40302	(900 μ A ~ 1 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 3) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (3 ~ 9) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (9 ~ 10) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (10 ~ 30) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (30 ~ 90) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (90 ~ 100) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 ~ 300) mA 40 Hz 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (300 ~ 900) mA 40 Hz 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (900 mA ~ 1 A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	 6.3×10^{-4} 7.2×10^{-4} 2.1×10^{-3} 3.8×10^{-4} 7.2×10^{-4} 3.8×10^{-3} 3.1×10^{-4} 5.8×10^{-4} 3.1×10^{-3} 6.3×10^{-4} 7.0×10^{-4} 2.0×10^{-3} 3.6×10^{-4} 6.8×10^{-4} 2.6×10^{-3} 2.9×10^{-4} 5.4×10^{-4} 2.2×10^{-3} 6.3×10^{-4} 6.9×10^{-4} 1.6×10^{-3} 3.3×10^{-4} 5.7×10^{-4} 1.2×10^{-3} 9.3×10^{-3} 5.1×10^{-4} 4.8×10^{-4} 1.0×10^{-3} 8.9×10^{-3} 7.0×10^{-4} 9.0×10^{-4} 5.1×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Current	40302	(1 ~ 2) A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 3) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (3 ~ 9) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (9 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (10 ~ 30) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (30 ~ 90) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (90 ~ 100) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (100 ~ 1 000) A 40 Hz 40 Hz ~ 1 kHz (1 000 ~ 2 500) A (40 ~ 60) Hz (2 500 ~ 3 000) A 60 Hz (3 000 ~ 10 000) A 60 Hz	 4.3×10^{-4} 6.6×10^{-4} 4.6×10^{-3} 5.7×10^{-4} 9.7×10^{-4} 4.5×10^{-3} 4.4×10^{-4} 8.2×10^{-4} 4.4×10^{-3} 6.4×10^{-4} 8.5×10^{-4} 4.3×10^{-3} 3.6×10^{-4} 7.0×10^{-4} 6.2×10^{-2} 2.5×10^{-4} 6.4×10^{-4} 5.5×10^{-2} 1.8×10^{-4} 5.4×10^{-4} 4.3×10^{-2} 1.3×10^{-3} 2.5×10^{-3} 1.3×10^{-3} 1.3×10^{-3} 3.6×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			Power Calibrator, Calibrator/ SICT-CP-40302
DC Current		0 μ A (0 ~ 0.1) μ A (0.1 ~ 0.2) μ A (0.2 ~ 0.3) μ A (0.3 ~ 0.9) μ A (0.9 ~ 1) μ A (1 ~ 2) μ A (2 ~ 3) μ A (3 ~ 9) μ A (9 ~ 10) μ A (10 ~ 20) μ A (20 ~ 30) μ A (30 ~ 50) μ A (50 ~ 90) μ A (90 ~ 100) μ A (100 ~ 200) μ A (200 ~ 700) μ A (700 ~ 900) μ A (0.9 ~ 1) mA (1 ~ 2) mA (2 ~ 7) mA (7 ~ 9) mA (9 ~ 10) mA (10 ~ 20) mA (20 ~ 70) mA (70 ~ 90) mA (90 ~ 100) mA (100 ~ 200) mA (200 ~ 700) mA (700 ~ 900) mA (0.9 ~ 1) A (1 ~ 2) A (2 ~ 3) A (3 ~ 7) A (7 ~ 9) A (9 ~ 10) A (10 ~ 30) A (30 ~ 70) A (70 ~ 100) A (100 ~ 2 500) A	9.2 nA 9.2×10^{-2} 4.6×10^{-2} 3.1×10^{-2} 2.3×10^{-2} 9.3×10^{-3} 4.7×10^{-3} 3.1×10^{-3} 2.3×10^{-3} 9.6×10^{-4} 5.1×10^{-4} 3.5×10^{-4} 2.7×10^{-4} 1.9×10^{-4} 6.2×10^{-4} 3.2×10^{-4} 2.2×10^{-4} 9.3×10^{-5} 6.1×10^{-4} 3.1×10^{-4} 2.2×10^{-4} 9.1×10^{-5} 6.1×10^{-4} 3.1×10^{-4} 2.2×10^{-4} 9.9×10^{-5} 6.1×10^{-4} 3.1×10^{-4} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-4} 3.4×10^{-4} 4.5×10^{-4} 3.6×10^{-4} 2.2×10^{-4} 6.4×10^{-4} 3.4×10^{-4} 2.6×10^{-4} 1.7×10^{-4} 1.3×10^{-3}	
AC Voltage		(1 mV) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (1 ~ 2) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	 4.8×10^{-3} 5.0×10^{-3} 6.5×10^{-3} 2.4×10^{-3} 2.6×10^{-3} 3.5×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Voltage	40302	(2 ~ 5) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (5 ~ 7) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (7 ~ 9) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (9 ~ 10) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (10 ~ 30) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (30 ~ 50) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (50 ~ 70) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (70 ~ 90) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (90 ~ 100) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	 1.7×10^{-3} 1.9×10^{-3} 2.7×10^{-3} 8.9×10^{-4} 1.0×10^{-3} 1.6×10^{-3} 6.9×10^{-4} 8.4×10^{-4} 1.4×10^{-3} 8.3×10^{-4} 9.4×10^{-4} 1.3×10^{-3} 4.5×10^{-4} 5.6×10^{-4} 1.1×10^{-3} 3.2×10^{-4} 4.0×10^{-4} 9.0×10^{-4} 2.3×10^{-4} 3.1×10^{-4} 7.1×10^{-4} 1.9×10^{-4} 2.6×10^{-4} 6.2×10^{-4} 1.6×10^{-4} 2.4×10^{-4} 5.7×10^{-4} 1.1×10^{-4} 1.8×10^{-4} 4.6×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Voltage	40302	(200 ~ 500) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (500 ~ 700) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (700 ~ 900) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (900 mV ~ 1 V) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (1 ~ 2) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 5) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (5 ~ 7) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (7 ~ 9) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (9 ~ 10) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (10 ~ 20) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	 8.8×10^{-5} 1.4×10^{-4} 2.2×10^{-4} 7.1×10^{-5} 1.0×10^{-4} 1.6×10^{-4} 6.5×10^{-5} 9.5×10^{-5} 1.4×10^{-4} 8.6×10^{-5} 1.1×10^{-4} 1.5×10^{-4} 6.4×10^{-5} 9.0×10^{-5} 1.2×10^{-4} 1.3×10^{-4} 9.7×10^{-5} 1.5×10^{-4} 2.2×10^{-4} 8.2×10^{-5} 6.8×10^{-5} 1.1×10^{-4} 1.5×10^{-4} 7.0×10^{-5} 6.2×10^{-5} 9.9×10^{-5} 1.3×10^{-4} 8.9×10^{-5} 8.4×10^{-5} 1.1×10^{-4} 1.4×10^{-4} 6.2×10^{-5} 9.0×10^{-5} 1.1×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			Power Calibrator, Calibrator/ SICT-CP-40302
AC Voltage		(20 ~ 50) V		
		40 Hz ~ 10 kHz	1.6×10^{-4}	
		(10 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	
		(50 ~ 70) V		
		40 Hz	9.8×10^{-5}	
		40 Hz ~ 10 kHz	8.3×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.4×10^{-4}	
		(70 ~ 90) V		
		40 Hz	8.5×10^{-5}	
		40 Hz ~ 10 kHz	7.5×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(90 ~ 100) V		
		40 Hz ~ 10 kHz	9.9×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 200) V		
		40 Hz ~ 10 kHz	7.3×10^{-5}	
		(10 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 500) V		
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(500 ~ 1 000) V		
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
DC Voltage		0 mV	61 μ V	
		(0 ~ 10) mV	6.1×10^{-3}	
		(10 ~ 20) mV	3.1×10^{-3}	
		(20 ~ 30) mV	2.0×10^{-3}	
		(30 ~ 60) mV	1.5×10^{-3}	
		(60 ~ 70) mV	8.7×10^{-4}	
		(70 ~ 80) mV	7.6×10^{-4}	
		(80 ~ 100) mV	6.8×10^{-4}	
		(100 ~ 200) mV	3.3×10^{-5}	
		(200 ~ 300) mV	2.2×10^{-5}	
		(300 ~ 800) mV	1.7×10^{-5}	
		(800 ~ 900) mV	9.6×10^{-6}	
		(0.9 ~ 1) V	6.1×10^{-5}	
		(1 ~ 2) V	3.1×10^{-5}	
		(2 ~ 3) V	2.1×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			Power Calibrator, Calibrator/ SICT-CP-40302
DC Voltage		(3 ~ 6) V (6 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 20) V (20 ~ 30) V (30 ~ 80) V (80 ~ 90) V (90 ~ 100) V (100 ~ 200) V (200 ~ 300) V (300 ~ 500) V (500 ~ 900) V (900 ~ 1 000) V	1.6×10^{-5} 9.8×10^{-6} 8.8×10^{-6} 6.1×10^{-5} 3.1×10^{-5} 2.2×10^{-5} 1.7×10^{-5} 9.3×10^{-6} 6.1×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.3×10^{-5} 6.2×10^{-5}	
Resistance	40303	0 Ω (0 ~ 9) Ω (9 ~ 100) Ω (100 ~ 900) Ω (0.9 ~ 9) k Ω (9 ~ 90) k Ω (0.090 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	0.61 m Ω 0.66 m Ω 6.2 m Ω 9.2 m Ω 92 m Ω 1.1 Ω 63 Ω 0.77 k Ω 13 k Ω	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC voltage/current calibrators				
AC Voltage		(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3} 6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3}	

403. AC voltage, current & power

[illegible]

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(50 ~ 200) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.4×10^{-5} 3.3×10^{-5} 8.5×10^{-5}	
		(200 ~ 1 000) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-5} 3.3×10^{-5} 5.8×10^{-4}	
AC Current		(10 μ A) 10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) μ A 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 μ A ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	
		(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Current	40303	(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 200) A 60 Hz	 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4} 4.5×10^{-4}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
Wattmeter calibrators AC Voltage	40304	(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	 1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3} 6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3} 1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3} 1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3}	Power Standard, Counter/ SICT-CP-40304

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			Power Standard, Counter/ SICT-CP-40304
AC Voltage		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(500 mV ~ 1 V)		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	
		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	
		(50 ~ 200) V		
		10 Hz	7.4×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	8.5×10^{-5}	
		(200 ~ 1 000) V		
		10 Hz	7.7×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators AC Current	40304	(10 μ A) 10 Hz ~ 10 kHz (10 ~ 100) μ A 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 μ A ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 200) A 60 Hz	 2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5} 8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5} 8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5} 9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5} 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4} 4.5×10^{-4}	Power Standard, Counter/ SICT-CP-40304

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304	(50 ~ 60) Hz		Power Standard, Counter/ SICT-CP-40304
AC Power		0 mW	0.05 mW	
		(0 ~ 0.22) mW	3.2×10^{-1}	
		(0.22 ~ 1.1) mW	6.3×10^{-2}	
		(1.1 ~ 2.2) mW	3.1×10^{-2}	
		(2.2 ~ 11) mW	6.3×10^{-3}	
		(11 ~ 22) mW	3.1×10^{-3}	
		(22 ~ 44) mW	1.6×10^{-3}	
		(44 ~ 66) mW	1.1×10^{-3}	
		(66 ~ 88) mW	8.0×10^{-4}	
		(88 ~ 110) mW	6.4×10^{-4}	
		(110 ~ 480) mW	2.6×10^{-4}	
		(480 ~ 550) mW	2.3×10^{-4}	
		(0.55 ~ 1.1) W	1.4×10^{-4}	
		(1.1 ~ 5.5) W	1.6×10^{-4}	
		5.5 W ~ 1.1 kW	1.4×10^{-4}	
		(1.1 ~ 2.2) kW	1.5×10^{-4}	
		(2.2 ~ 24) kW	1.4×10^{-4}	
Power Factor		(50 ~ 60) Hz		
		(0 ~ 550) mW		
		-1 ~ 1	2.3×10^{-4}	
		550 mW ~ 24 kW		
		-1 ~ 1	1.5×10^{-4}	
Harmonic Voltage		(50 ~ 60) Hz		
		(0.5 ~ 3) %	0.042 %	
		(3 ~ 10) %	0.052 %	
		(10 ~ 20) %	0.081 %	
Harmonic Current		(50 ~ 60) Hz		
		(0.5 ~ 3) %	0.042 %	
		(3 ~ 10) %	0.052 %	
		(10 ~ 20) %	0.055 %	
Flicker		P_{st} (0.25 ~ 5), (50 Hz)		
		Modulation Frequency		
		8.333 mHz	2.7×10^{-3}	
		16.667 mHz	2.7×10^{-3}	
		58.333 mHz	2.7×10^{-3}	
		325.000 mHz	2.7×10^{-3}	
		916.667 mHz	2.7×10^{-3}	
		13.500 Hz	2.7×10^{-3}	
		33.333 Hz	2.7×10^{-3}	
Frequency		(10 ~ 100) Hz	7.0×10^{-7}	
		(100 ~ 400) Hz	3.8×10^{-7}	
		400 Hz ~ 1 MHz	7.0×10^{-7}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC current shunts AC Resistance	40305	(100 ~ 200) A (60 Hz) 1 mΩ (40 ~ 60) Hz (1 ~ 10) mΩ (10 ~ 100) mΩ (60 Hz ~ 1 kHz) (1 ~ 10) mΩ (10 ~ 100) mΩ (40 Hz ~ 1 kHz) 100 mΩ ~ 100 Ω 100 Ω ~ 10 kΩ	8.4×10^{-4} 2.1×10^{-4} 3.5×10^{-4} 5.7×10^{-4} 3.4×10^{-4} 1.8×10^{-4} 2.3×10^{-4}	Reference Multimeter, Calibrator/ SICT-CP-40305
AC Voltage dividers		(50 Hz) (1 ~ 100) kV (60 Hz) (1 ~ 100) kV	2.6×10^{-4} 1.9×10^{-4}	
Phase angle generators, synchro resolve generators Phase	40306	(-180 ~ 180) ° 50 Hz (50 ~ 500) Hz (500 ~ 1 000) Hz	0.0016° 0.0031° 0.010°	전력 교정기/ SICT-CP-40307
Voltage/current phase angle meters/synchro resolve meters Phase	40307	(50 ~ 60) Hz (-180 ~ 180) °	0.0088°	Power Calibrator/ SICT-CP-40307
Potential transformer test set Ratio	40308	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) ' (1 100 ~ 22 900) V (-19.99 ~ 19.99) %	0.020% $0.70'$ 0.016%	Standard Potential transforme, Ratio transformers/ SICT-CP-40308
Phase		(-680 ~ 680) '	$0.50'$	
Potential transforme Ratio	40309	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) ' (1 100 ~ 22 900) V (-19.99 ~ 19.99) %	0.020% $0.70'$ 0.016%	Standard Potential transforme/ SICT-CP-40309
Phase		(-680 ~ 680) '	$0.50'$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters AC Power Factor	40310	(50 Hz, 60 Hz) -1 ~ 1	1.1×10^{-4}	Power Calibrator/ SICT-CP-40310
AC power meters AC Voltage	40311	(1 mV) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (1 ~ 2) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 5) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (5 ~ 7) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (7 ~ 9) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (9 ~ 10) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (10 ~ 30) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (30 ~ 60) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (60 ~ 200) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	4.8×10^{-3} 5.0×10^{-3} 6.5×10^{-3} 2.4×10^{-3} 2.6×10^{-3} 3.5×10^{-3} 1.7×10^{-3} 1.9×10^{-3} 2.7×10^{-3} 8.9×10^{-4} 1.0×10^{-3} 1.6×10^{-3} 6.9×10^{-4} 8.4×10^{-4} 1.4×10^{-3} 5.7×10^{-4} 7.1×10^{-4} 1.2×10^{-3} 3.6×10^{-4} 4.7×10^{-4} 1.1×10^{-3} 2.9×10^{-4} 3.7×10^{-4} 8.8×10^{-4} 1.9×10^{-4} 2.7×10^{-4} 6.5×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40311

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
AC Voltage		(200 ~ 300) mV		
		40 Hz ~ 10 kHz	8.6×10^{-5}	
		(10 ~ 50) kHz	1.3×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(300 ~ 600) mV		
		40 Hz ~ 10 kHz	7.6×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(600 mV ~ 1 V)		
		40 Hz ~ 10 kHz	6.7×10^{-5}	
		(10 ~ 50) kHz	9.8×10^{-5}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(1 ~ 2) V		
		40 Hz ~ 10 kHz	5.6×10^{-5}	
		(10 ~ 50) kHz	8.5×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(2 ~ 3) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 10 kHz	9.5×10^{-5}	
		(10 ~ 50) kHz	1.5×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(3 ~ 5) V		
		40 Hz	1.1×10^{-4}	
		40 Hz ~ 10 kHz	8.0×10^{-5}	
		(10 ~ 50) kHz	1.3×10^{-4}	
		(50 ~ 100) kHz	1.8×10^{-4}	
		(5 ~ 7) V		
		40 Hz	8.1×10^{-5}	
		40 Hz ~ 10 kHz	6.7×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(7 ~ 20) V		
		40 Hz	7.0×10^{-5}	
		40 Hz ~ 10 kHz	6.1×10^{-5}	
		(10 ~ 50) kHz	9.9×10^{-5}	
		(50 ~ 100) kHz	1.3×10^{-4}	
		(20 ~ 60) V		
		40 Hz	1.6×10^{-4}	
		40 Hz ~ 10 kHz	1.2×10^{-4}	
		(10 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
AC Voltage		(60 ~ 100) V		
		40 Hz	9.0×10^{-5}	
		40 Hz ~ 10 kHz	7.8×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.3×10^{-4}	
		(100 ~ 200) V		
		40 Hz	6.7×10^{-5}	
		40 Hz ~ 10 kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	9.9×10^{-5}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 400) V		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(400 ~ 500) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	5.4×10^{-4}	
		(500 ~ 600) V		
		40 Hz	1.5×10^{-4}	
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
		(600 ~ 700) V		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	8.5×10^{-4}	
		(700 ~ 900) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	7.5×10^{-4}	
		(900 ~ 1 000) V		
		40 Hz	1.2×10^{-4}	
		40 Hz ~ 1 kHz	1.0×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	6.1×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(100 μ A)		Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		40 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(100 ~ 300) μ A		
		40 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.6×10^{-4}	
		(5 ~ 10) kHz	1.8×10^{-3}	
		(300 μ A ~ 2 mA)		
		40 Hz ~ 1 kHz	1.9×10^{-4}	
		(1 ~ 5) kHz	3.8×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(2 ~ 4) mA		
		40 Hz ~ 1 kHz	3.2×10^{-4}	
		(1 ~ 5) kHz	6.9×10^{-4}	
		(5 ~ 10) kHz	3.8×10^{-3}	
		(4 ~ 7) mA		
		40 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.8×10^{-4}	
		(5 ~ 10) kHz	2.7×10^{-3}	
		(7 ~ 20) mA		
		40 Hz ~ 1 kHz	1.9×10^{-4}	
		(1 ~ 5) kHz	3.7×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(20 ~ 30) mA		
		40 Hz ~ 1 kHz	2.9×10^{-4}	
		(1 ~ 5) kHz	6.5×10^{-4}	
		(5 ~ 10) kHz	2.6×10^{-3}	
		(30 ~ 60) mA		
		40 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.2×10^{-4}	
		(5 ~ 10) kHz	2.2×10^{-3}	
		(60 ~ 200) mA		
		40 Hz ~ 1 kHz	1.8×10^{-4}	
		(1 ~ 5) kHz	3.7×10^{-4}	
		(5 ~ 10) kHz	1.7×10^{-3}	
		(200 ~ 300) mA		
		40 Hz	5.8×10^{-4}	
		(40 ~ 60) Hz	1.2×10^{-4}	
		60 Hz ~ 1 kHz	5.3×10^{-4}	
		(1 ~ 5) kHz	1.2×10^{-3}	
		(5 ~ 10) kHz	9.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(300 ~ 500) mA		Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		40 Hz	4.9×10^{-4}	
		(40 ~ 60) Hz	1.1×10^{-4}	
		60 Hz ~ 1 kHz	4.6×10^{-4}	
		(1 ~ 5) kHz	1.0×10^{-3}	
		(5 ~ 10) kHz	8.9×10^{-3}	
		(500 ~ 800) mA		
		40 Hz	4.0×10^{-4}	
		(40 ~ 60) Hz	1.3×10^{-4}	
		60 Hz ~ 1 kHz	3.9×10^{-4}	
		(1 ~ 5) kHz	8.3×10^{-4}	
		(5 ~ 10) kHz	8.6×10^{-3}	
		(800 mA ~ 2 A)		
		40 Hz	3.5×10^{-4}	
		(40 ~ 60) Hz	9.5×10^{-5}	
		60 Hz ~ 1 kHz	3.5×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	8.4×10^{-3}	
		(2 ~ 3) A		
		40 Hz	5.3×10^{-4}	
		(40 ~ 60) Hz	1.5×10^{-4}	
		(60 ~ 100) Hz	5.3×10^{-4}	
		100 Hz ~ 5 kHz	9.4×10^{-4}	
		(5 ~ 10) kHz	4.5×10^{-3}	
		(3 ~ 6) A		
		40 Hz	4.1×10^{-4}	
		(40 ~ 60) Hz	1.3×10^{-4}	
		(60 ~ 100) Hz	4.1×10^{-4}	
		100 Hz ~ 5 kHz	8.1×10^{-4}	
		(5 ~ 10) kHz	4.4×10^{-3}	
		(6 ~ 10) A		
		40 Hz	2.6×10^{-4}	
		(40 ~ 60) Hz	1.1×10^{-4}	
		(60 ~ 100) Hz	2.6×10^{-4}	
		100 Hz ~ 5 kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-3}	
		(10 ~ 50) A		
		40 Hz	2.4×10^{-4}	
		(40 ~ 60) Hz	1.1×10^{-4}	
		(60 ~ 100) Hz	2.4×10^{-4}	
		100 Hz ~ 5 kHz	6.6×10^{-4}	
		(5 ~ 10) kHz	6.2×10^{-2}	
		(50 ~ 100) A		
		(40 ~ 100) Hz	1.9×10^{-4}	
		100 Hz ~ 5 kHz	5.7×10^{-4}	
		(5 ~ 10) kHz	4.8×10^{-2}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		(100 ~ 1 000) A		
		(40 ~ 100) Hz	1.3×10^{-3}	
		100 Hz ~ 1 kHz	2.7×10^{-3}	
		(1 000 ~ 2 500) A		
		(40 ~ 60) Hz	1.2×10^{-3}	
		(2 500 ~ 3 000) A		
		60 Hz	1.3×10^{-3}	
AC Wattage		(50 ~ 60) Hz		
		0 mW	70 μ W	
		(0 ~ 0.22) mW	2.1×10^{-1}	
		(0.22 ~ 1.1) mW	4.1×10^{-2}	
		(1.1 ~ 2.2) mW	2.1×10^{-2}	
		(2.2 ~ 11) mW	4.1×10^{-3}	
		(11 ~ 22) mW	2.1×10^{-3}	
		(22 ~ 44) mW	1.0×10^{-3}	
		(44 ~ 66) mW	7.0×10^{-4}	
		(66 ~ 88) mW	5.3×10^{-4}	
		(88 ~ 110) mW	4.3×10^{-4}	
		(110 ~ 480) mW	2.1×10^{-4}	
		480 mW ~ 12 kW	1.2×10^{-4}	
		(12 ~ 24) kW	6.8×10^{-4}	
		(24 ~ 300) kW	1.2×10^{-3}	
		(300 ~ 600) kW	1.4×10^{-3}	
DC Voltage		0 mV	0.78 μ V	
		(0 ~ 1) mV	8.0×10^{-4}	
		(1 ~ 2) mV	4.0×10^{-4}	
		(2 ~ 3) mV	2.7×10^{-4}	
		(3 ~ 4) mV	2.0×10^{-4}	
		(4 ~ 8) mV	1.6×10^{-4}	
		(8 ~ 9) mV	9.4×10^{-5}	
		(9 ~ 10) mV	8.5×10^{-5}	
		(10 ~ 20) mV	4.5×10^{-5}	
		(20 ~ 30) mV	3.2×10^{-5}	
		(30 ~ 60) mV	2.6×10^{-5}	
		(60 ~ 200) mV	1.8×10^{-5}	
		(200 ~ 300) mV	9.3×10^{-6}	
		(300 ~ 400) mV	8.3×10^{-6}	
		(400 ~ 700) mV	7.8×10^{-6}	
		(700 ~ 900) mV	6.9×10^{-6}	
		(0.9 ~ 1) V	9.0×10^{-6}	
		(1 ~ 2) V	7.2×10^{-6}	
		(2 ~ 3) V	5.6×10^{-6}	
		(3 ~ 9) V	5.1×10^{-6}	
		(9 ~ 10) V	7.5×10^{-6}	
		(10 ~ 20) V	6.0×10^{-6}	
		(20 ~ 40) V	7.9×10^{-6}	
		(40 ~ 90) V	6.9×10^{-6}	
		(90 ~ 200) V	8.8×10^{-6}	
		(200 ~ 400) V	9.9×10^{-6}	
		(400 ~ 900) V	8.7×10^{-6}	
		(900 ~ 1 000) V	1.0×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
DC Current		0 μ A	9.0 nA	
		(0 ~ 0.1) μ A	9.0×10^{-2}	
		(0.1 ~ 0.2) μ A	4.5×10^{-2}	
		(0.2 ~ 0.3) μ A	3.0×10^{-2}	
		(0.3 ~ 0.4) μ A	2.3×10^{-2}	
		(0.4 ~ 0.9) μ A	1.8×10^{-2}	
		(0.9 ~ 1) μ A	9.1×10^{-3}	
		(1 ~ 2) μ A	4.5×10^{-3}	
		(2 ~ 3) μ A	3.0×10^{-3}	
		(3 ~ 4) μ A	2.3×10^{-3}	
		(4 ~ 9) μ A	1.8×10^{-3}	
		(9 ~ 10) μ A	9.4×10^{-4}	
		(10 ~ 20) μ A	5.0×10^{-4}	
		(20 ~ 30) μ A	3.4×10^{-4}	
		(30 ~ 50) μ A	2.7×10^{-4}	
		(50 ~ 90) μ A	1.9×10^{-4}	
		(90 ~ 100) μ A	1.3×10^{-4}	
		(100 ~ 200) μ A	8.6×10^{-5}	
		(200 ~ 300) μ A	6.5×10^{-5}	
		(300 ~ 500) μ A	5.5×10^{-5}	
		(0.5 ~ 2) mA	5.1×10^{-5}	
		(2 ~ 3) mA	7.3×10^{-5}	
		(3 ~ 4) mA	6.3×10^{-5}	
		(4 ~ 7) mA	5.7×10^{-5}	
		(7 ~ 20) mA	4.9×10^{-5}	
		(20 ~ 30) mA	8.2×10^{-5}	
		(30 ~ 50) mA	7.4×10^{-5}	
		(50 ~ 100) mA	6.6×10^{-5}	
		(100 ~ 200) mA	5.7×10^{-5}	
		(200 ~ 600) mA	1.5×10^{-4}	
		(0.6 ~ 2) A	1.1×10^{-4}	
		(2 ~ 3) A	4.0×10^{-4}	
		(3 ~ 4) A	3.2×10^{-4}	
		(4 ~ 7) A	2.8×10^{-4}	
		(7 ~ 10) A	2.1×10^{-4}	
		(10 ~ 20) A	1.5×10^{-4}	
		(20 ~ 30) A	2.5×10^{-4}	
		(30 ~ 80) A	2.1×10^{-4}	
		(80 ~ 100) A	1.5×10^{-4}	
		(100 ~ 2 500) A	1.3×10^{-3}	
DC Wattage		0 mW	61 nW	
		(0 ~ 1) mW	7.7×10^{-5}	
		(1 ~ 10) mW	4.8×10^{-5}	
		(10 ~ 100) mW	6.1×10^{-5}	
		(0.1 ~ 100) W	1.1×10^{-4}	
		(0.1 ~ 10) kW	1.9×10^{-4}	
		(10 ~ 20) kW	1.5×10^{-4}	
		(20 ~ 1 000) kW	1.0×10^{-3}	
		(1 000 ~ 2 500) kW	1.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
Harmonic Voltage		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 5) % (5 ~ 10) % (10 ~ 20) %	0.030 % 0.033 % 0.042 % 0.065 %	
Harmonic Current		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (5 ~ 20) %	0.030 % 0.032 % 0.038 %	
Flicker		P_{st} (0.25 ~ 5), (50 Hz) Modulation Frequency 8.333 mHz 16.667 mHz 58.333 mHz 325 mHz 916.667 mHz 13.5 Hz 33.333 Hz	3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2}	
Frequency		10 Hz ~ 10 MHz	1.3×10^{-4}	
Power Factor		(50 ~ 60) Hz -1 ~ 1	1.1×10^{-4}	
Current burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	2.9×10^{-3} 1.5×10^{-3} 8.0×10^{-4} 5.1×10^{-4} 4.0×10^{-4}	
Current burden factor		0.5 ~ 1	2.2×10^{-4}	
Voltage burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	7.4×10^{-4} 4.6×10^{-4} 3.7×10^{-4} 3.0×10^{-4} 3.1×10^{-4}	
Voltage burden factor		0.5 ~ 1	1.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312			Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
AC Voltage		(10 mV) 40 Hz ~ 5 kHz	2.2×10^{-4}	
		(10 ~ 100) mV 40 Hz ~ 5 kHz	7.9×10^{-5}	
		(100 mV ~ 1 V) 40 Hz ~ 5 kHz	6.7×10^{-5}	
		(1 ~ 10) V 40 Hz ~ 5 kHz	6.8×10^{-5}	
		(10 ~ 100) V 40 Hz ~ 5 kHz	7.0×10^{-5}	
		(100 ~ 600) V 40 Hz ~ 5 kHz	4.1×10^{-5}	
		(600 ~ 1 000) V 40 Hz ~ 5 kHz	7.1×10^{-5}	
Frequency		10 Hz	9.5×10^{-5}	
		(10 ~ 50) Hz	1.9×10^{-5}	
		(50 ~ 100) Hz	7.7×10^{-6}	
		(0.1 ~ 1) kHz	8.4×10^{-7}	
		(1 ~ 5) kHz	3.8×10^{-7}	
AC Current		(1 mA) (50 ~ 60) Hz	6.4×10^{-4}	
		(1 ~ 10) mA (50 ~ 60) Hz	3.6×10^{-4}	
		(10 ~ 100) mA (50 ~ 60) Hz	2.4×10^{-4}	
		(100 mA ~ 1 A) (50 ~ 60) Hz	2.1×10^{-4}	
		(1 ~ 10) A (50 ~ 60) Hz	2.3×10^{-4}	
		(10 ~ 20) A (50 ~ 60) Hz	4.0×10^{-4}	
		(20 ~ 30) A (50 ~ 60) Hz	6.4×10^{-4}	
		(30 ~ 50) A (50 ~ 60) Hz	4.2×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312			Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
DC Voltage		($\pm$) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μ V 5.8×10^{-4} 5.8×10^{-5} 7.7×10^{-6} 1.3×10^{-5} 6.6×10^{-5}	
DC Current		(1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 300) A (300 ~ 500) A (500 ~ 1 000) A (1 000 ~ 3 000) A	5.8×10^{-3} 5.9×10^{-4} 2.4×10^{-4} 3.1×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 4.7×10^{-5} 5.1×10^{-4}	
Load Regulation		(0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.16 mV 7.8×10^{-2} 8.2×10^{-3}	
Ripple		(0.1 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	3.8×10^{-1} 1.1×10^{-1} 7.3×10^{-2} 4.4×10^{-2} 7.1×10^{-2}	
Harmonic Voltage		(50 ~ 60) Hz 0.5 % (0.5 ~ 10) % (10 ~ 20) %	0.050 % 0.051 % 0.082 %	
Harmonic Current		(50 ~ 60) Hz 0.5 % (0.5 ~ 20) %	0.050 % 0.051 %	
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
DC Voltage		($\pm$) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}	
AC Voltage		(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
AC Breaking Current		(50 ~ 60) Hz 0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	5.3×10^{-4} 4.4×10^{-4} 7.3×10^{-4} 7.1×10^{-4} 4.4×10^{-4} 3.6×10^{-4} 7.1×10^{-4} 4.4×10^{-4} 7.3×10^{-4}	
DC Breaking Current		0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	3.9×10^{-4} 1.3×10^{-4} 6.4×10^{-4} 3.2×10^{-4} 1.3×10^{-4} 6.5×10^{-5} 3.3×10^{-4} 1.4×10^{-4} 6.4×10^{-4}	
Resistance		1 m Ω (1 ~ 10) m Ω 10 m Ω ~ 100 k Ω	8.6×10^{-4} 7.2×10^{-4} 6.8×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω 10 T Ω	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Leakage current(DC)		0 μ A (0 ~ 1) μ A (1 ~ 2) μ A (2 ~ 5) μ A (5 ~ 10) μ A (10 ~ 20) μ A (20 ~ 50) μ A (50 ~ 100) μ A (100 ~ 200) μ A (0.2 ~ 100) mA	7.0 nA 7.0×10^{-3} 3.6×10^{-3} 1.4×10^{-3} 7.4×10^{-4} 4.0×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 8.5×10^{-5} 6.1×10^{-4}	
Leakage current(AC)		(20 μ A) 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (20 ~ 50) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 ~ 200) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (200 ~ 500) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz 500 μ A ~ 1 mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.3×10^{-3} 8.5×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 5.5×10^{-3} 6.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 6.8×10^{-4} 2.8×10^{-3} 4.9×10^{-4} 3.2×10^{-4} 2.3×10^{-4} 4.9×10^{-4} 4.0×10^{-4} 3.9×10^{-4} 2.5×10^{-4} 1.7×10^{-4} 4.0×10^{-4} 1.7×10^{-3} 4.4×10^{-4} 3.2×10^{-4} 2.4×10^{-4} 5.4×10^{-4} 2.8×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.2×10^{-4} 2.1×10^{-3}	

403. AC voltage, current & power

[illegible]

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Current transformer test set Ratio	40315	(5 ~ 1 500) A (-19.99 ~ 19.99) %	0.020 %	Current transforme, Ratio transformers/ SICT-CP-40315
Phase		(-680 ~ 680)'	0.70'	
Current transformer Ratio	40316	(5 ~ 10 000) A (-19.99 ~ 19.99) %	0.020 %	Current transforme/ SICT-CP-40316
Phase		(-680 ~ 680)'	0.70'	
Current Coil		(AC) 2 ~ 50	0.10 %	
		(DC) 2 ~ 50	0.10 %	
transducers		(±) (10 A) 50 : 1 ~ 5 000 : 1	1.2×10^{-4}	
		(10 ~ 1 000) A 50 : 1 ~ 5 000 : 1	1.9×10^{-4}	
		(1 000 ~ 2 000) A 50 : 1 ~ 5 000 : 1	2.5×10^{-4}	
AC voltmeters AC Voltage	40318	(600 μV) 1 kHz	7.8×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318
		(600 μV ~ 1 mV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	5.0×10^{-3} 4.8×10^{-3} 6.5×10^{-3}	
		(1 ~ 3) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	2.0×10^{-3} 1.7×10^{-3} 2.8×10^{-3}	
		(3 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-4} 5.7×10^{-4} 1.2×10^{-3}	
		(10 ~ 30) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	8.0×10^{-4} 3.7×10^{-4} 1.1×10^{-3}	
		(30 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.3×10^{-4} 1.7×10^{-4} 5.7×10^{-4}	
		(100 mV ~ 10 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.9×10^{-4} 1.1×10^{-4} 2.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
AC Voltage		(10 ~ 100) V		
		10 Hz	5.3×10^{-4}	
		10 Hz ~ 10 kHz	1.3×10^{-4}	
		(10 ~ 100) kHz	3.6×10^{-4}	
		(100 ~ 1 000) V		
		50 Hz	3.7×10^{-4}	
		50 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 mV)		
		100 kHz	6.5×10^{-3}	
		100 kHz ~ 1 MHz	2.7×10^{-2}	
		(1 ~ 10) mV		
		100 kHz	1.2×10^{-3}	
		100 kHz ~ 1 MHz	5.6×10^{-3}	
		(10 ~ 100) mV		
		100 kHz	5.7×10^{-4}	
		100 kHz ~ 1 MHz	3.7×10^{-3}	
		(100 mV ~ 1 V)		
		100 kHz	1.5×10^{-4}	
		100 kHz ~ 1 MHz	2.3×10^{-3}	
		(1 ~ 10) V		
		100 kHz	4.7×10^{-5}	
		100 kHz ~ 1 MHz	7.0×10^{-4}	
		(10 ~ 20) V		
		100 kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(25 mV)		
		1 MHz	1.9×10^{-2}	
		(1 ~ 30) MHz	2.3×10^{-2}	
		(25 ~ 100) mV		
		1 MHz	2.3×10^{-2}	
		(1 ~ 30) MHz	2.8×10^{-2}	
		(100 ~ 300) mV		
		1 MHz	3.7×10^{-2}	
		(1 ~ 30) MHz	4.0×10^{-2}	
		(300 mV ~ 1 V)		
		1 MHz	2.4×10^{-2}	
		(1 ~ 30) MHz	2.7×10^{-2}	
		(1 ~ 2) V		
		1 MHz	1.5×10^{-2}	
		(1 ~ 30) MHz	1.8×10^{-2}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters AC Output Voltage	40318	(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 10) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz (10 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	 1.9×10^{-3} 3.1×10^{-3} 2.6×10^{-4} 3.7×10^{-4} 9.2×10^{-5} 4.8×10^{-5} 9.1×10^{-5} 7.2×10^{-5} 1.8×10^{-5} 5.9×10^{-5}	Reference Multimeter, Calibrator/ SICT-CP-40318
DC Output Voltage		1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V	7.5×10^{-4} 7.6×10^{-5} 9.4×10^{-6} 2.2×10^{-5}	
Watt hour meters Watt Hour	40319	(50 ~ 60) Hz 0 mWh (0 ~ 480) mWh 480 mWh ~ 12 kWh (12 ~ 24) kWh (24 ~ 300) kWh (300 ~ 600) kWh (DC) 0 mWh (0 ~ 1) mWh (1 ~ 100) mWh 100 mWh ~ 100 Wh 100 Wh ~ 10 kWh (10 ~ 20) kWh (20 ~ 1 000) kWh (1 000 ~ 2 500) kWh	 44 μ Wh 4.0×10^{-4} 3.6×10^{-4} 7.6×10^{-4} 1.3×10^{-3} 1.4×10^{-3} 61 nWh 3.5×10^{-4} 3.4×10^{-4} 3.5×10^{-4} 3.8×10^{-4} 3.7×10^{-4} 1.1×10^{-3} 1.3×10^{-3}	Power Calibrator/ SICT-CP-40319

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Ratio transformers	40321	PT (±) (0.1 ~ 1.000) % (1.000 ~ 19.00) % CT (±) (0.1 ~ 1.000) % (1.000 ~ 19.00) %	0.006 % 0.01 % 0.019 % 0.02 %	Calibrator/ SICT-CP-40321
Phase		PT (±) (0.040 ~ 1.999) ' (1.999 ~ 19.99) ' (19.99 ~ 199.9) ' (199.9 ~ 600) ' CT (±) (0.040 ~ 1.999) ' (1.999 ~ 19.99) ' (19.99 ~ 199.9) ' (199.9 ~ 600) '	0.060 ' 0.06 ' 0.2 ' 1 ' 0.060 ' 0.06 ' 0.2 ' 1 '	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers Amplifier	40401	(DC) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 ~ 100) kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V	0.4 μ V 7.4×10^{-5} 6.1×10^{-5} 6.0×10^{-5} 1.0×10^{-4} 1.7 μ V 2.6×10^{-4} 1.1×10^{-4} 9.0×10^{-5} 1.0×10^{-4} 3.1 μ V 3.7×10^{-4} 1.1×10^{-4} 8.0×10^{-5} 1.0×10^{-4}	Reference Multimeter/ SICT-CP-40401
DC/LF attenuators Attenuation	40402	10 Hz ~ 100 kHz (0 ~ -20) dB (-20 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.001 9 dB 0.001 7 dB 0.005 5 dB 0.008 7 dB	Reference Multimeter/ SICT-CP-40402
Multimeter calibrators DC Voltage DC Current AC Voltage	40403	($\pm$) 0 mV (0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V ($\pm$) 1 nA (1 ~ 100) nA 100 nA ~ 10 A (10 ~ 50) A (50 ~ 100) A (1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	0.05 μ V 3.3×10^{-6} 1.4×10^{-6} 1.0×10^{-6} 1.6×10^{-6} 2.1×10^{-6} 7.0 pA 4.7×10^{-3} 1.2×10^{-5} 4.0×10^{-5} 4.4×10^{-5} 1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3}	Reference Multimeter/ SICT-CP-40403

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators AC Voltage	40403	(2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (100 ~ 200) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (200 ~ 500) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (0.5 ~ 1) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	 6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3} 1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3} 1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3} 8.2×10^{-5} 3.9×10^{-5} 7.6×10^{-5} 1.1×10^{-3} 7.8×10^{-5} 3.6×10^{-5} 7.1×10^{-5} 1.1×10^{-3} 7.6×10^{-5} 3.3×10^{-5} 6.6×10^{-5} 1.1×10^{-3} 7.1×10^{-5} 2.7×10^{-5} 5.8×10^{-5} 1.0×10^{-3}	Reference Multimeter/ SICT-CP-40403

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Voltage		(2 ~ 5) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 20) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (50 ~ 200) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (200 ~ 1 000) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.2×10^{-5} 2.6×10^{-5} 7.5×10^{-5} 1.4×10^{-3} 7.2×10^{-5} 2.8×10^{-5} 7.5×10^{-5} 1.4×10^{-3} 7.2×10^{-5} 3.0×10^{-5} 8.0×10^{-5} 7.4×10^{-5} 3.3×10^{-5} 8.5×10^{-5} 7.7×10^{-5} 3.3×10^{-5} 5.8×10^{-4}	
AC Current		(10 μ A) 10 Hz ~ 10 kHz (10 ~ 100) μ A 10 Hz ~ 1 kHz (1 ~ 10) kHz (0.1 ~ 1) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (0.1 ~ 1) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Current		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 200) A 60 Hz	 8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5} 8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5} 9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5} 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4} 4.5×10^{-4}	
Resistance		0 Ω (0 ~ 1) Ω (1 ~ 1.9) Ω (1.9 ~ 10) Ω (10 ~ 19) Ω (19 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 1.9) kΩ (1.9 ~ 10) kΩ (10 ~ 19) kΩ (19 ~ 100) kΩ (100 ~ 190) kΩ (0.19 ~ 1) MΩ (1 ~ 1.9) MΩ (1.9 ~ 10) MΩ (10 ~ 19) MΩ (19 ~ 100) MΩ	0.14 μΩ 6.6×10^{-6} 8.4×10^{-6} 3.6×10^{-6} 2.6×10^{-6} 2.8×10^{-6} 2.5×10^{-6} 3.8×10^{-6} 2.0×10^{-6} 1.3×10^{-6} 1.9×10^{-6} 2.0×10^{-6} 2.9×10^{-6} 3.1×10^{-6} 3.6×10^{-6} 2.9×10^{-6} 1.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling) AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4}	
		(1 ~ 2) mV 0.1 Hz ~ 3 kHz	4.2×10^{-4}	
		(2 ~ 3) mV 0.1 Hz ~ 3 kHz	2.8×10^{-4}	
		(3 ~ 5) mV 0.1 Hz ~ 3 kHz	1.7×10^{-4}	
		(5 ~ 10) mV 0.1 Hz ~ 3 kHz	8.8×10^{-5}	
		(10 ~ 20) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(20 ~ 30) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) mV 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(50 ~ 100) mV 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(100 ~ 200) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 300) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 ~ 500) mV 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 ~ 2) V 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 ~ 3) V 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 ~ 5) V 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 ~ 10) V 0.1 Hz ~ 3 kHz	2.6×10^{-5}	

[illegible]

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators	40404			Calibrator/ SICT-CP-40404
Sine Wave Generator		(600 mV ~ 1 V) 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	
Time Marker Generator		(0.1 ~ 1) ns (1 ~ 10) ns (10 ~ 100) ns 0.1 μ s ~ 10 ms (10 ~ 100) ms (0.1 ~ 1) s (1 ~ 5) s	5.8×10^{-8} 6.5×10^{-9} 3.1×10^{-9} 5.8×10^{-8} 6.1×10^{-9} 5.8×10^{-8} 1.2×10^{-8}	
Impedance Mesurement		(50 ~ 75) Ω 75 Ω ~ 1 M Ω	1.7×10^{-4} 2.1×10^{-4}	
CD/DVD meters/analyzers	40405			Modulation Domain Analyzer/ SICT-CP-40405
Jitter		(1.0 ~ 60.0) ns 1 % 2 % 4 % 8 % 10 % 15 %	1.7×10^{-3} 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	
Video signal generators	40406			Video Measurement/ SICT-CP-40406
NTSC, PAL Multiburst		(0.1 ~ 1) MHz (1 ~ 2) MHz (2 ~ 6) MHz	6.0×10^{-2} 6.2×10^{-3} 3.1×10^{-3}	
NTSC, PAL, SECAM Pulse and Bar		(0 ~ 300) ns (0 ~ 1 000) mV	4.2×10^{-4} 3.5×10^{-3}	
NTSC, PAL, SECAM Frequency		1 Hz ~ 10 MHz	1.6×10^{-9}	
Video frequency		(10 ~ 100) Hz 100 Hz ~ 500 MHz	6.2×10^{-8} 6.2×10^{-9}	
Video level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	
TTL Sync level		(1 ~ 5) V	2.7×10^{-3}	
D-TV Level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal generators NTSC, PAL, H-Timing(Level) (Time) NTSC, PAL Color Bar(Luminance Level) NTSC, PAL Color Bar(Chrominance Level) NTSC, PAL Color Bar(Phase) SECAM Color Bar Frequency RF Output frequency RF Output level Sound Frequency	40406	(0 ~ 100) mV (100 ~ 1 000) mV (0 ~ 254) ns (254 ~ 300) ns 300 ns ~ 3 μ s (3 ~ 7) μ s (7 ~ 10) μ s (0 ~ 100) mV (100 ~ 1 000) mV (0 ~ 100) mV (100 ~ 1 000) mV (0 ~ 360) $^{\circ}$ (D'R & D'B) (3 ~ 5) MHz 10 kHz ~ 10 MHz (10 ~ 100) MHz (100 ~ 1 000) MHz (0.1 ~ 10) mV (10 ~ 500) mV 10 Hz ~ 100 kHz 100 kHz ~ 1 MHz	2.6×10^{-3} 3.4×10^{-3} 1.2×10^{-2} 3.8×10^{-3} 3.2×10^{-3} 7.4×10^{-3} 4.2×10^{-3} 0.06 mV 3.4×10^{-3} 0.06 mV 3.4×10^{-3} 0.13 $^{\circ}$ 1.2×10^{-3} 6.0×10^{-4} 6.0×10^{-5} 6.0×10^{-6} 1.4×10^{-2} 1.3×10^{-2} 6.1×10^{-8} 6.1×10^{-7}	Video Measurement/ SICT-CP-40406
Audio distortion analyzers/meters Input Frequency Input Level Flatness Test Input DC Voltage Input Distortion	40407	1 Hz ~ 200 kHz (10 ~ 100) kHz 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V (100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	6.1×10^{-7} 0.008 3 dB 0.27 μ V 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4} 0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	Calibrator/ SICT-CP-40407

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters	40407			Calibrator/ SICT-CP-40407
Input Distortion		(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	
Input AC Voltage		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4} 1.4×10^{-3} 7.6×10^{-4} 4.8×10^{-4} 4.1×10^{-4} 3.4×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB 0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 k Ω	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
(Distortion meter calibrator) Distortion		(400 Hz , 1 kHz) (-10 ~ -20) dB (-20 ~ -40) dB (-40 ~ -60) dB (-60 ~ -80) dB	0.15 dB 0.14 dB 0.17 dB 0.26 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF filters Filter	40408	10 Hz ~ 50 kHz (50 ~ 100) kHz (100 ~ 150) kHz	5.8×10^{-4} 1.2×10^{-3} 5.8×10^{-3}	Audio Analyzer/ SICT-CP-40408
LF/Audiosignalanalyzers Output Frequency AC Output Level AC Output Level Flatness Output Attenuation Output DC Offset Output Impedance Input Frequency AC Input Level Flatness DC Input Level Input Distortion	40409	1 Hz ~ 200 kHz (10 ~ 100) Hz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (100 Hz ~ 10 kHz) (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (10 ~ 100) kHz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm 10 Hz ~ 100 kHz (0 ~ -60) dB ($\pm$) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 50 V 5 Ω (10 ~ 600) Ω 1 Hz ~ 200 kHz 10 Hz ~ 100 kHz ($\pm$) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V (100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	5.8×10^{-6} 8.7×10^{-4} 9.4×10^{-5} 0.005 8 dB 8.7×10^{-4} 6.5×10^{-5} 0.005 8 dB 8.7×10^{-4} 9.4×10^{-5} 0.005 8 dB 0.007 1 dB 0.005 8 dB 0.7 μ V 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.2×10^{-3} 6.0×10^{-4} 6.1×10^{-7} 0.008 3 dB 0.27 μ V 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4} 0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	Calibrator, Reference Multimeter/ SICT-CP-40409

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/Audiosignalanalyzers	40409			Calibrator, Reference Multimeter/ SICT-CP-40409
Input Distortion		(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	
AC Input Level		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4} 1.4×10^{-3} 7.6×10^{-4} 4.1×10^{-4} 3.4×10^{-4} 2.6×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB 0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 k Ω	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
Line frequency meters	40410			Calibrator/ SICT-CP-40410
Frequency		16 Hz ~ 1 kHz	1.3×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411			Audio Analyzer, Digital Multimeter/ SICT-CP-40411
Frequency		(0.01 ~ 0.1) Hz (0.1 ~ 1) Hz 1 Hz ~ 1 GHz (1 ~ 4) GHz	5.8×10^{-6} 5.8×10^{-7} 5.8×10^{-9} 1.5×10^{-8}	
Output Level		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V (10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 8.0×10^{-5}	
DC Offset		($\pm$) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	 0.7 μ V 0.7 μ V 1.0×10^{-4} 6.0×10^{-5}	
Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	 0.099 dB 0.083 dB 0.095 dB 0.005 4 dB 0.001 1 dB 0.007 2 dB 0.021 dB 0.015 dB 0.027 dB	
Attenuation		(10 Hz ~ 100 kHz) (0 ~ 80) dB	 0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) % (1 ~ 100) kHz (3.16 ~ 0.010) %	 1.5×10^{-1} 3.2×10^{-1}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411			Audio Analyzer, Digital Multimeter/ SICT-CP-40411
Rise Fall Time		100 μ s ~ 100 ns	7.0×10^{-4}	
		(100 ~ 10) ns	7.8×10^{-4}	
		(10 ~ 1) ns	4.7×10^{-3}	
		1 ns ~ 100 ps	4.6×10^{-2}	
Duty cycle		(1 ~ 99) %	0.006 1 %	
High Frequency Flatness Test	40412	(100 kHz ~ 80 MHz)		Signal Generator/ SICT-CP-40412
		(0 ~ 20) dBm	0.11 dB	
FM Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
AM Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Genescopes	40413			Calibrator/ SICT-CP-40413
Marker Frequency		9 kHz ~ 10 MHz	2.8×10^{-6}	
		(10 ~ 200) MHz	6.4×10^{-7}	
RF Level	40413	9 kHz ~ 200 MHz		
		(100 ~ 50) dB μ V	0.31 dB	
AC/DC high voltages volt meters				
DC Voltage		($\pm$)		
		0 kV	0.58 V	
		(0 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.1×10^{-4}	
		(1 ~ 2) kV	4.4×10^{-4}	
		(2 ~ 100) kV	3.4×10^{-4}	
AC Voltage		(50 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	
		(1 ~ 2) kV	5.5×10^{-4}	
		(2 ~ 3) kV	5.3×10^{-4}	
		(3 ~ 15) kV	5.0×10^{-4}	
		(15 ~ 100) kV	5.7×10^{-4}	
		(60 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	
		(1 ~ 2) kV	5.5×10^{-4}	
		(2 ~ 3) kV	4.7×10^{-4}	
		(3 ~ 15) kV	4.5×10^{-4}	
		(15 ~ 100) kV	5.4×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Jitter meters	40415			Modulation Domain Analyzer/ SICT-CP-40415
CD/DVD Jitter		(1 ~ 20) ns (20 ~ 60) ns	1.7×10^{-3} 1.6×10^{-3}	
VTR Jitter		0.05 μ s (0.05 ~ 0.1) μ s (0.1 ~ 0.2) μ s (0.2 ~ 0.5) μ s (0.5 ~ 0.7) μ s	0.66 ns 0.77 ns 1.2 ns 2.8 ns 4.3 ns	
		1 % 2 % 4 % 8 % 10 % 15 %	0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	
Leakage current testers	40416			Calibrator/ SICT-CP-40416
DC Current		0 μ A (0 ~ 1) μ A (1 ~ 2) μ A (2 ~ 5) μ A (5 ~ 10) μ A (10 ~ 20) μ A (20 ~ 50) μ A (50 ~ 100) μ A (100 ~ 200) μ A (0.2 ~ 100) mA	7.0 nA 2.4×10^{-3} 3.6×10^{-3} 1.4×10^{-3} 7.4×10^{-4} 4.0×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 8.5×10^{-5} 6.1×10^{-4}	
AC Current		(20 μ A) 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	26 nA 8.5×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 5.5×10^{-3}	
		(20 ~ 50) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 6.8×10^{-4} 2.8×10^{-3}	
		(50 ~ 100) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	4.9×10^{-4} 3.2×10^{-4} 2.3×10^{-4} 4.9×10^{-4} 4.0×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	(100 ~ 200) μ A		Calibrator/ SICT-CP-40416
AC Current		10 Hz	3.9×10^{-4}	
		(10 ~ 20) Hz	2.5×10^{-4}	
		(0.02 ~ 1) kHz	1.7×10^{-4}	
		(1 ~ 5) kHz	4.0×10^{-4}	
		(5 ~ 10) kHz	1.7×10^{-3}	
		(200 ~ 500) μ A		
		10 Hz	4.4×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		(0.02 ~ 1) kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		(0.5 ~ 1) mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		(0.02 ~ 1) kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 100) mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		(0.02 ~ 1) kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	2.7×10^{-3}	
DC Voltage		0 V	0.06 mV	
		(0 ~ 0.1) V	6.0×10^{-4}	
		(0.1 ~ 0.2) V	3.0×10^{-4}	
		(0.2 ~ 0.5) V	1.2×10^{-4}	
		(0.5 ~ 1) V	6.0×10^{-5}	
		(1 ~ 2) V	3.1×10^{-4}	
		(2 ~ 5) V	1.2×10^{-4}	
		(5 ~ 10) V	6.1×10^{-5}	
		(10 ~ 20) V	3.1×10^{-5}	
		(20 ~ 50) V	1.4×10^{-5}	
		(50 ~ 100) V	8.8×10^{-6}	
		(100 ~ 200) V	3.1×10^{-5}	
		(200 ~ 300) V	2.3×10^{-5}	
		(300 ~ 500) V	1.2×10^{-4}	
		(500 ~ 1 000) V	5.8×10^{-5}	
AC Voltage		0.1 V		
		10 Hz	0.074 mV	
		(0.01 ~ 50) kHz	6.5×10^{-4}	
		(50 ~ 100) kHz	8.3×10^{-4}	
		(100 ~ 300) kHz	1.2×10^{-3}	
		(300 ~ 500) kHz	2.0×10^{-3}	
		(0.5 ~ 1) MHz	3.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	(0.1 ~ 0.2) V		Calibrator/ SICT-CP-40416
AC Voltage		10 Hz	4.5×10^{-4}	
		(0.01 ~ 50) kHz	3.5×10^{-4}	
		(50 ~ 100) kHz	5.5×10^{-4}	
		(100 ~ 300) kHz	9.5×10^{-4}	
		(300 ~ 500) kHz	1.8×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(0.2 ~ 0.5) V		
		10 Hz	4.0×10^{-4}	
		(10 ~ 20) Hz	2.0×10^{-4}	
		(0.02 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 300) kHz	6.0×10^{-4}	
		(300 ~ 500) kHz	1.7×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	3.3×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	9.1×10^{-5}	
		(20 ~ 100) kHz	1.5×10^{-4}	
		(100 ~ 300) kHz	4.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(1 ~ 2) V		
		10 Hz	4.3×10^{-4}	
		(0.01 ~ 100) kHz	3.3×10^{-4}	
		(100 ~ 300) kHz	5.5×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.1×10^{-3}	
		(2 ~ 5) V		
		10 Hz	4.0×10^{-4}	
		(0.01 ~ 100) kHz	2.0×10^{-4}	
		(100 ~ 300) kHz	5.0×10^{-4}	
		(300 ~ 500) kHz	1.7×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	(5 ~ 10) V		Calibrator/ SICT-CP-40416
AC Voltage		10 Hz	3.3×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	8.9×10^{-5}	
		(20 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	3.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(10 ~ 20) V		
		10 Hz	3.1×10^{-4}	
		(10 ~ 20) Hz	1.2×10^{-4}	
		(0.02 ~ 20) kHz	6.0×10^{-5}	
		(20 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(20 ~ 50) V		
		10 Hz	4.2×10^{-4}	
		(10 ~ 20) Hz	2.2×10^{-4}	
		(0.02 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	2.8×10^{-4}	
		(50 ~ 100) V		
		10 Hz	3.4×10^{-4}	
		(0.01 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 1 000) V		
		(0.05 ~ 1) kHz	1.1×10^{-4}	
Resistance		100 mΩ	7.7 μΩ	
		1 Ω ~ 10 kΩ	6.2×10^{-5}	
Input Voltage to Output Current Display(U1)		20 Hz		
		(4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz		
		(4.85 ~ 5.36) mA	0.006 1 mA	
		200 Hz		
		(5.11 ~ 5.65) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	500 Hz (6.64 ~ 7.34) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
Input Voltage to Output Current Display(U1)		1 kHz (9.70 ~ 10.73) mA	0.006 1 mA	
		2 kHz (14.07 ~ 15.56) mA	0.006 2 mA	
		5 kHz (17.82 ~ 19.70) mA	0.006 2 mA	
		10 kHz (18.66 ~ 20.63) mA	0.006 2 mA	
		20 kHz (18.92 ~ 20.92) mA	0.006 2 mA	
		50 kHz (19.00 ~ 21.00) mA	0.006 4 mA	
		100 kHz (19.00 ~ 21.00) mA	0.006 7 mA	
		200 kHz (19.00 ~ 21.00) mA	0.010 mA	
		500 kHz (19.00 ~ 21.00) mA	0.030 mA	
		1 MHz (19.00 ~ 21.00) mA	0.046 mA	
Input Voltage to Output Current Display(U2)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.92 ~ 5.44) mA	0.006 1 mA	
		500 Hz (5.37 ~ 5.93) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
Input Voltage to Output Current Display(U2)		1 kHz (5.56 ~ 6.14) mA	0.006 1 mA	
		2 kHz (4.68 ~ 5.17) mA	0.006 1 mA	
		5 kHz (2.53 ~ 2.80) mA	0.000 63 mA	
		10 kHz (1.35 ~ 1.49) mA	0.000 62 mA	
		20 kHz (0.683 ~ 0.755) mA	0.000 61 mA	
		50 kHz (274.57 ~ 303.47) mA	0.029 μ A	
		100 kHz (137.48 ~ 151.95) μ A	0.020 μ A	
		200 kHz (68.82 ~ 76.06) μ A	0.030 μ A	
		500 kHz (27.43 ~ 30.32) μ A	0.042 μ A	
		1 MHz (13.71 ~ 15.16) μ A	0.033 μ A	
Input Voltage to Output Current Display(U3)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.95 ~ 5.47) mA	0.006 1 mA	
		500 Hz (5.65 ~ 6.25) mA	0.006 1 mA	
		1 kHz (6.60 ~ 7.29) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
Input Voltage to Output Current Display(U3)		2 kHz (7.14 ~ 7.89) mA	0.006 1 mA	
		5 kHz (5.31 ~ 5.87) mA	0.006 1 mA	
		10 kHz (3.12 ~ 3.45) mA	0.000 64 mA	
		20 kHz (1.63 ~ 1.81) mA	0.000 62 mA	
		50 kHz (0.664 ~ 0.734) mA	0.000 62 mA	
		100 kHz (322.16 ~ 367.12) μ A	0.046 μ A	
		200 kHz (166.03 ~ 183.81) μ A	0.070 μ A	
		500 kHz (66.37 ~ 73.35) μ A	0.10 μ A	
		1 MHz (33.14 ~ 36.63) μ A	0.08 μ A	
Input Voltage to Output Voltage Ratio(U1)		4.00 (20 Hz)	1.3×10^{-4}	
		3.98 (50 Hz)	6.5×10^{-5}	
		3.97 (60 Hz)	6.5×10^{-5}	
		3.92 (100 Hz)	6.5×10^{-5}	
		3.72 (200 Hz)	6.5×10^{-5}	
		2.87 (500 Hz)	6.4×10^{-5}	
		1.96 (1 kHz)	6.4×10^{-5}	
		1.96 (2 kHz)	6.4×10^{-5}	
		1.96 (5 kHz)	6.4×10^{-5}	
		1.96 (10 kHz)	6.4×10^{-5}	
		1.00 (20 kHz)	6.7×10^{-5}	
		1.00 (50 kHz)	9.6×10^{-5}	
		1.00 (100 kHz)	1.2×10^{-4}	
		1.00 (200 kHz)	4.2×10^{-4}	
		1.00 (500 kHz)	1.5×10^{-3}	
		1.00 (1 MHz)	2.6×10^{-3}	

404. Other DC & LF measurements

404. Other DC & LF Measurements				
Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
Input Voltage to Output Voltage Ratio(U2)		4.00 (20 Hz)	1.3×10^{-4}	
		3.99 (50 Hz)	6.5×10^{-5}	
		3.99 (60 Hz)	6.5×10^{-5}	
		3.96 (100 Hz)	6.5×10^{-5}	
		3.87 (200 Hz)	6.5×10^{-5}	
		3.54 (500 Hz)	6.5×10^{-5}	
		3.43 (1 kHz)	6.5×10^{-5}	
		4.06 (2 kHz)	6.5×10^{-5}	
		7.50 (5 kHz)	6.6×10^{-5}	
		14.1 (10 kHz)	7.0×10^{-5}	
		27.8 (20 kHz)	3.6×10^{-5}	
		69.2 (50 kHz)	4.8×10^{-5}	
		138 (100 kHz)	1.0×10^{-4}	
		272 (200 kHz)	2.7×10^{-4}	
		691 (500 kHz)	1.1×10^{-3}	
		1 382 (1 MHz)	3.0×10^{-3}	
Input Voltage to Output Voltage Ratio(U3)		4.00 (20 Hz)	1.3×10^{-4}	
		3.99 (50 Hz)	6.5×10^{-5}	
		3.98 (60 Hz)	6.5×10^{-5}	
		3.95 (100 Hz)	6.6×10^{-5}	
		3.83 (200 Hz)	6.5×10^{-5}	
		2.36 (500 Hz)	6.5×10^{-5}	
		2.87 (1 kHz)	6.4×10^{-5}	
		2.65 (2 kHz)	6.4×10^{-5}	
		3.57 (5 kHz)	6.5×10^{-5}	
		6.09 (10 kHz)	6.5×10^{-5}	
		11.6 (20 kHz)	6.9×10^{-5}	
		28.7 (50 kHz)	3.8×10^{-5}	
		57.2 (100 kHz)	7.9×10^{-5}	
		114 (200 kHz)	1.9×10^{-4}	
		286 (500 kHz)	6.1×10^{-4}	
		572 (1 MHz)	2.3×10^{-3}	
mAs Meter		1 mAs	1.2×10^{-3}	
		(1 ~ 2 000) mAs	1.0×10^{-3}	
		(2 000 ~ 9 999) mAs	1.1×10^{-3}	
Electronic AC/DC loads	40417			Calibrator/ SICT-CP-40417
DC Voltage		0 mV	0.058 mV	
		(0 ~ 5) mV	5.8×10^{-2}	
		(5 ~ 20) mV	5.8×10^{-3}	
		(20 ~ 100) mV	1.2×10^{-3}	
		(0.1 ~ 1) V	6.2×10^{-5}	
		(1 ~ 2) V	3.2×10^{-5}	
		(2 ~ 4) V	2.1×10^{-5}	
		(4 ~ 7) V	1.3×10^{-5}	
		(7 ~ 9) V	9.1×10^{-6}	
		(9 ~ 10) V	7.9×10^{-6}	
		(10 ~ 50) V	3.1×10^{-5}	
		(50 ~ 100) V	1.0×10^{-5}	
		(100 ~ 200) V	3.4×10^{-5}	
		(200 ~ 400) V	2.5×10^{-5}	
		(400 ~ 1 000) V	1.6×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			Calibrator/ SICT-CP-40417
DC Current		1 mA	5.8 μ A	
		(1 ~ 2) mA	2.9×10^{-3}	
		(2 ~ 5) mA	1.2×10^{-3}	
		(5 ~ 20) mA	5.8×10^{-4}	
		(20 ~ 50) mA	1.2×10^{-4}	
		(50 ~ 100) mA	5.9×10^{-5}	
		(0.1 ~ 0.2) A	2.9×10^{-4}	
		(0.2 ~ 0.4) A	1.9×10^{-4}	
		(0.4 ~ 0.6) A	1.2×10^{-4}	
		(0.6 ~ 0.8) A	8.4×10^{-5}	
		(0.8 ~ 1) A	6.6×10^{-5}	
		(1 ~ 3) A	5.1×10^{-5}	
		(3 ~ 6) A	2.6×10^{-5}	
		(6 ~ 10) A	1.6×10^{-5}	
		(10 ~ 40) A	4.0×10^{-5}	
		(40 ~ 100) A	2.9×10^{-5}	
		(100 ~ 1 000) A	1.4×10^{-4}	
		(1 000 ~ 2 000) A	4.9×10^{-4}	
Charge voltage		0 mV	0.058 mV	
		(0 ~ 5) mV	5.8×10^{-2}	
		(5 ~ 20) mV	5.8×10^{-3}	
		(20 ~ 100) mV	1.2×10^{-3}	
		(0.1 ~ 1) V	6.2×10^{-5}	
		(1 ~ 2) V	3.2×10^{-5}	
		(2 ~ 4) V	2.1×10^{-5}	
		(4 ~ 7) V	1.3×10^{-5}	
		(7 ~ 9) V	9.1×10^{-6}	
		(9 ~ 10) V	7.9×10^{-6}	
		(10 ~ 50) V	3.1×10^{-5}	
		(50 ~ 100) V	1.0×10^{-5}	
		(100 ~ 200) V	3.4×10^{-5}	
		(200 ~ 400) V	2.5×10^{-5}	
		(400 ~ 1 000) V	1.6×10^{-5}	
		(1 000 ~ 1 200) V	1.0×10^{-3}	
		(1 200 ~ 1 400) V	9.2×10^{-4}	
		(1 400 ~ 1 500) V	8.7×10^{-4}	
Charge and Discharge Current		($\pm$)		
		1 mA	5.8 μ A	
		(1 ~ 2) mA	2.9×10^{-3}	
		(2 ~ 5) mA	1.2×10^{-3}	
		(5 ~ 20) mA	5.8×10^{-4}	
		(20 ~ 50) mA	1.2×10^{-4}	
		(50 ~ 100) mA	5.9×10^{-5}	
		(0.1 ~ 0.2) A	2.9×10^{-4}	
		(0.2 ~ 0.4) A	1.9×10^{-4}	
		(0.4 ~ 0.6) A	1.2×10^{-4}	
		(0.6 ~ 0.8) A	8.4×10^{-5}	
		(0.8 ~ 1) A	6.6×10^{-5}	
		(1 ~ 3) A	5.1×10^{-5}	
		(3 ~ 6) A	2.6×10^{-5}	
		(6 ~ 10) A	1.6×10^{-5}	
		(10 ~ 40) A	4.0×10^{-5}	
		(40 ~ 100) A	2.9×10^{-5}	
		(100 ~ 1 000) A	1.4×10^{-4}	
		(1 000 ~ 3 000) A	4.9×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			Calibrator/ SICT-CP-40417
Resistance		0.1 Ω (0.1 ~ 1) Ω (1 ~ 2) Ω (2 ~ 4) Ω (4 ~ 500) Ω (0.5 ~ 2) k Ω (2 ~ 10) k Ω	0.58 m Ω 2.1×10^{-3} 4.0×10^{-3} 2.3×10^{-3} 1.6×10^{-3} 9.0×10^{-4} 2.0×10^{-4}	
AC Voltage		(0.001 V) (40 ~ 400) Hz (0.001 ~ 0.1) V (40 ~ 400) Hz (0.1 ~ 0.2) V (40 ~ 400) Hz (0.2 ~ 0.5) V (40 ~ 400) Hz (0.5 ~ 2) V (40 ~ 400) Hz (2 ~ 3) V (40 ~ 400) Hz (3 ~ 7) V (40 ~ 400) Hz (7 ~ 20) V (40 ~ 50) Hz (50 ~ 400) Hz (20 ~ 80) V (40 ~ 50) Hz (50 ~ 400) Hz (80 ~ 200) V (40 ~ 400) Hz (200 ~ 500) V (50 ~ 400) Hz	0.61 mV 6.1×10^{-3} 3.1×10^{-3} 1.2×10^{-3} 6.2×10^{-4} 2.8×10^{-4} 2.3×10^{-4} 1.5×10^{-4} 9.8×10^{-5} 2.1×10^{-4} 1.2×10^{-4} 1.3×10^{-4} 1.8×10^{-4}	
AC Current		(1 mA) (40 ~ 400) Hz (1 ~ 100) mA (40 ~ 400) Hz (100 mA ~ 0.2 A) (40 ~ 400) Hz	0.58 mA 5.8×10^{-2} 5.8×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads AC Current	40417	(0.2 ~ 0.6) A (40 ~ 400) Hz	2.0×10^{-3}	Calibrator/ SICT-CP-40417
		(0.6 ~ 2) A (40 ~ 400) Hz	9.4×10^{-4}	
		(2 ~ 5) A (40 ~ 400) Hz	1.2×10^{-3}	
		(5 ~ 20) A) (40 ~ 400) Hz	1.0×10^{-3}	
AC Resistance		(1 Ω) (40 ~ 400) Hz	1.0 m Ω	
		(1 ~ 50) Ω (40 ~ 400) Hz	1.5×10^{-3}	
		(50 ~ 100) Ω (40 ~ 400) Hz	1.1×10^{-3}	
		(100 Ω ~ 10 k Ω) (40 ~ 400) Hz	1.7×10^{-3}	
Modulation meters Frequency Modulation	40418	0 kHz (0 ~ 400) kHz	1 Hz 1.2×10^{-2}	Measuring Receiver/ SICT-CP-40418
Amplitude Modulation		0 % (0 ~ 100) %	0.01 % 1.2×10^{-2}	
Phase Modulation		0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}	
Analogue/Digital multimeters DC Voltage	40419	($\pm$) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 15) mV (15 ~ 20) mV (20 ~ 50) mV (0.05 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	0.43 μ V 5.0×10^{-4} 2.5×10^{-4} 1.0×10^{-4} 5.0×10^{-5} 3.3×10^{-5} 2.5×10^{-6} 1.2×10^{-5} 8.0×10^{-6} 4.8×10^{-6} 3.8×10^{-6} 4.0×10^{-6} 2.6×10^{-6} 2.3×10^{-6} 6.0×10^{-6} 4.0×10^{-6} 3.5×10^{-6} 8.0×10^{-6} 5.2×10^{-6} 4.5×10^{-6}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(0.6 mV) 1 kHz (1 mV) 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 2) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 5) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (5 ~ 10) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 15) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	4.1 μ V 4.2 μ V 4.2 μ V 4.1 μ V 4.2 μ V 5.5 μ V 11 μ V 21 μ V 23 μ V 2.2×10^{-3} 2.2×10^{-3} 2.1×10^{-3} 2.2×10^{-3} 3.0×10^{-3} 6.0×10^{-3} 1.1×10^{-2} 1.3×10^{-2} 1.1×10^{-3} 9.2×10^{-4} 9.0×10^{-4} 1.0×10^{-3} 1.5×10^{-3} 3.0×10^{-3} 5.2×10^{-3} 6.8×10^{-3} 6.3×10^{-4} 5.0×10^{-4} 4.9×10^{-4} 5.9×10^{-4} 9.5×10^{-4} 1.9×10^{-3} 3.2×10^{-3} 4.3×10^{-3} 4.8×10^{-4} 3.6×10^{-4} 3.5×10^{-4} 4.5×10^{-4} 7.6×10^{-4} 1.5×10^{-3} 2.5×10^{-3} 3.7×10^{-3}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(15 ~ 20) mV		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	2.9×10^{-4}	
		(0.04 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 50) kHz	3.8×10^{-4}	
		(50 ~ 100) kHz	6.7×10^{-4}	
		(100 ~ 300) kHz	1.4×10^{-3}	
		(300 ~ 500) kHz	2.2×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.6×10^{-4}	
		(0.04 ~ 20) kHz	2.1×10^{-4}	
		(20 ~ 50) kHz	2.7×10^{-4}	
		(50 ~ 100) kHz	6.6×10^{-4}	
		(100 ~ 300) kHz	1.0×10^{-3}	
		(300 ~ 500) kHz	1.6×10^{-3}	
		(0.5 ~ 1) MHz	3.3×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	3.3×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		(0.04 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.9×10^{-4}	
		(50 ~ 100) kHz	4.8×10^{-4}	
		(100 ~ 300) kHz	7.6×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(100 ~ 150) mV		
		10 Hz	2.9×10^{-4}	
		(10 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 20) kHz	1.0×10^{-4}	
		(20 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	4.1×10^{-4}	
		(100 ~ 300) kHz	6.8×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	
		(150 ~ 200) mV		
		10 Hz	2.7×10^{-4}	
		(10 ~ 40) Hz	1.2×10^{-4}	
		(0.04 ~ 20) kHz	9.2×10^{-5}	
		(20 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	3.9×10^{-4}	
		(100 ~ 300) kHz	6.4×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(0.2 ~ 0.5) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (0.5 ~ 1) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 2) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 5) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (5 ~ 10) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	 3.0×10^{-4} 1.3×10^{-4} 9.0×10^{-5} 6.6×10^{-5} 1.0×10^{-4} 1.6×10^{-4} 4.6×10^{-4} 1.2×10^{-3} 2.7×10^{-3} 2.5×10^{-4} 1.0×10^{-4} 5.8×10^{-5} 4.9×10^{-5} 7.7×10^{-5} 1.2×10^{-4} 3.6×10^{-4} 1.0×10^{-3} 2.5×10^{-3} 2.2×10^{-4} 8.5×10^{-5} 4.5×10^{-5} 4.2×10^{-5} 6.8×10^{-5} 9.6×10^{-5} 3.2×10^{-4} 9.0×10^{-4} 2.4×10^{-3} 3.0×10^{-4} 1.3×10^{-4} 8.2×10^{-5} 6.2×10^{-5} 1.0×10^{-4} 1.4×10^{-4} 4.4×10^{-4} 1.3×10^{-3} 2.2×10^{-3} 2.5×10^{-4} 9.8×10^{-5} 5.4×10^{-5} 4.7×10^{-5} 7.7×10^{-5} 1.1×10^{-4} 3.2×10^{-4} 1.0×10^{-3} 1.7×10^{-3}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(10 ~ 15) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (15 ~ 20) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (20 ~ 50) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (50 ~ 100) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (200 ~ 500) V 50 Hz ~ 1 kHz (500 ~ 1 000) V 50 Hz ~ 1 kHz	 2.3×10^{-4} 8.7×10^{-5} 4.6×10^{-5} 4.3×10^{-5} 7.1×10^{-5} 9.5×10^{-5} 2.9×10^{-4} 9.4×10^{-4} 1.5×10^{-3} 2.2×10^{-4} 8.5×10^{-5} 4.3×10^{-5} 4.1×10^{-5} 6.8×10^{-5} 9.0×10^{-5} 2.8×10^{-4} 9.1×10^{-4} 1.4×10^{-3} 3.2×10^{-4} 1.4×10^{-4} 9.8×10^{-5} 7.4×10^{-5} 1.1×10^{-5} 2.1×10^{-4} 2.5×10^{-4} 1.0×10^{-4} 6.5×10^{-5} 5.6×10^{-5} 8.5×10^{-5} 1.6×10^{-4} 2.3×10^{-4} 9.3×10^{-5} 5.6×10^{-5} 5.1×10^{-5} 7.9×10^{-5} 1.4×10^{-4} 6.7×10^{-5} 6.3×10^{-5}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419			Calibrator/ SICT-CP-40419
Resistance		$0 \ \Omega$ $0 \ \Omega \sim 10 \ k\Omega$ $(10 \sim 100) \ k\Omega$ $(0.1 \sim 1) \ M\Omega$ $(1 \sim 10) \ M\Omega$ $(10 \sim 100) \ M\Omega$ $(0.1 \sim 1) \ G\Omega$	$0.001 \ 0 \ m\Omega$ 1.2×10^{-6} 1.4×10^{-6} 7.2×10^{-6} 7.7×10^{-6} 1.2×10^{-5} 3.2×10^{-4}	
DC Current		$(\pm)$ $0 \ nA$ $(0 \sim 1) \ nA$ $(1 \sim 100) \ nA$ $(0.1 \sim 1) \ \mu A$ $(1 \sim 2) \ \mu A$ $(2 \sim 5) \ \mu A$ $(5 \sim 10) \ \mu A$ $(10 \sim 20) \ \mu A$ $(20 \sim 50) \ \mu A$ $(50 \sim 100) \ \mu A$ $(100 \sim 200) \ \mu A$ $(0.2 \sim 0.5) \ mA$ $(0.5 \sim 1) \ mA$ $(1 \sim 1.5) \ mA$ $(1.5 \sim 2) \ mA$ $(2 \sim 5) \ mA$ $(5 \sim 10) \ mA$ $(10 \sim 15) \ mA$ $(15 \sim 20) \ mA$ $(20 \sim 50) \ mA$ $(50 \sim 100) \ mA$ $(100 \sim 150) \ mA$ $(150 \sim 200) \ mA$ $(0.2 \sim 0.5) \ A$ $(0.5 \sim 1) \ A$ $(1 \sim 1.5) \ A$ $(1.5 \sim 2) \ A$ $(2 \sim 3) \ A$ $(3 \sim 5) \ A$ $(5 \sim 10) \ A$ $(10 \sim 20) \ A$ $(20 \sim 30) \ A$	$6.0 \ nA$ 6.9×10^{-3} 4.6×10^{-3} 6.0×10^{-3} 3.1×10^{-3} 1.2×10^{-3} 6.3×10^{-4} 3.5×10^{-4} 1.6×10^{-4} 9.4×10^{-5} 6.3×10^{-5} 4.8×10^{-5} 3.5×10^{-5} 3.1×10^{-5} 3.0×10^{-5} 4.4×10^{-5} 3.2×10^{-5} 2.9×10^{-5} 2.8×10^{-5} 5.6×10^{-5} 4.4×10^{-5} 4.1×10^{-5} 3.9×10^{-5} 9.4×10^{-5} 6.9×10^{-5} 6.1×10^{-5} 5.8×10^{-5} 3.3×10^{-4} 2.4×10^{-4} 1.6×10^{-4} 1.2×10^{-4} 2.4×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Current	40419	(20 μ A) 1 kHz 10 kHz (20 ~ 50) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.1 ~ 0.2) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.2 ~ 0.5) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.5 ~ 1) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 2) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 5) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	11 nA 81 nA 1.4×10^{-4} 9.0×10^{-5} 7.3×10^{-5} 1.5×10^{-4} 5.5×10^{-4} 1.9×10^{-4} 1.2×10^{-4} 9.5×10^{-5} 2.0×10^{-4} 7.5×10^{-4} 3.0×10^{-4} 1.9×10^{-4} 1.4×10^{-5} 3.1×10^{-4} 1.2×10^{-3} 9.0×10^{-5} 7.0×10^{-5} 6.0×10^{-5} 1.2×10^{-4} 5.7×10^{-4} 1.4×10^{-4} 1.0×10^{-4} 8.0×10^{-5} 1.6×10^{-4} 7.6×10^{-4} 2.4×10^{-4} 1.6×10^{-4} 1.2×10^{-4} 2.4×10^{-4} 1.2×10^{-3} 9.0×10^{-5} 7.0×10^{-5} 5.2×10^{-5} 1.1×10^{-4} 5.4×10^{-4}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

[illegible]

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Current		(3 ~ 5) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 30) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz	 2.4×10^{-4} 2.5×10^{-4} 1.6×10^{-3} 4.2×10^{-4} 4.2×10^{-4} 3.1×10^{-3} 6.0×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 2.3×10^{-3} 8.0×10^{-4} 1.2×10^{-3} 3.9×10^{-3} 4.5×10^{-3}	
Frequency		10 Hz ~ 10 MHz	6.4×10^{-7}	
(Digital Smapling) AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz (1 mV ~ 2 mV) 0.1 Hz ~ 3 kHz (2 mV ~ 3 mV) 0.1 Hz ~ 3 kHz (3 mV ~ 5 mV) 0.1 Hz ~ 3 kHz (5 mV ~ 10 mV) 0.1 Hz ~ 3 kHz (10 mV ~ 20 mV) 0.1 Hz ~ 3 kHz (20 mV ~ 30 mV) 0.1 Hz ~ 3 kHz (30 mV ~ 50 mV) 0.1 Hz ~ 3 kHz (50 mV ~ 100 mV) 0.1 Hz ~ 3 kHz	 8.4×10^{-4} 4.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 8.8×10^{-5} 4.8×10^{-5} 3.6×10^{-5} 3.0×10^{-5} 2.6×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters (Digital Sampling) AC Voltage	40419	(100 mV ~ 200 mV) 0.1 Hz ~ 3 kHz (200 mV ~ 300 mV) 0.1 Hz ~ 3 kHz (300 mV ~ 500 mV) 0.1 Hz ~ 3 kHz (500 mV ~ 1 V) 0.1 Hz ~ 3 kHz (1 V ~ 2 V) 0.1 Hz ~ 3 kHz (2 V ~ 3 V) 0.1 Hz ~ 3 kHz (3 V ~ 5 V) 0.1 Hz ~ 3 kHz (5 V ~ 10 V) 0.1 Hz ~ 3 kHz (10 V ~ 30 V) 10 Hz ~ 3 kHz (30 V ~ 50 V) 10 Hz ~ 3 kHz (50 V ~ 100 V) 10 Hz ~ 3 kHz (100 V ~ 200 V) 10 Hz ~ 3 kHz (200 V ~ 1 000 V) 50 Hz ~ 1 kHz	 4.8×10^{-5} 3.6×10^{-5} 2.8×10^{-5} 2.4×10^{-5} 4.8×10^{-5} 3.6×10^{-5} 3.0×10^{-5} 2.6×10^{-5} 3.6×10^{-5} 2.8×10^{-5} 2.4×10^{-5} 4.8×10^{-5} 2.4×10^{-5}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420	(600 μ V) 1 kHz	7.8×10^{-3}	Calibrator/ SICT-CP-40420
AC Voltage Test		(600 μ V ~ 20 mV) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	2.6×10^{-3} 2.5×10^{-3} 3.5×10^{-3} 7.0×10^{-3} 1.5×10^{-2}	
		(20 ~ 200) mV (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	5.2×10^{-4} 3.4×10^{-4} 8.8×10^{-4} 4.3×10^{-3}	
		(200 mV ~ 2 V) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.5×10^{-4} 1.5×10^{-4} 4.6×10^{-4} 8.8×10^{-4} 3.4×10^{-3}	
		(2 ~ 20) V (10 ~ 40) Hz 40 Hz ~ 100 kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.0×10^{-4} 1.2×10^{-4} 4.4×10^{-4} 2.2×10^{-3}	
		(20 ~ 200) V (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz	3.0×10^{-4} 1.2×10^{-4} 1.8×10^{-4}	
		(200 ~ 500) V 50 Hz ~ 1 kHz	3.8×10^{-4}	
		(500 ~ 1 000) V 50 Hz ~ 1 kHz	3.7×10^{-4}	
		(25 ~ 500) mV (1 ~ 30) MHz	2.1×10^{-2}	
		(500 mV ~ 2 V) (0.1 ~ 30) MHz	2.1×10^{-2}	
Weighting Test		(DIN/NOISE) 31.5 Hz ~ 10 kHz (JIS A) 31.5 Hz ~ 16 kHz (CCIR) 31.5 Hz ~ 31.5 kHz (CCIR/ARM) 31.5 Hz ~ 31.5 kHz	0.12 dB 0.12 dB 0.12 dB 0.12 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420			Calibrator/ SICT-CP-40420
AC Voltage Output		(10 mV) 1 kHz	2.8×10^{-3}	
		(10 mV ~ 1 V) 20 Hz ~ 1 kHz	8.5×10^{-5}	
		(1 ~ 10) kHz	1.3×10^{-4}	
		(10 ~ 100) kHz	6.9×10^{-4}	
DC Voltage Output		0 mV 100 mV ~ 1 V	0.99 μ V 1.1×10^{-5}	
Oscilloscopes	40421			Calibration Generator/ SICT-CP-40421
Impedance Measure		50 Ω 75 Ω 1 M Ω	3.5×10^{-5} 2.7×10^{-5} 2.5×10^{-5}	
DC Voltage		($\pm$) 0 mV (0 ~ 1) mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 900) mV (0.9 ~ 9) V (9 ~ 200) V	0.79 μ V 8.0×10^{-4} 4.1×10^{-4} 1.7×10^{-4} 8.5×10^{-5} 1.5×10^{-5} 9.1×10^{-6} 9.5×10^{-6}	
AC Voltage(Square wave)		(1 kHz) 1 mV (1 ~ 25) mV (0.025 ~ 0.5) V (0.5 ~ 2.2) V (2.2 ~ 130) V	6.5×10^{-3} 8.8×10^{-4} 9.1×10^{-4} 6.8×10^{-4} 8.4×10^{-4}	
Time Marker		100 ps (100 ~ 200) ps 200 ps ~ 20 ms 20 ms ~ 5 s	6.2×10^{-7} 3.1×10^{-7} 1.7×10^{-6} 1.6×10^{-5}	
CAL Output Amplitude		(40 Hz ~ 20 kHz) 100 mV 100 mV ~ 12 V	3.2×10^{-5} 1.9×10^{-5}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.2×10^{-7}	
Sinewave Signal Generator Level		50 kHz 50 kHz ~ 1 MHz 1 MHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 18) GHz (18 ~ 25) GHz (25 ~ 33) GHz (33 ~ 40) GHz	2.3×10^{-2} 4.7×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 3.2×10^{-2} 5.5×10^{-2} 5.8×10^{-2} 6.0×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes AC Voltage(Sine wave)	40421	(10 ~ 40) Hz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V (40 Hz ~ 1 kHz) 2 mV (2 ~ 20) mV (20 ~ 800) mV (0.8 ~ 20) V (20 ~ 200) V (1 ~ 50) kHz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V (50 ~ 100) kHz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V	5.3 μ V 1.9 $\times 10^{-3}$ 6.8 $\times 10^{-4}$ 4.3 $\times 10^{-4}$ 4.5 $\times 10^{-4}$ 5.0 μ V 1.7 $\times 10^{-3}$ 3.3 $\times 10^{-4}$ 1.7 $\times 10^{-4}$ 1.8 $\times 10^{-4}$ 5.2 μ V 1.9 $\times 10^{-3}$ 4.8 $\times 10^{-4}$ 1.3 $\times 10^{-4}$ 1.7 $\times 10^{-4}$ 7.1 μ V 2.7 $\times 10^{-3}$ 9.0 $\times 10^{-4}$ 1.9 $\times 10^{-4}$ 3.0 $\times 10^{-4}$	Calibration Generator/ SICT-CP-40421
LF phase meters Phase Test	40422	(1 Hz ~ 200 kHz) (-180 ~ 180) °	0.074 °	Multi Function Generator/ SICT-CP-40422
Volt/Current recorders DC Voltage DC Current	40424	(±) (0 ~ 100) μ V (0.1 ~ 1) mV (1 ~ 10) mV (0.01 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V (±) (0 ~ 1) nA (1 ~ 100) nA (0.1 ~ 1) μ A (1 ~ 10) μ A (10 ~ 100) μ A (0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 100) A	0.51 μ V 5.2 $\times 10^{-4}$ 5.9 $\times 10^{-5}$ 6.7 $\times 10^{-6}$ 4.3 $\times 10^{-6}$ 6.3 $\times 10^{-6}$ 8.7 $\times 10^{-6}$ 6.9 pA 4.6 $\times 10^{-3}$ 2.3 $\times 10^{-3}$ 7.2 $\times 10^{-4}$ 1.4 $\times 10^{-4}$ 7.6 $\times 10^{-5}$ 8.4 $\times 10^{-5}$ 1.2 $\times 10^{-4}$ 2.1 $\times 10^{-4}$	Calibrator/ SICT-CP-40424

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425			Digital Multimeter/ SICT-CP-40425
DC Voltage		1 mV 1 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	6 μ V 7.0×10^{-4} 7.0×10^{-5} 2.2×10^{-5}	
DC Current		1 mA 1 mA ~ 1 A (1 ~ 20) A (20 ~ 100) A	58 μ A 6.0×10^{-4} 2.5×10^{-4} 4.0×10^{-4}	
AC Voltage		(1 mV) 20 Hz ~ 100 kHz (1 ~ 100) mV 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) V 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 100) V 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 1 000) V 50 Hz ~ 10 kHz (10 ~ 30) kHz	58 μ V 6.1×10^{-4} 1.2×10^{-3} 7.6×10^{-4} 1.3×10^{-3} 2.0×10^{-2} 2.2×10^{-4} 1.1×10^{-3} 2.1×10^{-2} 1.8×10^{-4} 1.1×10^{-3} 1.9×10^{-4} 4.5×10^{-4}	
AC Current		(1 mA) 40 Hz ~ 10 kHz (1 ~ 100) mA 40 Hz ~ 10 kHz (100 mA ~ 1 A) 40 Hz ~ 10 kHz (1 ~ 10) A 40 Hz ~ 10 kHz (10 ~ 100) A 40 Hz ~ 10 kHz	58 μ A 8.6×10^{-4} 9.1×10^{-4} 9.9×10^{-4} 2.3×10^{-4}	
Timer		(1 ~ 100) s	5.8×10^{-6}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF signal generators	40426			Audio Analyzer, Digital Multimeter/ SICT-CP-40426
Frequency Test		(0.1 ~ 1) Hz 1 Hz ~ 100 MHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V (10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 8.0×10^{-5}	
DC Offset		($\pm$) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	 0.7 μ V 0.7 μ V 1.0×10^{-4} 6.0×10^{-5}	
Output Level Flatness Test		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	 0.099 dB 0.083 dB 0.095 dB 0.005 4 dB 0.001 1 dB 0.007 2 dB 0.021 dB 0.015 dB 0.027 dB	
Attenuator Test		(10 Hz ~ 100 kHz) (0 ~ 80) dB	 0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) % (1 ~ 100) kHz (3.16 ~ 0.010) %	 1.5×10^{-1} 3.2×10^{-1}	
Rise/Fall Time		100 μ s ~ 100 ns (100 ~ 10) ns (10 ~ 1) ns 1 ns ~ 100 ps	7.0×10^{-4} 7.8×10^{-4} 4.7×10^{-3} 4.6×10^{-2}	
Duty cycle		(1 ~ 99) %	0.006 1 %	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Reference Frequency		10 MHz	7.7×10^{-12}	
Center Frequency		10 Hz	6.1×10^{-5}	
		(10 ~ 100) Hz	6.1×10^{-6}	
		100 Hz ~ 1 kHz	6.1×10^{-7}	
		1 kHz ~ 1 MHz	6.1×10^{-8}	
		(1 ~ 100) MHz	6.1×10^{-9}	
		100 MHz ~ 1 GHz	6.1×10^{-8}	
Frequency Range		10 Hz	1.1×10^{-3}	
		(10 ~ 100) Hz	1.1×10^{-4}	
		100 Hz ~ 1 kHz	1.1×10^{-5}	
		(1 ~ 100) kHz	1.1×10^{-4}	
		100 kHz ~ 1 MHz	1.1×10^{-6}	
		(1 ~ 100) MHz	1.1×10^{-7}	
		100 MHz ~ 1 GHz	1.1×10^{-9}	
Resolution Bandwidth		100 Hz	3.3×10^{-2}	
		100 Hz ~ 3 kHz	3.1×10^{-2}	
		(3 ~ 300) kHz	3.3×10^{-2}	
		300 kHz ~ 1 MHz	3.4×10^{-2}	
Absolute Amplitude		(-60 dBV)		
		10 Hz	0.043 dB	
		10 Hz ~ 10 kHz	0.042 dB	
		(10 ~ 100) kHz	0.056 dB	
		(-60 ~ -50) dBV		
		10 Hz	0.017 dB	
		10 Hz ~ 10 kHz	0.015 dB	
		(10 ~ 100) kHz	0.022 dB	
		(-50 ~ -40) dBV		
		10 Hz	0.009 dB	
		10 Hz ~ 10 kHz	0.009 8 dB	
		(10 ~ 100) kHz	0.012 dB	
		(-40 ~ -30) dBV		
		10 Hz	0.009 dB	
		10 Hz ~ 10 kHz	0.006 8 dB	
		(10 ~ 100) kHz	0.011 dB	
		(-30 ~ 30) dBV		
		10 Hz	0.016 dB	
		10 Hz ~ 10 kHz	0.006 3 dB	
		(10 ~ 100) kHz	0.007 3 dB	
Referency Level		(-60 dBV)		
		10 Hz ~ 100 kHz	0.17 dB	
		(-60 ~ 30) dBV		
		10 Hz ~ 100 kHz	0.16 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Frequency Response		10 Hz 10 Hz ~ 100 kHz	0.009 1 dB 0.008 7 dB	
Logscale Fidelity		(0 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB	0.009 2 dB 0.012 dB 0.016 dB 0.042 dB	
Output frequency		10 Hz ~ 300 MHz	6.1×10^{-11}	
Input Impedance		1 M Ω (50 ~ 75) Ω	0.000 12 M Ω 0.000 7 Ω	
Output Voltage		10 mV 10 mV ~ 5 V	0.000 38 mV 8.8×10^{-5}	
Output Offset Voltage		(-30 ~ 30) V	6.7×10^{-6}	
Output Voltage Flatness		10 Hz ~ 100 kHz	0.000 67 dB	
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Frequency		(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 10 V (100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 10 V (10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	 0.099 dB 0.083 dB 0.095 dB 0.005 4 dB 0.001 1 dB 0.007 2 dB 0.010 dB 0.011 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Attenuation		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	
Sweep generators	40429			Audio Analyzer, Digital Multimeter/ SICT-CP-40429
Frequency		(0.1 ~ 1) Hz	5.8×10^{-7}	
		1 Hz ~ 100 kHz	5.8×10^{-9}	
Output Level Test		(10 ~ 100) Hz		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		100 mV ~ 10 V	7.0×10^{-5}	
		(100 Hz ~ 10 kHz)		
		1 mV	1.0×10^{-3}	
		(1 ~ 100) mV	1.0×10^{-4}	
		100 mV ~ 10 V	3.0×10^{-5}	
		(10 ~ 100) kHz		
		1 mV	1.0×10^{-3}	
		(1 ~ 100) mV	1.0×10^{-4}	
		100 mV ~ 10 V	7.0×10^{-5}	
Output Level Flatness		(100 mV)		
		(10 ~ 100) Hz	0.099 dB	
		100 Hz ~ 10 kHz	0.083 dB	
		(10 ~ 100) kHz	0.095 dB	
		(100 mV ~ 1 V)		
		(10 ~ 100) Hz	0.005 4 dB	
		100 Hz ~ 10 kHz	0.001 1 dB	
		(10 ~ 100) kHz	0.007 2 dB	
		(1 ~ 10) V		
		(10 ~ 100) Hz	0.010 dB	
		100 Hz ~ 100 kHz	0.011 dB	
Attenuation		(10 Hz ~ 10 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Signal transducers	40430			Digital Multimeter/ SICT-CP-40430
Voltage		1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 ~ 300) V	5.2×10^{-4} 8.8×10^{-5} 3.1×10^{-5} 4.3×10^{-3}	
Current		10 μ A (10 ~ 100) μ A 100 μ A ~ 100 mA 100 mA ~ 20 A	9.5×10^{-4} 9.7×10^{-5} 7.0×10^{-5} 2.0×10^{-4}	
Frequency		(1 ~ 10) Hz 10 Hz ~ 100 kHz	3.1×10^{-4} 7.2×10^{-5}	
Transistor curve tracers	40432			Digital Multimeter/ SICT-CP-40432
DC Voltage(Source)		0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	1.0 μ V 1.1×10^{-5} 1.0×10^{-5} 9.8×10^{-6} 1.1×10^{-5} 8.0×10^{-6}	
DC Current(Source)		0 nA (0 ~ 1) nA (1 ~ 100) nA (0.1 ~ 1) μ A (1 ~ 10) μ A 10 μ A ~ 10 mA (10 ~ 100) mA 100 mA ~ 10 A	0.12 nA 1.0×10^{-2} 8.0×10^{-3} 8.0×10^{-4} 9.0×10^{-5} 1.6×10^{-5} 5.0×10^{-5} 2.5×10^{-4}	
DC Voltage(Measure)		0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	0.78 μ V 1.5×10^{-5} 6.0×10^{-4} 6.4×10^{-4} 6.2×10^{-5}	
AC/DC high voltage generators	40434			High Voltage Digital Meter/ SICT-CP-40434
DC Voltage		($\pm$) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	 0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV (60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	 0.58 V 1.2×10^{-3} 6.1×10^{-4} 5.7×10^{-4} 1.2×10^{-2} 0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.2×10^{-4} 1.2×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC/DC high voltage probes	40435			DC Power Supply/ SICT-CP-40435
DC Voltage		($\pm$) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV	0.06 V 4.2×10^{-4} 2.1×10^{-4} 3.5×10^{-4}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV (60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4} 2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
Logic analyzers	40436			Calibrator/ SICT-CP-40436
DC Voltage		(0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	1.5 μ V 8.8×10^{-6} 7.2×10^{-6}	
Clock frequency		10 MHz	7.7×10^{-12}	
Telephone testers	40437			Tone Pulse Simulator/ SICT-CP-40437
L1, L2 Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 1 000) V	9.9×10^{-4} 1.1×10^{-5} 7.1×10^{-6} 8.5×10^{-6}	
Loop Current		(0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	2.5×10^{-5} 5.2×10^{-5} 2.2×10^{-4} 4.7×10^{-4}	
Ring Output Voltage		(10 Hz ~ 20 kHz) 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	 4.7×10^{-4} 2.9×10^{-4} 3.1×10^{-4}	
Ring Frequency		(1 ~ 1 000) Hz	7.0×10^{-5}	
D.T.M.F & Pulse		(+10 ~ -39.9) dBm	0.09 dB	
D.T.M.F & Frequency		(697 ~ 1 477) Hz	0.59 Hz	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal analyzers	40438			Video Amplitude Calibration Fixture/ SICT-CP-40438
Color Bar Decoding Accuracy(Gain)		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (500 ~ 1 000) mV	2.5×10^{-1} 5.0×10^{-2} 8.2×10^{-2} 4.1×10^{-2} 2.1×10^{-2} 8.5×10^{-3}	
Frequency		20 Hz ~ 5 MHz	5.8×10^{-6}	
Color Bar Decoding Accuracy(Phase)		(0 ~ 360)°	0.70°	
Measure Square Wave		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 400) mV (400 ~ 600) mV (600 ~ 999.9) mV	9.4×10^{-2} 2.0×10^{-2} 9.9×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 1.2×10^{-3} 9.4×10^{-4}	
Measure Sine Wave		No Filter, PAL NTS BW Lim, NTSC,PAL Chroma BP, NTSC,PAL (10 kHz ~ 10 MHz) 500 mV	7.0×10^{-3}	
Burst Frequency		(3 ~ 5) MHz	4.0×10^{-7}	
vertical Gain		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 600) mV (600 ~ 999.9) mV	9.4×10^{-2} 2.0×10^{-2} 9.9×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 9.4×10^{-4}	
Horizontal Frequency		(20 ~ 100) Hz 100 Hz ~ 10 kHz 10 kHz ~ 10 MHz	3.1×10^{-3} 6.1×10^{-4} 6.1×10^{-5}	
Gain Frequency Response		Flat, Luminance, Chroma at (20 Hz ~ 20 MHz) 700 mV	7.0×10^{-3}	
Transient Response		(0 ~ 1 000) mV	1.3×10^{-2}	
(Video Noise) Luminance Volt Level		(0 ~ -30) dB	4.8×10^{-1}	
Chrominance AM/PM Level		(0 ~ -30) dB	6.7×10^{-1}	
Luminance Volt Level		(0 ~ 1 000) mV	1.7×10^{-5}	
Luminance Input Level		(0 ~ 1 000) mV	1.8×10^{-5}	
Chrominance Input Level		(0 ~ 1 000) mV	1.7×10^{-5}	

405. Low frequency electric & magnetic fields

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Flux meters Flux	40503	0.1 mWb ~ 10 Wb	5.8×10^{-3}	Flux sources/ SICT-CP-40503
Flux sources Flux	40504	(0.1 ~ 50) mWb (0.05 ~ 0.1) Wb (0.1 ~ 10) Wb	6.6×10^{-4} 2.3×10^{-5} 1.4×10^{-5}	GPS receiver, Frequency counter/ SICT-CP-40504
Magnetometers Gauss	40508	(0 ~ 0.1) mT (0.1 ~ 0.5) mT (0.5 ~ 3) mT (3 ~ 5) mT (5 ~ 20) mT (20 ~ 30) mT (30 ~ 1 700) mT	7.1×10^{-2} 1.4×10^{-2} 7.0×10^{-3} 4.0×10^{-3} 3.0×10^{-3} 6.7×10^{-3} 6.4×10^{-3}	Helmholtz coil, Standard magnets/ SICT-CP-40508
Reference/standard magnets Gauss	40510	(1.5 ~ 30) mT (30 ~ 1 000) mT	7.3×10^{-3} 2.6×10^{-3}	Gaussmeters/ SICT-CP-40510

400. Radio Frequency Measurements				
Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers	40601			Power Sensor, Attenuator/ SICT-CP-40601
Gain		(0 ~ 80) dB 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 40) GHz	0.01 dB 0.02 dB 0.10 dB 0.11 dB 0.14 dB 0.19 dB	
Harmonics		(9 kHz ~ 26.5 GHz) (0 ~ -100) dBc	0.98 dB	
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.0×10^{-3} 2.1×10^{-2} 3.3×10^{-2}	
Coaxial attenuators	40602			Power Sensor, Directional Coupler/ SICT-CP-40602
Attenuation		(DC ~ 26.5 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB (26.5 ~ 34) GHz (0 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (34 ~ 40) GHz (0 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB	0.040 dB 0.042 dB 0.045 dB 0.050 dB 0.055 dB 0.074 dB 0.076 dB 0.089 dB 0.12 dB 0.14 dB 0.15 dB 0.26 dB 0.32 dB 0.37 dB 0.35 dB 0.45 dB 0.53 dB	
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Burst pulse generators	40605			Digital Oscilloscope/ SICT-CP-40605
Burst Volatage		50 Ω ($\pm$)		
		5 V	2.0×10^{-2}	
		(5 ~ 100) V	1.6×10^{-2}	
		(0.1 ~ 8) kV	2.0×10^{-2}	
		1 k Ω ($\pm$)		
		5 V ~ 8 kV	4.0×10^{-2}	
Rise/Fall Time		1 ns	2.0×10^{-2}	
		(1 ~ 2) ns	6.8×10^{-3}	
		(2 ~ 4) ns	2.6×10^{-3}	
		4 ns ~ 1 μ s	1.5×10^{-3}	
		(1 ~ 2) μ s	6.2×10^{-3}	
		(2 ~ 4) μ s	2.6×10^{-3}	
		(4 ~ 10) μ s	1.3×10^{-3}	
Pulse Width		1 ns	6.0×10^{-3}	
		(1 ~ 2) ns	3.1×10^{-3}	
		2 ns ~ 200 ms	1.5×10^{-3}	
Time measurement by section		1 ns	6.0×10^{-3}	
		(1 ~ 2) ns	3.1×10^{-3}	
		2 ns ~ 200 ms	1.5×10^{-3}	
Repeat Frequency		1 Hz ~ 25 MHz	1.6×10^{-3}	
Attenuator calibrators	40606			Verification Kit/ SICT-CP-40606
Attenuation		(0 ~ 10) dB	0.024 dB	
		(10 ~ 20) dB	0.025 dB	
		(20 ~ 30) dB	0.027 dB	
		(30 ~ 40) dB	0.029 dB	
		(40 ~ 50) dB	0.031 dB	
		(50 ~ 60) dB	0.034 dB	
		(60 ~ 70) dB	0.036 dB	
		(70 ~ 80) dB	0.039 dB	
		(80 ~ 90) dB	0.042 dB	
		(90 ~ 100) dB	0.045 dB	
		(100 ~ 110) dB	0.048 dB	
		(110 ~ 120) dB	0.052 dB	
RF power meter calibrators	40607			Digital Multimeter/ SICT-CP-40607
Output Power		3 μ W	0.1 nW	
		10 μ W	0.2 nW	
		30 μ W	0.5 nW	
		100 μ W	1 nW	
		300 μ W	4 nW	
		1 mW	0.18 μ W	
		3 mW	0.19 μ W	
		10 mW	0.2 μ W	
		30 mW	0.3 μ W	
		100 mW	1 μ W	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC transducers ; current probes, absorbing clamps, etc	40608			Power Senso, Network analyzer/ SICT-CP-40608
Transfer Impedance		10 Hz ~ 50 MHz	0.60 dB	
		(50 ~ 200) MHz	1.1 dB	
		200 MHz ~ 3 GHz	1.9 dB	
Insertion Loss		30 MHz ~ 1 GHz	1.9 dB	
Electric Magnetic Near-Field		100 kHz ~ 1 GHz	1.9 dB	
Reflection coefficient		(0 ~ 1)		
		10 Hz ~ 1 GHz	4.2×10^{-3}	
		(1 ~ 3) GHz	6.0×10^{-3}	
SWR		(1 ~ ∞)		
		10 Hz ~ 1 GHz	9.0×10^{-3}	
		(1 ~ 3) GHz	1.3×10^{-2}	
Coaxial directional couplers/ splitters	40610			Power Sensor, Synthesized Sweeper/ SICT-CP-40610
Coupling Factor		(0 ~ 30) dB		
		10 Hz ~ 10 kHz	0.005 9 dB	
		(10 ~ 100) kHz	0.007 3 dB	
		100 kHz ~ 3 GHz	0.087 dB	
		(3 ~ 8) GHz	0.095 dB	
		(8 ~ 18) GHz	0.11 dB	
		(18 ~ 26.5) GHz	0.14 dB	
		(26.5 ~ 40) GHz	0.20 dB	
		(30 ~ 80) dB		
		10 Hz ~ 10 kHz	0.006 5 dB	
		(10 ~ 100) kHz	0.021 dB	
		100 kHz ~ 4 GHz	0.088 dB	
		(4 ~ 8) GHz	0.092 dB	
		(8 ~ 10) GHz	0.10 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 26.5) GHz	0.13 dB	
		(26.5 ~ 40) GHz	0.20 dB	
Reflection coefficient		(0 ~ 1)		
		20 Hz ~ 1 GHz	4.2×10^{-3}	
		(1 ~ 20) GHz	9.4×10^{-3}	
		(20 ~ 40) GHz	1.5×10^{-2}	
SWR		(1 ~ ∞)		
		20 Hz ~ 1 GHz	9.7×10^{-3}	
		(1 ~ 20) GHz	2.4×10^{-2}	
		(20 ~ 40) GHz	3.8×10^{-2}	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators Peak Current(Ip)	40613	(±) (3.75 ~ 7.5) A (7.5 ~ 15) A (15 ~ 22.5) A (22.5 ~ 56.3) A (56.3 ~ 93.8) A (93.8 ~ 112.5) A (112.5 ~ 150) A	5.1×10^{-2} 5.3×10^{-2} 4.6×10^{-2} 5.2×10^{-2} 4.9×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	Digital Oscilloscope/ SICT-CP-40613
Current I1 (30 ~ 60) ns		(±) 2 A (2 ~ 4) A (4 ~ 8) A (8 ~ 16) A (16 ~ 36) A (36 ~ 50) A (50 ~ 60) A (60 ~ 80) A	4.5×10^{-2} 5.0×10^{-2} 5.3×10^{-2} 4.9×10^{-2} 5.0×10^{-2} 4.4×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I2 (60 ~ 130) ns		(±) 1 A (1 ~ 2) A (2 ~ 4) A (4 ~ 6) A (6 ~ 8) A (8 ~ 15) A (15 ~ 25) A (25 ~ 30) A (30 ~ 40) A	5.0×10^{-2} 5.4×10^{-2} 5.7×10^{-2} 4.9×10^{-2} 5.4×10^{-2} 6.5×10^{-2} 5.2×10^{-2} 6.7×10^{-2} 6.1×10^{-2}	
Current I3 (360 ~ 800) ns		(±) 0.275 A (0.275 ~ 0.55) A (0.55 ~ 1.1) A (1.1 ~ 1.65) A (1.65 ~ 4.13) A (4.13 ~ 6.88) A (6.88 ~ 8.25) A (8.25 ~ 11) A	1.5×10^{-1} 2.2×10^{-1} 1.9×10^{-1} 1.6×10^{-1} 2.3×10^{-1} 1.5×10^{-1} 2.4×10^{-1} 1.9×10^{-1}	
Current I4 (30 ~ 65) ns		(±) 0.15 A (0.15 ~ 0.3) A (0.3 ~ 0.6) A (0.6 ~ 1.2) A (1.2 ~ 2.25) A (2.25 ~ 2.7) A (2.7 ~ 3.75) A (3.75 ~ 4.5) A (4.5 ~ 6) A	1.3×10^{-1} 1.9×10^{-1} 3.0×10^{-1} 2.1×10^{-1} 2.5×10^{-1} 2.0×10^{-1} 1.5×10^{-1} 2.9×10^{-1} 2.0×10^{-1}	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators	40613	($\pm$)		Digital Oscilloscope/ SICT-CP-40613
Semiconductor Peak Current HBM		(0.083 ~ 0.17) A	4.8×10^{-2}	
		(0.17 ~ 0.33) A	3.6×10^{-2}	
		(0.33 ~ 0.67) A	3.4×10^{-2}	
		(0.67 ~ 1.33) A	3.2×10^{-2}	
		(1.33 ~ 6.66) A	2.7×10^{-2}	
Semiconductor Peak Current MM		($\pm$)		
		(0.88 ~ 1.75) A	6.6×10^{-2}	
		(1.75 ~ 14) A	3.5×10^{-2}	
		(14 ~ 17.5) A	3.1×10^{-2}	
		(17.5 ~ 26.25) A	2.7×10^{-2}	
		(26.25 ~ 35) A	3.0×10^{-2}	
Time		0.1 ns	2.7×10^{-2}	
		0.1 ns ~ 1 ms	2.4×10^{-2}	
Peak Voltage		($\pm$)		
		1 kV	3.0×10^{-2}	
		(1 ~ 35) kV	2.5×10^{-2}	
EMC receivers	40614	100 kHz ~ 1 GHz	6.1×10^{-10}	Network Analyzer, Pulse Generator/ SICT-CP-40614
Frequency accuracy				
SWR		9 kHz ~ 50 MHz	0.011	
		50 MHz ~ 8 GHz	0.028	
		(8 ~ 19) GHz	0.035	
		(19 ~ 26) GHz	0.045	
		(26 ~ 40) GHz	0.064	
IF Band Accuracy		1 Hz ~ 10 MHz	65 mHz	
IF Band Selectivity		1 Hz ~ 10 MHz	6.7×10^{-4}	
IF Band Linearity		1 Hz ~ 1 GHz	0.12 dB	
Frequency Response		10 Hz ~ 5 kHz	0.01 dB	
		5 kHz ~ 1 GHz	0.07 dB	
		(1 ~ 10) GHz	0.08 dB	
		(10 ~ 18) GHz	0.09 dB	
		(18 ~ 26.5) GHz	0.11 dB	
		(26.5 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.20 dB	
Frequency Response (CISPR)		9 kHz ~ 1 GHz	0.80 dB	
Display linearity accuracy		(80 ~ 50) dB μ V	0.10 dB	
		(50 ~ 0) dB μ V	0.14 dB	
Input Attenuation		(0 ~ 30) dB	0.15 dB	
		(30 ~ 70) dB	0.12 dB	
		(70 ~ 110) dB	0.10 dB	
Noise Indicator		DC ~ 26.5 GHz	0.16 dB	
Interference Immunity		9 kHz ~ 40 GHz	0.67 dB	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF filters	40615	Reject Frequency (9 ~ 90) kHz (90 ~ 900) kHz 900 kHz ~ 900 MHz 900 MHz ~ 18 GHz (18 ~ 50) GHz Insertion Loss (9 kHz ~ 8 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (8 ~ 18) GHz (0 ~ 10) dB (10 ~ 30) dB (30 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (18 ~ 50) GHz (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB	0.024 kHz 0.24 kHz 0.025 MHz 0.068 MHz 0.12 MHz 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.18 dB 0.23 dB 0.66 dB 1.7 dB 4.3 dB 0.23 dB 0.24 dB 0.25 dB 0.26 dB 0.31 dB 0.73 dB 1.7 dB 4.3 dB 0.48 dB 0.51 dB 0.52 dB 0.53 dB 0.54 dB 0.59 dB 0.78 dB 1.6 dB 2.8 dB 6.0 dB	Network Analyzer/ SICT-CP-40615
RF impedance meters	40616	RF Level (100 kHz ~ 18 GHz) (35 ~ 20) dBm Frequency (20 Hz ~ 18 GHz) (20 ~ -70) dBm 9 kHz ~ 0.1 MHz 0.1 MHz ~ 18 GHz Load Measurement DC 10 Hz ~ 100 MHz (100 ~ 500) MHz 500 MHz ~ 1.8 GHz (1.8 ~ 3.0) GHz (3.0 ~ 18) GHz	0.11 dB 0.12 dB 6.8×10^{-10} 6.2×10^{-11} 0.02 Ω 0.06 Ω 0.15 Ω 0.21 Ω 0.41 Ω 1.1 Ω	Performance Kit/ SICT-CP-40616

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF impulse generators Pulse Level	40617	9 kHz ~ 1 GHz	0.33 dB	Digital Oscilloscope/ SICT-CP-40617
Line impedance stabilization networks; LISN, CDN, ISN, etc. Impedance Phase Insertion Loss Decoupling attenuation(Isolation) Coupling/Decoupling network(Impedance) Coupling/Decoupling network (Insertion loss)	40618	9 kHz ~ 1 000 MHz 9 kHz ~ 1 000 MHz (0 ~ 100) dB 9 kHz ~ 100 MHz (100 ~ 1 000) MHz (0 ~ 100) dB (9 ~ 30) kHz (0.03 ~ 20) MHz (20 ~ 1 000) MHz 9 kHz ~ 1 000 MHz (0 ~ 100) dB 9 kHz ~ 1 000 MHz	2.0×10^{-2} 1.2° 0.07 dB 0.08 dB 0.21 dB 0.20 dB 0.21 dB 2.0×10^{-2} 0.10 dB	Impedance/Gain-Phase Analyzer, Calibration Kit/ SICT-CP-40618
Coaxial standard mismatches Reflection coefficient(Γ) SWR	40619	(0 ~ 1) (9 ~ 100) kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (1 ~ ∞) (9 ~ 100) kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz	0.004 3 0.006 0 0.009 5 0.016 0.008 6 0.012 0.019 0.032	Network Analyzer, Calibration Kit/ SICT-CP-40619
Mobile communication test sets RF Output Level	40621	(35 ~ 20) dBm 100 kHz ~ 18 GHz) (20 ~ -20) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-60 ~ -70) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-70 ~ -120) dBm 9 kHz ~ 26.5 GHz	0.08 dB 0.06 dB 0.11 dB 0.14 dB 0.18 dB 0.06 dB 0.11 dB 0.16 dB 0.21 dB 0.08 dB 0.10 dB 0.18 dB 0.21 dB 0.15 dB	Measuring Receiver, RF Signal Generator/ SICT-CP-40621

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621			Measuring Receiver, RF Signal Generator/ SICT-CP-40621
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion Harmonics of Modulation Rate Signal		≤ 20 %	2.3×10^{-2}	
Hamonics		(0 ~ -90) dB	0.36 dB	
Frequency Output Accuracy		9 kHz ~ 40 GHz	7.9×10^{-11}	
AC Output Level		(10 Hz ~ 100 kHz)		
		(1 ~ 100) mV	5.2×10^{-4}	
		100 mV ~ 100 V	5.8×10^{-4}	
DC Output Level		1 mV ~ 1 V	1.5×10^{-6}	
		(1 ~ 10) V	1.3×10^{-6}	
		(10 ~ 100) V	1.8×10^{-6}	
AC Input Level		(10 Hz ~ 100 kHz)		
		(1 ~ 10) mV	7.6×10^{-3}	
		(10 ~ 100) mV	1.3×10^{-3}	
		(0.1 ~ 1) V	6.7×10^{-4}	
		(1 ~ 10) V	1.7×10^{-4}	
		(10 ~ 100) V	2.0×10^{-4}	
DC Input Level		(1 ~ 10) mV	1.3×10^{-3}	
		(10 ~ 100) mV	3.1×10^{-4}	
		(0.1 ~ 1) V	6.2×10^{-5}	
		(1 ~ 100) V	6.1×10^{-5}	
RF Input Level		(9 kHz ~ 18 GHz)		
		(10 ~ -70) dBm	0.10 dB	
		(18 ~ 40) GHz		
		(10 ~ -70) dBm	0.16 dB	
Modulation meters	40622			Measuring Receiver/ SICT-CP-40622
Amplitude Modulation		0 kHz	0.01 %	
		(0 ~ 400) kHz	1.2×10^{-2}	
Frequency Modulation		0 %	1 Hz	
		(0 ~ 100) %	1.2×10^{-2}	
Phase Modulation		0 rad	1.2 mrad	
		(0 ~ 400) rad	1.2×10^{-2}	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623	10 Hz ~ 40 GHz	6.8×10^{-10}	Power Sensor, Verification Kit/ SICT-CP-40623
Frequency		(20 ~ -30) dBm		
Source Power Level		10 Hz ~ 100 Hz	0.01 dB	
		100 Hz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.21 dB	
		(50 ~ 80) GHz	0.45 dB	
		(80 ~ 110) GHz	0.53 dB	
		(-30 ~ -60) dBm		
		9 kHz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
Dynamic Range		(100 kHz ~ 18 GHz)		
		(0 ~ 10) dB	0.086 dB	
		(10 ~ 20) dB	0.087 dB	
		(20 ~ 30) dB	0.088 dB	
		(30 ~ 40) dB	0.091 dB	
		(40 ~ 50) dB	0.099 dB	
		(50 ~ 60) dB	0.11 dB	
		(60 ~ 70) dB	0.12 dB	
		(70 ~ 80) dB	0.14 dB	
		(80 ~ 90) dB	0.15 dB	
		(90 ~ 100) dB	0.18 dB	
Attenuation		(20 dB)		
		300 kHz ~ 1.5 GHz	0.050 dB	
		(1.5 ~ 8) GHz	0.051 dB	
		(8 ~ 18) GHz	0.055 dB	
		(18 ~ 26.5) GHz	0.067 dB	
		(40 dB)		
		300 kHz ~ 1.5 GHz	0.054 dB	
		(1.5 ~ 8) GHz	0.055 dB	
		(8 ~ 18) GHz	0.059 dB	
		(18 ~ 26.5) GHz	0.082 dB	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623	($\pm 180^\circ$) 300 kHz ~ 45 MHz 45 MHz ~ 2.0 GHz (2.0 ~ 3.0) GHz (3.0 ~ 4.5) GHz (4.5 ~ 6.0) GHz (6.0 ~ 7.5) GHz (7.5 ~ 8.0) GHz (8.0 ~ 9.0) GHz (9.0 ~ 10.5) GHz (10.5 ~ 12.0) GHz (12.0 ~ 13.5) GHz (13.5 ~ 15.0) GHz (15.0 ~ 16.5) GHz (16.5 ~ 18.0) GHz (18.0 ~ 21.0) GHz (21.0 ~ 22.5) GHz (22.5 ~ 24.0) GHz (24.0 ~ 25.5) GHz (25.5 ~ 26.5) GHz	0.04° 0.09° 0.10° 0.11° 0.15° 0.19° 0.21° 0.22° 0.29° 0.28° 0.26° 0.27° 0.26° 0.29° 0.31° 0.28° 0.39° 0.33° 0.44°	Power Sensor, Verification Kit/ SICT-CP-40623
Phase				
SWR		(30 kHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0.021 0.021 0.021 0.021 0.018 0.018 0.018 0.024	
Noise figure meters	40624	10 MHz ~ 26.5 GHz	4.3×10^{-5}	Noise Source/ SICT-CP-40624
Tuning Accuracy				
Referency frequency		10 MHz	6.1×10^{-10}	
Input VSWR		9 kHz ~ 1 GHz (1 ~ 20) GHz (20 ~ 26.5) GHz	0.008 0.019 0.03	
DC voltage		(0 ~ 28) V	0.000 18 V	
Range		(0 ~ 30) dB	0.052 dB	
Noise Figure		10 MHz ~ 8 GHz (8 ~ 18) GHz (18 ~ 26.5) GHz	0.12 dB 0.16 dB 0.37 dB	
Noise generators	40625	(-80 ~ -130) dBm/Hz	0.10 dB	Spectrum Analyzer/ SICT-CP-40625
Noise Power				
Scale Fidelity		(0 ~ 50) dB	0.27 dB	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise impulse simulators Peak Voltage	40626	($\pm$) 0.1 kV (0.1 ~ 5) kV	4.0×10^{-2} 3.5×10^{-2}	Digital Oscilloscope/ SICT-CP-40626
Rise/Fall Time		1 ns (1 ~ 2) ns (2 ~ 4) ns	6.0×10^{-3} 3.1×10^{-3} 1.5×10^{-3}	
Pulse Width		10 ns (10 ~ 1 000) ns	2.0×10^{-3} 1.5×10^{-3}	
RF phase noise meters RF phase noise	40627	(Carrier Frequency) 100 MHz ~ 18 GHz (Offset Frequency) 10 Hz ~ 100 MHz	1.0 dB 1.0 dB	RF Signal analyzer/ SICT-CP-40627
Coaxial noise sources ENR	40628	(4.5 ~ 16) dB (0.01 ~ 1) GHz (1 ~ 7) GHz (7 ~ 8) GHz (8 ~ 14) GHz (14 ~ 18) GHz (12 ~ 17) dB (0.01 ~ 1) GHz (1 ~ 2) GHz (2 ~ 6) GHz (6 ~ 7) GHz (7 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz	0.28 dB 0.27 dB 0.30 dB 0.31 dB 0.32 dB 0.31 dB 0.28 dB 0.30 dB 0.29 dB 0.40 dB 0.41 dB 0.47 dB	Coaxial noise sources, Noise figure analyzer/ SICT-CP-40628
SWR		(0 ~1) (0.01 ~ 3) GHz (3 ~ 20) GHz (20 ~ 26.5) GHz	0.006 8 0.010 0.015	
RF power meters High power	40635	(0.1 ~ 500) W 10 kHz ~ 250 MHz (0.1 ~ 150) W (80 ~ 1 000) MHz (0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.6×10^{-2} 2.6×10^{-2} 2.7×10^{-2}	Range Calibrator/ SICT-CP-40635
Zero Carryover		10 μ W ~ 1 mW (1 ~ 100) mW	3 nW 0.01 mW	
Power		3 μ W ~ 100 mW	1.6×10^{-3}	
Calibration Factor		(88 ~ 100) %	0.5×10^{-3}	
Power Ref. Output		50 MHz, 1 mW	8 μ W	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Diode power sensors Cal factor	40636	(1 ~ 10) μ W 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (10 μ W ~ 10 mW) 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	 0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.8×10^{-2} 4.0×10^{-2} 0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.6×10^{-2} 3.5×10^{-2}	Therimistor Mount, Synthesized Sweeper/ SICT-CP-40636
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	 4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	 9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	
Thermocouple power sensors Cal Factor	40637	(1 ~ 10) μ W 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (10 μ W ~ 10 mW) 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	 0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.8×10^{-2} 4.0×10^{-2} 0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.6×10^{-2} 3.5×10^{-2}	Therimistor Mount, Synthesized Sweeper/ SICT-CP-40637
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	 4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	 9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Pulse generators	40638			Digital Oscilloscope/ SICT-CP-40638
Period		100 ps ~ 1 s	5.8×10^{-9}	
Frequency		1 Hz ~ 1 GHz (1 ~ 3.35) GHz	5.8×10^{-9} 1.7×10^{-8}	
Width		100 ps ~ 1 s	1.2×10^{-3}	
Delay Time		100 ps ~ 1 s	1.2×10^{-3}	
Double Pulse		100 ps ~ 1 s	1.2×10^{-3}	
Duty Cycle		(1 ~ 99) %	0.006 2 %	
DC Level		$\pm$ (10 mV ~ 100 V)	5.8×10^{-4}	
Output Level		(100 Hz ~ 10 kHz) (10 ~ -20) dBm	0.018 dB	
Radar test sets	40639			Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
(Ship) RF Level		(20 ~ -20) dBm 20 Hz ~ 1 GHz (1 ~ 18) GHz	0.09 dB 0.13 dB	
		(-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz	0.06 dB 0.11 dB	
		(-60 ~ -120) dBm 10 MHz ~ 18 GHz	0.25 dB	
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion of Modulation		(0 ~ 2) %	1.2×10^{-3}	
Hamonics		(9 kHz ~ 18 GHz) (0 ~ -110) dB	0.25 dB	
Frequency		9 kHz ~ 18 GHz	6.2×10^{-11}	
Pulse Period		1 ns ~ 10 ms	1.2×10^{-2}	
High power		(0.1 ~ 500) W 10 kHz ~ 250 MHz	2.6×10^{-2}	
		(0.1 ~ 150) W (80 ~ 1 000) MHz	2.6×10^{-2}	
		(0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.7×10^{-2}	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets (flight) Frequency(VOR/ILS/DME) Amplitude Modulation(VOR/ILS)	40639	(74.6 ~ 1 150) MHz Localizer (108.1 ~ 111.95) MHz (0.1 ~ 20) % Glideslope (330.95 ~ 334.70) MHz (20 ~ 40) % Marker Beacon (74.6 ~ 75.4) MHz (40 ~ 95) % VOR (108 ~ 117.95) MHz (0.1 ~ 30) %	8.2×10^{-8} 0.62 % 0.84 % 1.4 % 0.62 %	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
고주파 레벨(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm Glideslope (330.95 ~ 334.70) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB 0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
DDM(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz LEFT(-0.200 ~ -0.155) LEFT(-0.155 ~ -0.093) CENTER(0.000) RIGHT(0.093 ~ 0.155) RIGHT(0.155 ~ 0.200) Glideslope (330.95 ~ 334.70) MHz DOWN(0.400 ~ 0.175) DOWN(0.175 ~ 0.091) CENTER(0.000) UP(-0.091 ~ -0.175) UP(-0.175 ~ -0.400)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5 0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets SDM(VOR/ILS) BEARING ANGLE(VOR)	40639	Localizer (108.1 ~ 111.95) MHz (0 ~ 40) % Glideslope (330.95 ~ 334.70) MHz (40 ~ 80) % VOR (108 ~ 117.95) MHz (0 ~ 360) °	0.87 % 1.2 % 0.04 °	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
RF signal generators Frequency Modulation Amplitude Modulation Phase Modulation Pulse Modulation Distortion RF Level	40640	(0.1 ~ 400) kHz (0.1 ~ 100) % (0.1 ~ 400) rad (100 kHz ~ 12 000 MHz) Period (1 μs ~ 1 s) ton (100 ns ~ 100 μs) PRR ≤ 20 % (54 ~ 57) dBm 100 kHz ~ 500 MHz (51 ~ 54) dBm 100 kHz ~ 2 GHz (35 ~ 51) dBm 100 kHz ~ 18 GHz (20 ~ 35) dBm 100 kHz ~ 10 GHz (10 ~ 18) GHz (-30 ~ 20) dBm 20 Hz ~ 10 GHz (10 ~ 18) GHz (18 ~ 28) GHz (28 ~ 40) GHz (40 ~ 50) GHz (50 ~ 70) GHz (70 ~ 110) GHz (-30 ~ -60) dBm 20 Hz ~ 10 GHz (10 ~ 18) GHz (18 ~ 28) GHz (28 ~ 40) GHz (40 ~ 50) GHz	1.2×10^{-2} 1.2×10^{-2} 1.2×10^{-2} 1.2×10^{-3} 1.2×10^{-3} 3.1×10^{-3} 2.3×10^{-2} 0.35 dB 0.32 dB 0.32 dB 0.26 dB 0.27 dB 0.09 dB 0.11 dB 0.12 dB 0.16 dB 0.20 dB 0.29 dB 0.38 dB 0.10 dB 0.11 dB 0.12 dB 0.16 dB 0.21 dB	Measuring Receiver/ SICT-CP-40640

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators	40640	(-120 ~ -60) dBm		Measuring Receiver/ SICT-CP-40640
RF Level		20 Hz ~ 4.2 GHz	0.21 dB	
		(4.2 ~ 8) GHz	0.23 dB	
		(8 ~ 12.4) GHz	0.24 dB	
		(12.4 ~ 18) GHz	0.27 dB	
		(18 ~ 26.5) GHz	0.31 dB	
Harmonic		(-10 ~ -110) dBc	0.37 dB	
Frequency		9 kHz ~ 40 GHz	2.1×10^{-11}	
RF spectrum analyzers	40641	(3 ~ 100) Hz	2.0×10^{-4}	Power Sensor, Synthesized Sweeper/ SICT-CP-40641
Center Frequency		(100 ~ 500) Hz	6.1×10^{-6}	
		(500 ~ 900) Hz	1.2×10^{-6}	
		900 Hz ~ 100 kHz	6.8×10^{-7}	
		0.1 MHz ~ 40 GHz	6.2×10^{-9}	
Frequency Counter		(3 ~ 100) Hz	2.0×10^{-4}	
		(100 ~ 500) Hz	6.1×10^{-6}	
		(500 ~ 900) Hz	1.2×10^{-6}	
		900 Hz ~ 100 kHz	6.8×10^{-7}	
		0.1 MHz ~ 40 GHz	6.2×10^{-9}	
Span		10 Hz ~ 100 kHz	7.6×10^{-3}	
		0.1 MHz ~ 40 GHz	7.7×10^{-6}	
RBW		1 Hz ~ 100 MHz	6.2×10^{-6}	
RBW Selectivity		1 Hz ~ 100 MHz	3.2×10^{-2}	
RBW Switching		1 Hz ~ 100 MHz	0.022 dB	
Scale Switching		1 dB ~ 10 dB scale/div	0.022 dB	
Scale Fidelity		(0 ~ -30) dB	0.073 dB	
		(-30 ~ -40) dB	0.077 dB	
		(-40 ~ -50) dB	0.082 dB	
		(-50 ~ -60) dB	0.095 dB	
		(-60 ~ -80) dB	0.13 dB	
		(-80 ~ -100) dB	0.17 dB	
Frequency Response		(10 ~ 100) Hz	0.01 dB	
		100 Hz ~ 1 GHz	0.15 dB	
		(1 ~ 6) GHz	0.16 dB	
		(6 ~ 10) GHz	0.17 dB	
		(10 ~ 18) GHz	0.19 dB	
		(18 ~ 26.5) GHz	0.21 dB	
		(26.5 ~ 30) GHz	0.23 dB	
		(30 ~ 35) GHz	0.28 dB	
		(35 ~ 40) GHz	0.32 dB	
Average Noise Level		DC ~ 40 GHz	0.17 dB	
Sideband Noise Level		(-30 ~ 30) kHz	0.33 dB	
CAL Output Freq. & Int. Frequency		DC ~ 1 GHz	6.2×10^{-9}	
CAL Output Level		(-20 ~ 20) dBm	0.09 dB	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Surge generators	40643			Digital Oscilloscope/ SICT-CP-40643
Surge Voltage		($\pm$) 2 V (2 ~ 5) V 5 V ~ 200 kV	7.0×10^{-2} 4.0×10^{-2} 3.5×10^{-2}	
Surge Current		($\pm$) 5 A ~ 200 kA	3.6×10^{-2}	
Rise/Fall Time		1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	
Pulse Width		1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	
Time measurement by section		1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	
Frequency measurement by section		1 Hz ~ 25 MHz	1.6×10^{-3}	
Phase Shifting		at 50 Hz (0 ~ 360) $^{\circ}$ at 60 Hz (0 ~ 360) $^{\circ}$	1.2° 1.4°	
SWR meters	40644			Coaxial Mismatch/ SICT-CP-40644
Frequency		9 kHz ~ 18 GHz	6.4×10^{-5}	
Output Level		30 kHz ~ 100 MHz 100 MHz ~ 10 GHz (10 ~ 18) GHz	0.06 dB 0.08 dB 0.09 dB	
SWR		(30 kHz ~ 30 MHz) 1.05 1.20 1.50 2.00 (30 MHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0.019 0.019 0.019 0.020 0.021 0.021 0.021 0.021 0.018 0.018 0.018 0.024	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Open,Short,Phase)	40645	($\pm 180^\circ$)		Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
		10 Hz ~ 2 GHz	0.49°	
		(2 ~ 20) GHz	0.61°	
		(20 ~ 40) GHz	0.95°	
		(40 ~ 50) GHz	1.2°	
(Reflection coefficient)		(0 ~ 1)		
		10 Hz ~ 100 kHz	0.004 3	
		100 kHz ~ 500 MHz	0.006 0	
		500 MHz ~ 20 GHz	0.009 5	
		(20 ~ 40) GHz	0.016	
		(40 ~ 50) GHz	0.019	
(SWR)		(1 ~ ∞)		
		10 Hz ~ 100 kHz	0.008 6	
		100 kHz ~ 500 MHz	0.012	
		500 MHz ~ 20 GHz	0.019	
		(20 ~ 40) GHz	0.032	
		(40 ~ 50) GHz	0.038	
(Impedance)		(0.000 0 ~ 0.047 6)		
		10 Hz ~ 500 MHz	0.64 Ω	
		500 MHz ~ 20 GHz	1.0 Ω	
		(20 ~ 40) GHz	1.6 Ω	
		(40 ~ 50) GHz	2.0 Ω	
		(0.047 6 ~ 0.090 9)		
		10 Hz ~ 500 MHz	0.71 Ω	
		500 MHz ~ 20 GHz	1.1 Ω	
		(20 ~ 40) GHz	1.8 Ω	
		(40 ~ 50) GHz	2.2 Ω	
		(0.090 9 ~ 0.166 7)		
		10 Hz ~ 500 MHz	0.84 Ω	
		500 MHz ~ 20 GHz	1.3 Ω	
		(20 ~ 40) GHz	2.1 Ω	
		(40 ~ 50) GHz	2.7 Ω	
		(0.166 7 ~ 0.230 8)		
		10 Hz ~ 500 MHz	0.99 Ω	
		500 MHz ~ 20 GHz	1.6 Ω	
		(20 ~ 40) GHz	2.5 Ω	
		(40 ~ 50) GHz	3.1 Ω	
		(0.230 8 ~ 0.285 7)		
		10 Hz ~ 500 MHz	1.1 Ω	
		500 MHz ~ 20 GHz	1.8 Ω	
		(20 ~ 40) GHz	2.9 Ω	
		(40 ~ 50) GHz	3.6 Ω	
		(0.285 7 ~ 0.333 4)		
		10 Hz ~ 500 MHz	1.3 Ω	
		500 MHz ~ 20 GHz	2.1 Ω	
		(20 ~ 40) GHz	3.3 Ω	
		(40 ~ 50) GHz	4.2 Ω	

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Impedance Phase)	40645	(0.000 0 ~ 0.047 6, $\pm 180^\circ$) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.047 6 ~ 0.090 9, $\pm 180^\circ$) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.090 9 ~ 0.166 7, $\pm 180^\circ$) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.166 7 ~ 0.230 8, $\pm 180^\circ$) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.230 8 ~ 0.285 7, $\pm 180^\circ$) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.285 7 ~ 0.333 4, $\pm 180^\circ$) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.68° 1.1° 1.7° 2.2° 0.67° 1.1° 1.7° 2.1° 0.66° 1.1° 1.7° 2.1° 0.65° 1.0° 1.6° 2.1° 0.64° 1.0° 1.6° 2.0° 0.62° 1.0° 1.6° 2.0°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
Coaxial thermistor mounts Cal Factor	40646	(1 ~ 10) μ W (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (10 μ W ~ 10 mW) (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.4 $\times 10^{-2}$ 0.9 $\times 10^{-2}$ 1.1 $\times 10^{-2}$ 1.9 $\times 10^{-2}$ 2.8 $\times 10^{-2}$ 4.0 $\times 10^{-2}$ 0.4 $\times 10^{-2}$ 0.9 $\times 10^{-2}$ 1.1 $\times 10^{-2}$ 1.9 $\times 10^{-2}$ 2.6 $\times 10^{-2}$ 3.5 $\times 10^{-2}$	Therimistor Mount, Synthesized Sweeper/ SICT-CP-40646

406. Radio frequency measurements

400. Radio Frequency Measurements				
Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial thermistor mounts Reflection coefficient SWR	40646	(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz (1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	 4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2} 9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	Synthesized Sweeper/ SICT-CP-40646
RF voltmeters RF Voltage	40650	3 V 1 V 300 mV 270 mV 240 mV 210 mV 180 mV 150 mV 120 mV 100 mV 90 mV 60 mV 30 mV 10 mV 3 mV 1 mV	4.2 mV 1.4 mV 0.36 mV 0.32 mV 0.28 mV 0.26 mV 0.24 mV 0.22 mV 0.17 mV 0.16 mV 0.13 mV 0.11 mV 0.048 mV 0.020 mV 0.018 mV 0.013 mV	RF Millivolt Meter Calibrator/ SICT-CP-40650
Vector voltmeters RF Voltage RF Phase	40651	3 V 1 V 300 mV 100 mV 30 mV 10 mV 3 mV 1 mV (0 ~ 270)°	4.2 mV 1.4 mV 0.36 mV 0.16 mV 0.048 mV 0.046 mV 0.022 mV 0.024 mV 0.006°	Signal Generator/ SICT-CP-40651
Field strength meters Center frequency Scale Fidelity Frequency response	40652	(9 ~ 100) kHz 0.1 MHz ~ 18 GHz (0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -100) dB 9 kHz ~ 500 MHz 500 MHz ~ 18 GHz	 6.8×10^{-8} 6.2×10^{-9} 0.11 dB 0.12 dB 0.13 dB 0.18 dB 0.05 dB 0.08 dB	Signal Generator/ SICT-CP-40652
AM/FM test sources Output frequency	40653	(10 ~ 560) MHz	6.2×10^{-10}	Measuring Receiver/ SICT-CP-40653

406. Radio frequency measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip simulators	40654			Digital Oscilloscope/ SICT-CP-40654
DC Voltage		1 V (1 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V	1.1×10^{-5} 6.4×10^{-6} 5.4×10^{-6} 9.3×10^{-6} 8.0×10^{-6} 1.1×10^{-5}	
AC Voltage		(50 ~ 60) Hz 50 V (50 ~ 100) V (100 ~ 500) V	5.0×10^{-5} 2.5×10^{-5} 1.6×10^{-4}	
Frequency		50 Hz 60 Hz	8.4×10^{-6} 8.3×10^{-6}	
Dip DC Voltage		(0 ~ 50) V 0 % (0 ~ 120) %	0.2 V 3.4×10^{-2}	
Dip AC Voltage		(50 Hz ~ 60 Hz, 0 V ~ 400 V) 0 % (0 ~ 120) %	0.9 V 3.4×10^{-2}	
Time measurement by section		100 ns ~ 2 μ s (2 ~ 4) μ s (4 ~ 400) μ s (0.4 ~ 2) ms 2 ms ~ 5 s	1.8×10^{-3} 2.0×10^{-3} 1.6×10^{-3} 2.0×10^{-3} 1.6×10^{-3}	
Inrush Current		(5 ~ 1 000) A	3.6×10^{-2}	
Phase Shifting		at 50 Hz (0 ~ 360) $^{\circ}$ at 60 Hz (0 ~ 360) $^{\circ}$	1.2° 1.4°	

407. Field strength & antennas

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Loop antennas Antenna Factor	40704	10 Hz ~ 30 MHz	1.2 dB	Signal generator1, Signal analyzer/ SICT-CP-40704
Monopole antennas Antenna Factor	40705	10 Hz ~ 30 MHz	1.4 dB	Signal generator1, Signal analyzer/ SICT-CP-40705

501. Contact temperature

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 ℃ (-196 ~ -95) ℃ (-95 ~ -90) ℃ (-90 ~ 250) ℃ (250 ~ 550) ℃ (550 ~ 660) ℃ (660 ~ 1 100) ℃ (1 100 ~ 1 600) ℃	0.010 ℃ 0.060 ℃ 0.030 ℃ 0.017 ℃ 0.020 ℃ 0.060 ℃ 0.7 ℃ 1.7 ℃	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators (Temperature indicators/recorders/controllers) With Sensor	50102	(-196 ~ 500) ℃ (500 ~ 660) ℃ (660 ~ 700) ℃ (700 ~ 900) ℃ (900 ~ 1 100) ℃ (1 100 ~ 1 400) ℃ (1 400 ~ 1 600) ℃	0.020 ℃ 0.045 ℃ 0.59 ℃ 0.60 ℃ 0.61 ℃ 2.0 ℃ 2.1 ℃	SPRT, STANDARD TC/ SICT-CP-50102
Without Sensor		(-196 ~ 0) ℃ (0 ~ 100) ℃ (100 ~ 200) ℃ (200 ~ 300) ℃ (300 ~ 400) ℃ (400 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 700) ℃ (700 ~ 800) ℃ (800 ~ 1 300) ℃ (1 300 ~ 1 600) ℃	0.010 ℃ 0.013 ℃ 0.018 ℃ 0.022 ℃ 0.025 ℃ 0.029 ℃ 0.033 ℃ 0.040 ℃ 0.044 ℃ 0.07 ℃ 0.09 ℃	
(temperature calibrators) Output		(-196 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 800) ℃ (800 ~ 1 300) ℃ (1 300 ~ 1 600) ℃	0.005 ℃ 0.006 ℃ 0.007 ℃ 0.08 ℃ 0.10 ℃	
Input		(-196 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 800) ℃ (800 ~ 1 300) ℃ (1 300 ~ 1 600) ℃	0.03 ℃ 0.04 ℃ 0.05 ℃ 0.07 ℃ 0.09 ℃	
Glass thermometers: liquid-in-glass, Beckmann liquid-in-glass	50103	(-90 ~ -58) ℃ (-58 ~ 400) ℃ (400 ~ 500) ℃	0.15 ℃ 0.04 ℃ 0.15 ℃	SPRT/ SICT-CP-50103

501. Contact temperature

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance thermometers; SPRT, IPRT, thermistors, etc. (Fixed point) Ar TP Hg TP H ₂ O TP Ga MP In FP Sn FP Zn FP Al FP Ag FP	50104	(-196 ~ 100) °C (100 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C -189.344 2 °C -38.834 4 °C 0.01 °C 29.764 6 °C 156.598 5 °C 231.928 °C 419.527 °C 660.323 °C 961.78 °C	0.019 °C 0.020 °C 0.021 °C 0.044 °C 0.87 mK 1.1 mK 0.38 mK 1.1 mK 2.2 mK 2.2 mK 2.3 mK 2.7 mK 16 mK	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers; bimetal, gas or liquid type bimetal	50105	(-196 ~ -70) °C (-70 ~ 100) °C (100 ~ 200) °C (200 ~ 500) °C (500 ~ 650) °C	0.6 °C 0.2 °C 0.3 °C 0.6 °C 1.2 °C	SPRT/ SICT-CP-50105
Thermocouples: noble metal, base metal, pure metal, special type, etc. Base metal Noble metal (Fixed point) H ₂ O ICE Point Sn FP Zn FP Al FP Ag FP Cu FP Co-C MP Fe MP	50106	(0 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C (-196 ~ -90) °C (-90 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (660 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C 0.00 °C 231.928 °C 419.527 °C 660.323 °C 961.78 °C 1 084.62 °C 1 324 °C 1 534 °C	0.5 °C 1.7 °C 1.8 °C 0.4 °C 0.2 °C 0.3 °C 0.4 °C 1.1 °C 1.4 °C 1.8 °C 0.2 °C 0.2 °C 0.2 °C 0.2 °C 0.3 °C 0.3 °C 1.1 °C 1.6 °C	SPRT, Fixed point, STANDARD TC/ SICT-CP-50106

501. Contact temperature

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature transducers	50107	(-196 ~ 400) °C (400 ~ 500) °C (500 ~ 660) °C (660 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C	0.031 °C 0.043 °C 0.072 °C 0.6 °C 0.7 °C 2.1 °C 2.2 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107
Primary fixed-point cells and apparatus H ₂ O TP	50108	0.01 °C	0.24 mK	Triple-Point Cell SICT-CP-50108

502. non contact temperature

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical pyrometers	50203	(900 ~ 1 800) ℃	5 ℃	Standard Lamp/ SICT-CP-50203
Standard radiation thermometers	50204	(-40 ~ -20) ℃ (-20 ~ 0) ℃ (0 ~ 10) ℃ (10 ~ 50) ℃ (50 ~ 100) ℃ (100 ~ 200) ℃ (200 ~ 300) ℃ (300 ~ 400) ℃ (400 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 700) ℃ (700 ~ 800) ℃ (800 ~ 900) ℃ (900 ~ 1 200) ℃ (1 200 ~ 1 400) ℃ (1 400 ~ 1 500) ℃ (1 500 ~ 1 600) ℃ (1 600 ~ 1 800) ℃ (1 800 ~ 2 000) ℃ (2 000 ~ 2 100) ℃ (2 100 ~ 2 200) ℃ (2 200 ~ 2 300) ℃ (2 300 ~ 2 400) ℃	0.9 ℃ 0.7 ℃ 0.4 ℃ 0.3 ℃ 0.4 ℃ 0.5 ℃ 0.7 ℃ 0.8 ℃ 1.0 ℃ 1.1 ℃ 1.3 ℃ 1.6 ℃ 1.7 ℃ 1.8 ℃ 1.9 ℃ 2.0 ℃ 2.1 ℃ 2.5 ℃ 2.6 ℃ 4.2 ℃ 4.4 ℃ 4.6 ℃ 4.7 ℃	Transfer Standard Pyrometer/ SICT-CP-50204
Thermal image apparatus	50205	(-40 ~ -20) ℃ (-20 ~ 0) ℃ (0 ~ 10) ℃ (10 ~ 50) ℃ (50 ~ 100) ℃ (100 ~ 200) ℃ (200 ~ 300) ℃ (300 ~ 400) ℃ (400 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 700) ℃ (700 ~ 800) ℃ (800 ~ 900) ℃ (900 ~ 1 200) ℃ (1 200 ~ 1 400) ℃ (1 400 ~ 1 500) ℃ (1 500 ~ 1 600) ℃ (1 600 ~ 1 800) ℃ (1 800 ~ 2 000) ℃ (2 000 ~ 2 100) ℃ (2 100 ~ 2 200) ℃ (2 200 ~ 2 300) ℃ (2 300 ~ 2 400) ℃	0.9 ℃ 0.7 ℃ 0.4 ℃ 0.3 ℃ 0.4 ℃ 0.5 ℃ 0.7 ℃ 0.8 ℃ 1.0 ℃ 1.1 ℃ 1.3 ℃ 1.6 ℃ 1.7 ℃ 1.8 ℃ 1.9 ℃ 2.0 ℃ 2.1 ℃ 2.5 ℃ 2.6 ℃ 4.2 ℃ 4.4 ℃ 4.6 ℃ 4.7 ℃	Transfer Standard Pyrometer/ SICT-CP-50205 SICT-CP-50205

502. non contact temperature

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Blackbody furnaces	50206	(-40 ~ 0) ℃ (0 ~ 10) ℃ (10 ~ 50) ℃ (50 ~ 100) ℃ (100 ~ 200) ℃ (200 ~ 300) ℃ (300 ~ 400) ℃ (400 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 700) ℃ (700 ~ 1 100) ℃ (1 100 ~ 1 300) ℃ (1 300 ~ 1 500) ℃ (1 500 ~ 1 700) ℃ (1 700 ~ 1 800) ℃ (1 800 ~ 1 900) ℃ (1 900 ~ 2 000) ℃ (2 000 ~ 2 100) ℃ (2 100 ~ 2 200) ℃ (2 200 ~ 2 300) ℃ (2 300 ~ 2 400) ℃	0.6 ℃ 0.4 ℃ 0.3 ℃ 0.4 ℃ 0.5 ℃ 0.7 ℃ 0.8 ℃ 0.9 ℃ 1.0 ℃ 1.1 ℃ 1.4 ℃ 1.5 ℃ 1.6 ℃ 1.8 ℃ 1.9 ℃ 2.0 ℃ 2.1 ℃ 3.7 ℃ 3.9 ℃ 4.1 ℃ 4.3 ℃	Transfer Standard Pyrometer/ SICT-CP-50206
Others ; ear thermometers, etc.	50207	(30 ~ 45) ℃	0.07 ℃	Standard prt/ SICT-CP-50207

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dew-point hygrometers; chilled mirror, alumina thinfilm, etc.	50301	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -20) °C D.P. (-20 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.20 °C D.P. 0.19 °C D.P. 0.13 °C D.P. 0.15 °C D.P.	Dewpoint Meter/ SICT-CP-50301
Relative humidity hygrometers; polimer thinfilm, hair, etc. humidity Temperature	50302	(3 ~ 60) % R.H. (60 ~ 90) % R.H. (90 ~ 98) % R.H. (-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.6 °C 0.3 °C 0.5 °C 1.5 °C	Dewpoint Meter/ SICT-CP-50302
Psychrometers; assmann ventilated, PRT type, etc. assmann ventilated (humidity) (Temperature) PRT type (humidity) (Temperature)	50303	(10 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 95) % R.H. (0 ~ 50) °C (10 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H. (0 ~ 80) °C (80 ~ 100) °C	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 0.3 °C 1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.3 °C 0.5 °C	Dewpoint Meter/ SICT-CP-50303
Temperature humidity recorders; Hygrothermograph, etc Humidity Temperature	50304	(5 ~ 70) % R.H. (70 ~ 95) % R.H. (-20 ~ 80) °C	2.1 % R.H. 2.2 % R.H. 0.7 °C	Dewpoint Meter/ SICT-CP-50304

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transducers; dew-point /relative humidity (Dew-point Transducers) Dew point (Relative humidity Transducers) Humidity Temperature	50305	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -60) °C D.P. (-60 ~ -40) °C D.P. (-40 ~ -20) °C D.P. (-20 ~ 0) °C D.P. (0 ~ 50) °C D.P. (50 ~ 90) °C D.P. (90 ~ 95) °C D.P. (3 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H. (-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.60 °C D.P. 0.33 °C D.P. 0.22 °C D.P. 0.21 °C D.P. 0.20 °C D.P. 0.15 °C D.P. 0.14 °C D.P. 0.15 °C D.P. 0.17 °C D.P. 1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.7 °C 0.3 °C 0.5 °C 1.5 °C	Dewpoint Meter/ SICT-CP-50305
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc. Dew point Humidity Temperature	50306	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -30) °C D.P. (-30 ~ -10) °C D.P. (-10 ~ 60) °C D.P. (60 ~ 80) °C D.P. (80 ~ 95) °C D.P. (3 ~ 20) % R.H. (20 ~ 30) % R.H. (30 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 98) % R.H. (-90 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C	0.60 °C D.P. 0.32 °C D.P. 0.19 °C D.P. 0.17 °C D.P. 0.16 °C D.P. 0.13 °C D.P. 0.14 °C D.P. 0.15 °C D.P. 1.8 % R.H. 1.7 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H. 0.4 °C 0.5 °C 0.6 °C	Dewpoint Meter/ SICT-CP-50306

504. Moisture

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cereal moisture meters Moisture	50401	(9 ~ 20) % M.C.	0.7 % M.C.	Balance/ SICT-CP-50401
Wood moisture meters Moisture	50402	(8 ~ 25) % M.C.	2.5 % M.C.	Balance/ SICT-CP-50402
Paper moisture meters Moisture	50403	(8 ~ 20) % M.C.	3.4 % M.C.	Balance/ SICT-CP-50403

601. Sound in air

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Soundcalibrators	60102	(250 Hz) (94 ~ 114) dB (1 000 Hz) (94 ~ 114) dB	0.10 dB 0.10 dB	Reference microphone/ SICT-CP-60102
Microphones	60104	(250 Hz) (-60 ~ -20) dB	0.15 dB	Reference microphone/ SICT-CP-60104
Sound level meters	60106	(31.5 ~ 12 500) Hz 31.5 Hz 63 Hz 125 Hz 250 Hz 500 Hz 1 000 Hz 2 000 Hz 4 000 Hz 8 000 Hz 12 500 Hz	0.4 dB 0.3 dB 0.3 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.4 dB 0.6 dB	Multifunction Acoustic Calibrator/ SICT-CP-60107

603. Vibration

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators	60301	10 Hz (10 ~ 2 500) Hz	1.6×10^{-2} 1.5×10^{-2}	Reference Accelerometer/ SICT-CP-60301
Vibration transducers	60302	0.5 Hz (0.5 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz (2 500 ~ 5 000) Hz	1.5×10^{-2} 1.4×10^{-2} 1.1×10^{-2} 1.2×10^{-2} 2.4×10^{-2}	Reference Accelerometer/ SICT-CP-60302
Vibration measuring instruments	60303			Reference Accelerometer/ SICT-CP-60303
Acceleration		10 Hz 20 Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz	1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2}	
Velocity		10 Hz 20 Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz	1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2}	
Displacement		(10 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz	1.4×10^{-2} 2.1×10^{-2} 5.9×10^{-2}	

701. Photometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters Illuminance	70101	(0.5 ~ 20 000) lx	1.7×10^{-2}	Illuminance Meters/ SICT-CP-70101
Luminance meters Luminance	70102	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1×10^{-2} 1.7×10^{-2} 1.4×10^{-2} 1.6×10^{-2}	Luminance Standard Sources/ SICT-CP-70102
Total luminous flux meters Total luminous flux	70103	70 lm (70 ~ 4 650) lm	3.2×10^{-2} 1.5×10^{-2}	Total Luminous Flux Standard Lamps/ SICT-CP-70103
Luminance intensity meters Luminance	70104	(72 ~ 3 200) cd	3.7×10^{-2}	Luminous Intensity Standard Lamps, Illuminance Meters / SICT-CP-70104

702. Properties of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color temperature meters Color temperature Chromaticity	70202	(2 677 ~ 3 333) K x y	25 K 0.004 0.004	Color Temperature Standard Lamps/ SICT-CP-70202
Color temperature standard lamps Color temperature Chromaticity	70203	(2 677 ~ 3 333) K x y	27 K 0.005 0.005	Spectroradiometers/ SICT-CP-70203
Colorimeters; source color Luminance Chromaticity	70204	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	2.1×10^{-2} 1.7×10^{-2} 1.4×10^{-2} 1.6×10^{-2} 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.003 0.003	Luminance Standard Sources/ SICT-CP-70204
Laser power meters	70207	(405 nm) (0.75 ~ 9) mW (660 nm) (0.7 ~ 47) mW (785 nm) (0.7 ~ 46) mW (1 080 nm) (1 ~ 40) W	1.2×10^{-2} 1.2×10^{-2} 1.2×10^{-2} 3.3×10^{-2}	Optical Power Meters/ SICT-CP-70207
Standard LED light sources Total luminous flux	70208	(68.4 ~ 72.6) lm	3.8×10^{-2}	Total Spectral Radiant Flux Meters/ SICT-CP-70208
Total luminous flux standard lamps Total luminous flux	70209	(320 ~ 10 000) lm	4.7×10^{-2}	Total Luminous Flux Standard Lamps/ SICT-CP-70209
Optical detectors Relative spectral responsivity	70210	(0 ~ 1) (300) nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm	8.8×10^{-2} 8.0×10^{-2} 7.2×10^{-2} 6.5×10^{-2}	Photodiodes/ SICT-CP-70210

702. Properties of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical detectors Relative spectral responsivity	70210	(315 ~ 320) nm (320 ~ 325) nm (325 ~ 330) nm (330 ~ 375) nm (375 ~ 435) nm (435 ~ 955) nm (955 ~ 965) nm (965 ~ 970) nm (970 ~ 975) nm (975 ~ 980) nm (980 ~ 985) nm (985 ~ 990) nm (990 ~ 995) nm (995 ~ 1 035) nm (1 035 ~ 1 055) nm (1 055 ~ 1 085) nm (1 090 ~ 1 100) nm	2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2}	Photodiodes/ SICT-CP-70210
Pyranometers and pyrhemimeters Irradiance responsivity	70211	(250 ~ 2 500) nm (1 000 $\pm$ 150) W/m ²	2.9×10^{-2}	Standard pyranometers/ SICT-CP-70211
Display color analyzers; luminance, chromaticity, white balance, etc. Luminance Chromaticity	70213	1 cd/m ² (1 ~ 5) cd/m ² (5 ~ 200) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y	3.8×10^{-2} 1.8×10^{-2} 1.7×10^{-2} 0.004 4 0.006 1 0.003 6 0.003 3 0.003 5 0.004 2 0.003 5 0.003 2	Luminance Meters/ SICT-CP-70213
Luminous intensity standard lamps Luminous intensity	70214	(10 ~ 20 000) cd	4.0×10^{-2}	Spectroradiometers/ SICT-CP-70214
Spectral irradiance standard lamps Spectral irradiance Illuminance	70215	250 nm (250 ~ 255) nm (255 ~ 300) nm (300 ~ 380) nm (380 ~ 450) nm (450 ~ 555) nm (555 ~ 660) nm (660 ~ 900) nm (900 ~ 1 600) nm (1 600 ~ 2 000) nm (2 000 ~ 2 300) nm (2 300 ~ 2 400) nm (800 ~ 7 200) lx	4.5×10^{-2} 4.4×10^{-2} 4.2×10^{-2} 3.5×10^{-2} 3.1×10^{-2} 2.9×10^{-2} 2.8×10^{-2} 2.7×10^{-2} 2.6×10^{-2} 2.9×10^{-2} 3.0×10^{-2} 3.4×10^{-2} 2.5×10^{-2}	Spectral Irradiance Standard Lamps/ SICT-CP-70215

702. Properties of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Total spectral radiant flux standard lamps Total spectral radiant flux	70216	350 nm (350 ~ 365) nm (365 ~ 380) nm (380 ~ 400) nm (400 ~ 455) nm (455 ~ 850) nm	6.7×10^{-2} 6.3×10^{-2} 5.8×10^{-2} 4.2×10^{-2} 3.9×10^{-2} 3.6×10^{-2}	Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70216
Luminance standard lamps Luminance Chromaticity	70217	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	2.2×10^{-2} 1.8×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.004 0.004	Luminance Standard Sources/ SICT-CP-70217
Spectral radiance standard lamps Spectral radiance	70218	300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm (315 ~ 320) nm (320 ~ 330) nm (330 ~ 340) nm (340 ~ 425) nm (425 ~ 470) nm (470 ~ 1 050) nm (1 050 ~ 1 600) nm	2.0×10^{-1} 1.7×10^{-1} 1.2×10^{-1} 8.9×10^{-2} 7.4×10^{-2} 4.8×10^{-2} 4.1×10^{-2} 3.5×10^{-2} 3.0×10^{-2} 2.8×10^{-2} 3.0×10^{-2}	Spectral Radiance Standard Sources/ SICT-CP-70218
UV irradiance meters Irradiance (UV Meter)	70219	(254 nm) 50 μW/cm ² ~ 3 mW/cm ² (365 nm) 10 μW/cm ² ~ 230 mW/cm ² (405 nm) 10 μW/cm ² ~ 230 mW/cm ²	 3.7×10^{-2} 3.8×10^{-2} 3.8×10^{-2}	UV Meter Standard Detectors/ SICT-CP-70219

702. Properties of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Spectral irradiance meters	70220	(350 ~ 1 694) nm	0.25 nm	Spectral Irradiance Standard Lamps/ SICT-CP-70220		
Wavelength						
Spectral irradiance					250 nm	4.5×10^{-2}
					(250 ~ 255) nm	4.2×10^{-2}
					(255 ~ 300) nm	4.1×10^{-2}
					(300 ~ 350) nm	3.7×10^{-2}
					(350 ~ 380) nm	3.5×10^{-2}
					(380 ~ 450) nm	3.1×10^{-2}
					(450 ~ 550) nm	2.9×10^{-2}
					(550 ~ 900) nm	2.7×10^{-2}
					(900 ~ 1 615) nm	2.6×10^{-2}
					(1 615 ~ 2 315) nm	2.9×10^{-2}
					(2 315 ~ 2 365) nm	3.4×10^{-2}
(2 365 ~ 2 400) nm	4.0×10^{-2}					
Illuminance	(800 ~ 7 200) lx	1.9×10^{-2}				
Total spectral radiant flux meters	70221	(350 ~ 850) nm	0.25 nm	Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70221		
Wavelength						
Total spectral radiant flux					350 nm	2.0×10^{-2}
					(350 ~ 365) nm	1.8×10^{-2}
					(365 ~ 375) nm	1.7×10^{-2}
					(375 ~ 390) nm	1.6×10^{-2}
					(390 ~ 445) nm	1.5×10^{-2}
(445 ~ 850) nm	1.4×10^{-2}					
Spectral radiance meters	70222	(350 ~ 1 694) nm	0.25 nm	Spectral Radiance Standard Sources/ SICT-CP-70222		
Wavelength						
Spectral radiance					300 nm	2.0×10^{-1}
					(300 ~ 305) nm	1.7×10^{-1}
					(305 ~ 310) nm	1.2×10^{-1}
					(310 ~ 315) nm	8.8×10^{-2}
					(315 ~ 320) nm	7.2×10^{-2}
					(320 ~ 325) nm	5.5×10^{-2}
					(325 ~ 335) nm	4.6×10^{-2}
					(335 ~ 345) nm	3.7×10^{-2}
					(345 ~ 405) nm	3.5×10^{-2}
					(405 ~ 455) nm	3.0×10^{-2}
					(455 ~ 755) nm	2.6×10^{-2}
					(755 ~ 1 400) nm	2.7×10^{-2}
					(1 400 ~ 1 525) nm	3.0×10^{-2}
					(1 525 ~ 1 600) nm	2.8×10^{-2}

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source T _{max} A (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.94	
		Y	0.85	
		Z	0.30	
2. L,Gray		X	0.64	
		Y	0.59	
		Z	0.22	
3. M,Gray		X	0.30	
		Y	0.27	
		Z	0.11	
4. D,Gray		X	0.11	
		Y	0.10	
		Z	0.04	
5. Red		X	0.36	
		Y	0.22	
		Z	0.06	
6. Yellow		X	0.78	
		Y	0.68	
		Z	0.08	
7. Green		X	0.19	
		Y	0.22	
		Z	0.08	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.16	
Included Reflectance Std. Light Source T _{max} A (2°)				
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.23	
		b*	0.26	
6. Yellow		L*	0.34	
		a*	0.12	
		b*	0.37	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source T _{max} A (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0.000 5	
	y		0.000 5	
2. L,Gray	x		0.000 5	
	y		0.000 5	
3. M,Gray	x		0.000 5	
	y		0.000 5	
4. D,Gray	x		0.000 5	
	y		0.000 5	
5. Red	x		0.000 8	
	y		0.000 5	
6. Yellow	x		0.000 5	
	y		0.000 5	
7. Green	x		0.000 5	
	y		0.000 5	
8. Cyan	x		0.000 5	
	y		0.000 5	
Included Reflectance Std. Light Source T _{max} A (10 °)				
1. White	X		0.95	
	Y		0.85	
	Z		0.30	
2. L,Gray	X		0.65	
	Y		0.59	
	Z		0.21	
3. M,Gray	X		0.30	
	Y		0.27	
	Z		0.11	
4. D,Gray	X		0.11	
	Y		0.10	
	Z		0.04	
5. Red	X		0.35	
	Y		0.22	
	Z		0.06	
6. Yellow	X		0.78	
	Y		0.66	
	Z		0.08	
7. Green	X		0.20	
	Y		0.22	
	Z		0.08	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.15	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source Type A (10 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.22	
		b*	0.26	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.39	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.13	
		b*	0.14	
Include Reflectance Std. Light Source Type A (10 °)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.000 7	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 5	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source T _{max} C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.83	
		Y	0.85	
		Z	0.97	
2. L,Gray		X	0.58	
		Y	0.59	
		Z	0.69	
3. M,Gray		X	0.27	
		Y	0.28	
		Z	0.32	
4. D,Gray		X	0.10	
		Y	0.10	
		Z	0.11	
5. Red		X	0.25	
		Y	0.18	
		Z	0.15	
6. Yellow		X	0.56	
		Y	0.63	
		Z	0.22	
7. Green		X	0.18	
		Y	0.24	
		Z	0.21	
8. Cyan		X	0.20	
		Y	0.23	
		Z	0.49	
Included Reflectance Std. Light Source T _{max} C (2°)				
1. White		L*	0.36	
		a*	0.09	
		b*	0.08	
2. L,Gray		L*	0.32	
		a*	0.08	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.06	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.21	
6. Yellow		L*	0.33	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.13	
		b*	0.13	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source T _{max} C (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0.000 5	
	y		0.000 5	
2. I,Gray	x		0.000 5	
	y		0.000 5	
3. M,Gray	x		0.000 5	
	y		0.000 5	
4. D,Gray	x		0.000 5	
	y		0.000 5	
5. Red	x		0.001 3	
	y		0.000 5	
6. Yellow	x		0.000 5	
	y		0.000 7	
7. Green	x		0.000 5	
	y		0.000 5	
8. Cyan	x		0.000 5	
	y		0.000 5	
Included Reflectance Std. Light Source T _{max} C (10 °)				
1. White	X		0.82	
	Y		0.85	
	Z		0.95	
2. L,Gray	X		0.57	
	Y		0.59	
	Z		0.68	
3. M,Gray	X		0.27	
	Y		0.28	
	Z		0.32	
4. D,Gray	X		0.10	
	Y		0.10	
	Z		0.11	
5. Red	X		0.24	
	Y		0.18	
	Z		0.14	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.20	
7. Green	X		0.18	
	Y		0.24	
	Z		0.19	
8. Cyan	X		0.20	
	Y		0.24	
	Z		0.48	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source C (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.07	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.23	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.42	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.12	
8. Cyan		L*	0.24	
		a*	0.11	
		b*	0.15	
Included Reflectance Std. Light Source C (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 3	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 7	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source D65 (2°) 1. White	70301	X Y Z	0.81 0.85 0.89	Color Standard Tiles/ SICT-CP-70301
2. L,Gray		X Y Z	0.56 0.59 0.64	
3. M,Gray		X Y Z	0.26 0.28 0.30	
4. D,Gray		X Y Z	0.09 0.10 0.10	
5. Red		X Y Z	0.25 0.18 0.14	
6. Yellow		X Y Z	0.58 0.63 0.20	
7. Green		X Y Z	0.17 0.25 0.20	
8. Cyan		X Y Z	0.19 0.23 0.46	
Included Reflectance Std. Light Source D65 (2°) 1. White		L* a* b*	0.36 0.07 0.07	
2. L,Gray		L* a* b*	0.32 0.07 0.06	
3. M,Gray		L* a* b*	0.25 0.06 0.05	
4. D,Gray		L* a* b*	0.17 0.04 0.04	
5. Red		L* a* b*	0.24 0.25 0.21	
6. Yellow		L* a* b*	0.33 0.16 0.40	
7. Green		L* a* b*	0.24 0.13 0.08	
8. Cyan		L* a* b*	0.24 0.14 0.13	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source D65 (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0.000 5	
	y		0.000 5	
2. I,Gray	x		0.000 5	
	y		0.000 5	
3. M,Gray	x		0.000 5	
	y		0.000 5	
4. D,Gray	x		0.000 5	
	y		0.000 5	
5. Red	x		0.001 3	
	y		0.000 5	
6. Yellow	x		0.000 6	
	y		0.000 6	
7. Green	x		0.000 5	
	y		0.000 5	
8. Cyan	x		0.000 5	
	y		0.000 5	
Included Reflectance Std. Light Source D65 (10 °)				
1. White	X		0.80	
	Y		0.85	
	Z		0.88	
2. L,Gray	X		0.56	
	Y		0.59	
	Z		0.63	
3. M,Gray	X		0.26	
	Y		0.28	
	Z		0.29	
4. D,Gray	X		0.09	
	Y		0.10	
	Z		0.10	
5. Red	X		0.24	
	Y		0.18	
	Z		0.13	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.19	
7. Green	X		0.18	
	Y		0.25	
	Z		0.18	
8. Cyan	X		0.19	
	Y		0.24	
	Z		0.45	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source D65 (10°) 1. White	70301	L*	0.36	Color Standard Tiles/ SICT-CP-70301
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.12	
		b*	0.12	
Included Reflectance Std. Light Source D65 (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 3	
		y	0.000 5	
6. Yellow		x	0.000 6	
		y	0.000 6	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Exclude Reflectance Std. Light Source Type A (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.90	
		Y	0.81	
		Z	0.29	
2. L,Gray		X	0.60	
		Y	0.55	
		Z	0.20	
3. M,Gray		X	0.26	
		Y	0.24	
		Z	0.10	
4. D,Gray		X	0.07	
		Y	0.06	
		Z	0.03	
5. Red		X	0.31	
		Y	0.20	
		Z	0.10	
6. Yellow		X	0.73	
		Y	0.64	
		Z	0.07	
7. Green		X	0.15	
		Y	0.19	
		Z	0.07	
8. Cyan		X	0.13	
		Y	0.16	
		Z	0.14	
Exclude Reflectance Std. Light Source Type A (2 °)				
1. White		L*	0.35	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.45	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.16	
		b*	0.15	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Exclude Reflectance Std. Light Source Type A (2 °)				
1. White	x		0.000 5	
	y		0.000 5	
2. I,Gray	x		0.000 5	
	y		0.000 5	
3. M,Gray	x		0.000 5	
	y		0.000 5	
4. D,Gray	x		0.000 5	
	y		0.000 5	
5. Red	x		0.001 1	
	y		0.000 5	
6. Yellow	x		0.000 5	
	y		0.000 5	
7. Green	x		0.000 5	
	y		0.000 5	
8. Cyan	x		0.000 5	
	y		0.000 5	
Exclude Reflectance Std. Light Source Type A (10 °)				
1. White	X		0.91	
	Y		0.81	
	Z		0.28	
2. L,Gray	X		0.61	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.09	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.30	
	Y		0.20	
	Z		0.10	
6. Yellow	X		0.74	
	Y		0.62	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.06	
8. Cyan	X		0.13	
	Y		0.17	
	Z		0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type A (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.35	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.05	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.49	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.14	
		b*	0.15	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 1	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 5	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type C (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.79	
	Y		0.81	
	Z		0.92	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.65	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.28	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.21	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.56	
	Y		0.60	
	Z		0.19	
7. Green	X		0.14	
	Y		0.20	
	Z		0.16	
8. Cyan	X		0.16	
	Y		0.19	
	Z		0.45	
Exclude Reflectance Std. Light Source Type C (2 °)				
1. White	L*		0.35	
	a*		0.09	
	b*		0.07	
2. L,Gray	L*		0.31	
	a*		0.07	
	b*		0.06	
3. M,Gray	L*		0.24	
	a*		0.06	
	b*		0.05	
4. D,Gray	L*		0.14	
	a*		0.04	
	b*		0.04	
5. Red	L*		0.25	
	a*		0.40	
	b*		0.60	
6. Yellow	L*		0.32	
	a*		0.17	
	b*		0.52	
7. Green	L*		0.23	
	a*		0.14	
	b*		0.08	
8. Cyan	L*		0.23	
	a*		0.14	
	b*		0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type C (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0.000 5	
	y		0.000 5	
2. I,Gray	x		0.000 5	
	y		0.000 5	
3. M,Gray	x		0.000 5	
	y		0.000 5	
4. D,Gray	x		0.000 5	
	y		0.000 5	
5. Red	x		0.002 8	
	y		0.000 5	
6. Yellow	x		0.000 7	
	y		0.000 7	
7. Green	x		0.000 5	
	y		0.000 5	
8. Cyan	x		0.000 5	
	y		0.000 5	
Exclude Reflectance Std. Light Source Type C (10 °)				
1. White	X		0.79	
	Y		0.81	
	Z		0.90	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.63	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.27	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.55	
	Y		0.56	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.44	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type C (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.35	
		a*	0.08	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.16	
Exclude Reflectance Std. Light Source Type C (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.002 9	
		y	0.000 5	
6. Yellow		x	0.000 6	
		y	0.000 8	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type D65 (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.77	
		Y	0.81	
		Z	0.85	
2. L,Gray		X	0.52	
		Y	0.55	
		Z	0.60	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.26	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.06	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.55	
		Y	0.60	
		Z	0.18	
7. Green		X	0.14	
		Y	0.21	
		Z	0.15	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.41	
Exclude Reflectance Std. Light Source Type D65 (2 °)				
1. White		L*	0.35	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.51	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.15	
		b*	0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type D65 (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0.000 5	
	y		0.000 5	
2. L,Gray	x		0.000 5	
	y		0.000 5	
3. M,Gray	x		0.000 5	
	y		0.000 5	
4. D,Gray	x		0.000 5	
	y		0.000 5	
5. Red	x		0.002 7	
	y		0.000 5	
6. Yellow	x		0.000 6	
	y		0.000 7	
7. Green	x		0.000 5	
	y		0.000 5	
8. Cyan	x		0.000 5	
	y		0.000 5	
Exclude Reflectance Std. Light Source Type D65 (10 °)				
1. White	X		0.77	
	Y		0.81	
	Z		0.84	
2. L,Gray	X		0.52	
	Y		0.55	
	Z		0.59	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.25	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.06	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.54	
	Y		0.57	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.16	
	Y		0.20	
	Z		0.41	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type D65 (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.35	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.09	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.13	
Exclude Reflectance Std. Light Source Type D65 (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.002 8	
		y	0.000 5	
6. Yellow		x	0.000 7	
		y	0.000 7	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source T _{max} A (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.94	
		Y	0.85	
		Z	0.30	
2. L,Gray		X	0.64	
		Y	0.59	
		Z	0.22	
3. M,Gray		X	0.30	
		Y	0.27	
		Z	0.11	
4. D,Gray		X	0.11	
		Y	0.10	
		Z	0.04	
5. Red		X	0.36	
		Y	0.22	
		Z	0.06	
6. Yellow		X	0.78	
		Y	0.68	
		Z	0.08	
7. Green		X	0.19	
		Y	0.22	
		Z	0.08	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.16	
Included Reflectance Std. Light Source T _{max} A (2°)				
1. White		L*	0.37	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.23	
		b*	0.26	
6. Yellow		L*	0.35	
		a*	0.12	
		b*	0.37	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source T _{max} A (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 8	
	y		0.000 7	
2. L,Gray	x		0.000 8	
	y		0.000 7	
3. M,Gray	x		0.000 8	
	y		0.000 7	
4. D,Gray	x		0.000 8	
	y		0.000 7	
5. Red	x		0.001 0	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Included Reflectance Std. Light Source T _{max} A (10 °)				
1. White	X		0.95	
	Y		0.85	
	Z		0.30	
2. L,Gray	X		0.65	
	Y		0.59	
	Z		0.21	
3. M,Gray	X		0.30	
	Y		0.27	
	Z		0.11	
4. D,Gray	X		0.11	
	Y		0.10	
	Z		0.04	
5. Red	X		0.35	
	Y		0.22	
	Z		0.06	
6. Yellow	X		0.78	
	Y		0.66	
	Z		0.08	
7. Green	X		0.20	
	Y		0.22	
	Z		0.08	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.15	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source Type A (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.37	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.22	
		b*	0.26	
6. Yellow		L*	0.34	
		a*	0.12	
		b*	0.39	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.13	
		b*	0.14	
Include Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 8	
		y	0.000 7	
2. I,Gray		x	0.000 8	
		y	0.000 7	
3. M,Gray		x	0.000 8	
		y	0.000 7	
4. D,Gray		x	0.000 8	
		y	0.000 7	
5. Red		x	0.001 0	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 8	
8. Cyan		x	0.000 7	
		y	0.000 7	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source C (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.83	
		Y	0.85	
		Z	0.97	
2. L,Gray		X	0.58	
		Y	0.59	
		Z	0.69	
3. M,Gray		X	0.27	
		Y	0.28	
		Z	0.32	
4. D,Gray		X	0.10	
		Y	0.10	
		Z	0.11	
5. Red		X	0.25	
		Y	0.18	
		Z	0.15	
6. Yellow		X	0.57	
		Y	0.63	
		Z	0.22	
7. Green		X	0.18	
		Y	0.24	
		Z	0.21	
8. Cyan		X	0.20	
		Y	0.23	
		Z	0.49	
Included Reflectance Std. Light Source C (2°)				
1. White		L*	0.37	
		a*	0.09	
		b*	0.08	
2. L,Gray		L*	0.32	
		a*	0.08	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.06	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.21	
6. Yellow		L*	0.33	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.13	
		b*	0.13	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source T _{max} C (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I,Gray	x		0.000 7	
	y		0.000 7	
3. M,Gray	x		0.000 7	
	y		0.000 7	
4. D,Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.001 4	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 9	
7. Green	x		0.000 7	
	y		0.000 7	
8. Cyan	x		0.000 7	
	y		0.000 7	
Included Reflectance Std. Light Source T _{max} C (10 °)				
1. White	X		0.82	
	Y		0.85	
	Z		0.95	
2. L,Gray	X		0.57	
	Y		0.59	
	Z		0.68	
3. M,Gray	X		0.27	
	Y		0.28	
	Z		0.32	
4. D,Gray	X		0.10	
	Y		0.10	
	Z		0.11	
5. Red	X		0.24	
	Y		0.18	
	Z		0.14	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.20	
7. Green	X		0.18	
	Y		0.24	
	Z		0.19	
8. Cyan	X		0.20	
	Y		0.24	
	Z		0.48	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source T _{max} C (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.37	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.07	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.23	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.42	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.12	
8. Cyan		L*	0.24	
		a*	0.11	
		b*	0.15	
Included Reflectance Std. Light Source T _{max} C (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source D65 (2°) 1. White	70304	X Y Z	0.81 0.85 0.89	Color Standard Tiles/ SICT-CP-70304
2. L,Gray		X Y Z	0.56 0.59 0.64	
3. M,Gray		X Y Z	0.26 0.28 0.30	
4. D,Gray		X Y Z	0.09 0.10 0.10	
5. Red		X Y Z	0.25 0.18 0.14	
6. Yellow		X Y Z	0.58 0.63 0.20	
7. Green		X Y Z	0.17 0.25 0.20	
8. Cyan		X Y Z	0.19 0.23 0.46	
Included Reflectance Std. Light Source D65 (2°) 1. White		L* a* b*	0.37 0.07 0.07	
2. L,Gray		L* a* b*	0.32 0.07 0.06	
3. M,Gray		L* a* b*	0.25 0.06 0.05	
4. D,Gray		L* a* b*	0.17 0.04 0.04	
5. Red		L* a* b*	0.24 0.25 0.21	
6. Yellow		L* a* b*	0.33 0.16 0.40	
7. Green		L* a* b*	0.24 0.13 0.08	
8. Cyan		L* a* b*	0.24 0.14 0.13	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source D65 (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I,Gray	x		0.000 7	
	y		0.000 7	
3. M,Gray	x		0.000 7	
	y		0.000 7	
4. D,Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.001 4	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 7	
8. Cyan	x		0.000 7	
	y		0.000 7	
Included Reflectance Std. Light Source D65 (10 °)				
1. White	X		0.80	
	Y		0.85	
	Z		0.88	
2. L,Gray	X		0.56	
	Y		0.59	
	Z		0.63	
3. M,Gray	X		0.26	
	Y		0.28	
	Z		0.29	
4. D,Gray	X		0.09	
	Y		0.10	
	Z		0.10	
5. Red	X		0.24	
	Y		0.18	
	Z		0.13	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.19	
7. Green	X		0.18	
	Y		0.25	
	Z		0.18	
8. Cyan	X		0.19	
	Y		0.24	
	Z		0.45	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source D65 (10°) 1. White	70304	L*	0.37	Color Standard Tiles/ SICT-CP-70304
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.12	
		b*	0.12	
Included Reflectance Std. Light Source D65 (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A (2 °)				
1. White	X		0.90	
	Y		0.81	
	Z		0.29	
2. L,Gray	X		0.60	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.10	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.31	
	Y		0.2	
	Z		0.1	
6. Yellow	X		0.73	
	Y		0.64	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.07	
8. Cyan	X		0.13	
	Y		0.16	
	Z		0.14	
Exclude Reflectance Std. Light Source Type A (2 °)				
1. White	L*		0.36	
	a*		0.07	
	b*		0.07	
2. L,Gray	L*		0.31	
	a*		0.06	
	b*		0.06	
3. M,Gray	L*		0.24	
	a*		0.05	
	b*		0.05	
4. D,Gray	L*		0.14	
	a*		0.04	
	b*		0.04	
5. Red	L*		0.25	
	a*		0.30	
	b*		0.70	
6. Yellow	L*		0.33	
	a*		0.12	
	b*		0.45	
7. Green	L*		0.22	
	a*		0.12	
	b*		0.08	
8. Cyan	L*		0.22	
	a*		0.16	
	b*		0.15	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A (2 °)				
1. White	x		0.000 8	
	y		0.000 7	
2. I,Gray	x		0.000 8	
	y		0.000 7	
3. M,Gray	x		0.000 8	
	y		0.000 7	
4. D,Gray	x		0.000 8	
	y		0.000 7	
5. Red	x		0.001 3	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Exclude Reflectance Std. Light Source Type A (10 °)				
1. White	X		0.91	
	Y		0.81	
	Z		0.28	
2. L,Gray	X		0.61	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.09	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.30	
	Y		0.20	
	Z		0.10	
6. Yellow	X		0.74	
	Y		0.62	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.06	
8. Cyan	X		0.13	
	Y		0.17	
	Z		0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type A (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.36	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.05	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.49	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.14	
		b*	0.15	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 6	
		y	0.000 6	
2. I,Gray		x	0.000 6	
		y	0.000 6	
3. M,Gray		x	0.000 6	
		y	0.000 6	
4. D,Gray		x	0.000 6	
		y	0.000 6	
5. Red		x	0.001 1	
		y	0.000 6	
6. Yellow		x	0.000 6	
		y	0.000 6	
7. Green		x	0.000 6	
		y	0.000 6	
8. Cyan		x	0.000 6	
		y	0.000 6	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type C (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.79	
		Y	0.81	
		Z	0.92	
2. L,Gray		X	0.54	
		Y	0.55	
		Z	0.65	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.28	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.07	
5. Red		X	0.21	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.56	
		Y	0.60	
		Z	0.19	
7. Green		X	0.14	
		Y	0.20	
		Z	0.16	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.45	
Exclude Reflectance Std. Light Source Type C (2 °)				
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.52	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type C (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I,Gray	x		0.000 7	
	y		0.000 7	
3. M,Gray	x		0.000 7	
	y		0.000 7	
4. D,Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.002 8	
	y		0.000 7	
6. Yellow	x		0.000 9	
	y		0.000 9	
7. Green	x		0.000 7	
	y		0.000 7	
8. Cyan	x		0.000 7	
	y		0.000 7	
Exclude Reflectance Std. Light Source Type C (10 °)				
1. White	X		0.79	
	Y		0.81	
	Z		0.90	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.63	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.27	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.55	
	Y		0.56	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.44	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type C (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.36	
		a*	0.08	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.16	
Exclude Reflectance Std. Light Source Type C (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.002 9	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.001 0	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type D65 (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.77	
		Y	0.81	
		Z	0.85	
2. L,Gray		X	0.52	
		Y	0.55	
		Z	0.60	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.26	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.06	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.55	
		Y	0.60	
		Z	0.18	
7. Green		X	0.14	
		Y	0.21	
		Z	0.15	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.41	
Exclude Reflectance Std. Light Source Type D65 (2 °)				
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.51	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.15	
		b*	0.14	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type D65 (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I,Gray	x		0.000 7	
	y		0.000 7	
3. M,Gray	x		0.000 7	
	y		0.000 7	
4. D,Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.002 7	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 9	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Exclude Reflectance Std. Light Source Type D65 (10 °)				
1. White	X		0.77	
	Y		0.81	
	Z		0.84	
2. L,Gray	X		0.52	
	Y		0.55	
	Z		0.59	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.25	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.06	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.54	
	Y		0.57	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.16	
	Y		0.20	
	Z		0.41	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type D65 (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.09	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.13	
Exclude Reflectance Std. Light Source Type D65 (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.002 8	
		y	0.000 7	
6. Yellow		x	0.000 9	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 8	
8. Cyan		x	0.000 7	
		y	0.000 7	
Absolute Spectral Reflectance White Plate (Include, Exclude Reflectance)		(360 ~ 830) nm	0.007 6	Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70304

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Gloss meters Gloss	70306	20 ° 60 ° 85 °	8.9×10^{-3} 9.5×10^{-3} 8.0×10^{-3}	Gloss Standard/ SICT-CP-70306
Gloss standard plates Gloss	70307	20° 60° 85°	9.5×10^{-3} 9.8×10^{-3} 8.3×10^{-3}	Gloss Meter/ SICT-CP-70307
Haze meters Haze Transmittance	70308	H-1 H-5 H-10 H-20 H-30 T-30 T-50 T-70 T-90	0.30 0.26 0.4 0.6 0.8 0.50 0.50 0.50 0.50	Haze Standard Plate, Transmittance Standard Plates/ SICT-CP-70308
Lens meters Vertex diopter	70312	-25 D ~ 25 D	0.03 D	Reference Lens/ SICT-CP-70312
Optical densitometers Density	70315	1 Step ~ 10 Step 11 Step 11 Step ~ 15 Step	0.03 0.06 0.11	Density CRM/ SICT-CP-70315
Reflectance meters Reflectance	70319	380 nm ~ 780 nm	1.1×10^{-2}	Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70319
Refractometers Refracto	70321	(1.332 99 ~ 1.505 80) nD 1.51 nD 1.62 nD	0.000 04 nD 0.000 2 nD 0.000 2 nD	Reference Refract CRM/ SICT-CP-70321
Transmittance meters	70323	(0.1) (250 ~ 750) nm (0.5) (250 ~ 750) nm (0.9) (250 ~ 750) nm	6.1×10^{-3} 3.8×10^{-3} 2.2×10^{-3}	Transmittance Filter/ SICT-CP-70323

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
spectrophotometers				
Wavelength		(240 ~ 750) nm	0.4 nm	
Transmittance		(0.1 ~ 0.3)		
		250 nm	8.1×10^{-3}	
		300 nm	7.8×10^{-3}	
		350 nm	7.7×10^{-3}	
		400 nm	5.4×10^{-3}	
		450 nm	5.2×10^{-3}	
		500 nm	5.2×10^{-3}	
		550 nm	5.2×10^{-3}	
		600 nm	5.2×10^{-3}	
		650 nm	5.2×10^{-3}	
		700 nm	5.2×10^{-3}	
		750 nm	5.2×10^{-3}	
		(0.3 ~ 0.5)		
		250 nm	8.1×10^{-3}	
		300 nm	7.8×10^{-3}	
		350 nm	7.7×10^{-3}	
		400 nm	5.2×10^{-3}	
		450 nm	5.2×10^{-3}	
		500 nm	5.2×10^{-3}	
		550 nm	5.1×10^{-3}	
		600 nm	5.2×10^{-3}	
		650 nm	5.2×10^{-3}	
		700 nm	5.2×10^{-3}	
		750 nm	5.2×10^{-3}	
		(0.5 ~ 0.9)		
		250 nm	7.9×10^{-3}	
		300 nm	8.2×10^{-3}	
		350 nm	7.8×10^{-3}	
		400 nm	5.2×10^{-3}	
		450 nm	5.2×10^{-3}	
		500 nm	5.2×10^{-3}	
		550 nm	5.1×10^{-3}	
		600 nm	5.1×10^{-3}	
		650 nm	5.1×10^{-3}	
		700 nm	5.2×10^{-3}	
		750 nm	5.2×10^{-3}	
		(0.01)		
		440 nm	1.3×10^{-2}	
		465 nm	8.4×10^{-3}	
		546 nm	8.9×10^{-3}	
		590 nm	1.0×10^{-2}	
		635 nm	8.1×10^{-3}	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Transmittance		(0.03)		
		440 nm	7.6×10^{-3}	
		465 nm	5.7×10^{-3}	
		546 nm	5.8×10^{-3}	
		590 nm	6.3×10^{-3}	
		635 nm	5.9×10^{-3}	
Absorbance		(0.1 ~ 0.3)		
		250 nm	0.003 6	
		300 nm	0.003 4	
		350 nm	0.003 5	
		400 nm	0.002 3	
		450 nm	0.002 3	
		500 nm	0.002 3	
		550 nm	0.002 3	
		600 nm	0.002 3	
		650 nm	0.002 3	
		700 nm	0.002 3	
		750 nm	0.002 3	
		(0.3 ~ 0.5)		
		250 nm	0.003 6	
		300 nm	0.003 4	
		350 nm	0.003 5	
		400 nm	0.002 3	
		450 nm	0.002 3	
		500 nm	0.002 3	
		550 nm	0.002 3	
		600 nm	0.002 3	
		650 nm	0.002 3	
		700 nm	0.002 3	
		750 nm	0.002 3	
		(0.5 ~ 0.9)		
		250 nm	0.003 5	
		300 nm	0.003 5	
		350 nm	0.003 5	
		400 nm	0.002 3	
		450 nm	0.002 3	
		500 nm	0.002 3	
		550 nm	0.002 3	
		600 nm	0.002 3	
		650 nm	0.002 3	
		700 nm	0.002 3	
		750 nm	0.002 3	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325	Absorbance (0.01) 440 nm 465 nm 546 nm 590 nm 635 nm (0.03) 440 nm 465 nm 546 nm 590 nm 635 nm Reflectance (250 ~ 2 500) nm Wavenumber 544.9 cm ⁻¹ 842.1 cm ⁻¹ 906.82 cm ⁻¹ 1 028.42 cm ⁻¹ 1 069.27 cm ⁻¹ 1 154.62 cm ⁻¹ 1 583.04 cm ⁻¹ 1 601.38 cm ⁻¹ 2 850.20 cm ⁻¹ 3 001.40 cm ⁻¹ 3 026.44 cm ⁻¹ 3 060.14 cm ⁻¹ 3 082.22 cm ⁻¹	 0.003 8 0.002 7 0.002 9 0.003 1 0.002 7 0.003 1 0.002 4 0.002 6 0.002 6 0.002 6 1.4 × 10 ⁻² 2.5 cm ⁻¹ 1.3 cm ⁻¹ 0.12 cm ⁻¹ 0.28 cm ⁻¹ 0.78 cm ⁻¹ 0.11 cm ⁻¹ 0.11 cm ⁻¹ 0.12 cm ⁻¹ 0.13 cm ⁻¹ 0.11 cm ⁻¹ 0.11 cm ⁻¹ 0.11 cm ⁻¹ 0.11 cm ⁻¹	Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326	Wavelength (240 ~750) nm Transmittance (0.1 ~ 0.3) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	 0.5 nm 8.5 × 10 ⁻³ 8.1 × 10 ⁻³ 8.1 × 10 ⁻³ 5.9 × 10 ⁻³ 5.7 × 10 ⁻³ 5.7 × 10 ⁻³ 5.7 × 10 ⁻³ 5.7 × 10 ⁻³ 5.7 × 10 ⁻³ 5.7 × 10 ⁻³ 5.7 × 10 ⁻³	Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Transmittance		(0.3 ~ 0.5)		
		250 nm	8.3×10^{-3}	
		300 nm	8.1×10^{-3}	
		350 nm	8.0×10^{-3}	
		400 nm	5.7×10^{-3}	
		450 nm	5.7×10^{-3}	
		500 nm	5.7×10^{-3}	
		550 nm	5.7×10^{-3}	
		600 nm	5.7×10^{-3}	
		650 nm	5.7×10^{-3}	
		700 nm	5.7×10^{-3}	
		750 nm	5.7×10^{-3}	
		(0.5 ~ 0.9)		
		250 nm	8.2×10^{-3}	
		300 nm	8.1×10^{-3}	
		350 nm	8.0×10^{-3}	
		400 nm	5.7×10^{-3}	
		450 nm	5.7×10^{-3}	
		500 nm	5.7×10^{-3}	
		550 nm	5.7×10^{-3}	
		600 nm	5.6×10^{-3}	
		650 nm	5.6×10^{-3}	
		700 nm	5.7×10^{-3}	
		750 nm	5.9×10^{-3}	
Absorbance		(0.1 ~ 0.3)		
		250 nm	0.003 7	
		300 nm	0.003 7	
		350 nm	0.003 8	
		400 nm	0.002 7	
		450 nm	0.002 7	
		500 nm	0.002 8	
		550 nm	0.002 8	
		600 nm	0.002 7	
		650 nm	0.002 7	
		700 nm	0.002 7	
		750 nm	0.002 8	
		(0.3 ~ 0.5)		
		250 nm	0.003 6	
		300 nm	0.003 6	
		350 nm	0.003 6	
		400 nm	0.002 4	
		450 nm	0.002 4	
		500 nm	0.002 4	
		550 nm	0.002 4	
		600 nm	0.002 4	
		650 nm	0.002 4	
		700 nm	0.002 4	
		750 nm	0.002 5	

703. Properties of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Absorbance		(0.5 ~ 0.9)		
		250 nm	0.003 6	
		300 nm	0.003 5	
		350 nm	0.003 5	
		400 nm	0.002 4	
		450 nm	0.002 4	
		500 nm	0.002 4	
		550 nm	0.002 4	
		600 nm	0.002 4	
		650 nm	0.002 4	
		700 nm	0.002 4	
		750 nm	0.002 4	
Reflectance		(360 ~ 830) nm	1.0×10^{-2}	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength output Optical power output	70402	1 310 nm, 1 550 nm 1 310 nm, 1 550 nm (0 ~ -60) dBm	0.058 nm 0.070 dB	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70402
Optical attenuators Optical Attenuation	70410	1 310 nm, 1 550 nm (-60 ~ 0) dB	0.08 dB	Optical powermeter, Optical power stabilized lasers and LDs/ SICT-CP-70410
Fiber-optic power meters Absolute optical power Optical Linearity	70412	1 310 nm, 1 550 nm (0 ~ -60) dBm 1 310 nm, 1 550 nm (0 ~ -60) dB	0.072 dB 0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70412
Optical loss testers Optical Attenuation	70413	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical attenuator/ SICT-CP-70413
Optical multimeters Absolute optical power measure Linearity measure	70415	1 310 nm, 1 550 nm (0 ~ -60) dBm 1 310 nm, 1 550 nm (0 ~ -60) dB	0.072 dB 0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70415
Optical network analyzer (Optical multimeter) Absolute optical power (광)Optical spectrum analyzer Wavelength measure Resolution measure Absolute optical power measure (Optical attenuator) Optical Attenuation Return loss (Optical time domain reflectometer) Wavelength output Optical Length measure	70416	1 310 nm, 1 550 nm (-60 ~ 0) dBm 1 310 nm 1 550 nm Resolution: (0.1 ~ 1) nm 1 310 nm 1 550 nm 1 310 nm, 1 550 nm (-60 ~ 0) dBm 1 310 nm, 1 550 nm (-60 ~ 0) dB 1 310 nm, 1 550 nm (20 ~ 40) dB 1 310 nm, 1 550 nm 1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber	0.072 dB 0.058 nm 0.058 nm 0.058 nm 0.058 nm 0.072 dB 0.07 dB 0.8 dB 0.082 nm 0.081 m 0.34 m 0.080 m 0.34 m	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical network analyzer Optical loss measure	70416	1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	 0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Optical spectrum analyzers Wavelength measure Resolution measure Absolute optical power measure Linearity measure	70417	1 310 nm 1 550 nm 분해능: (0.1 ~ 1) nm 1 310 nm 1 550 nm 1 310 nm, 1 550 nm (-60 ~ 0) dBm 1 310 nm, 1 550 nm (-60 ~ 0) dB	 0.058 nm 0.058 nm 0.058 nm 0.058 nm 0.072 dB 0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator, Optical spectrum analyzer/ SICT-CP-70417
Optical time domain reflectometers, OTDR Wavelength output Optical Length measure Optical loss measure	70418	1 310 nm, 1 550 nm 1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber 1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	 0.08 nm 0.081 m 0.34 m 0.080 m 0.34 m 0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical length fiber reference, Optical fiberloss reference, Optical spectrum analyzer/ SICT-CP-70418
Return loss meters Return loss measure	70423	1 310 nm, 1 550 nm 20 dB ~ 40 dB	 0.8 dB	Optical Returnloss generator SICT-CP-70423
Frequency stabilized lasers and LDs Wavelength optical power	70429	1 310 nm 1 550 nm 1 310 nm, 1 550 nm (-60 ~ 0) dBm	4 pm 4 pm 0.07 dB	Wavelength meter, Optical powermeter/ SICT-CP-70429

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
ASE light sources	70430	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70430
Wavelength output		1 310 nm, 1 550 nm	0.07 dB	
Optical power output		(-60 ~ 0) dBm		
Optical power stabilized lasers and LDs	70433	1 310 nm	4 pm	Wavelength meter, Optical powermeter/ SICT-CP-70433
Wavelength output		1 550 nm	4 pm	
Optical power output		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath alcohol analyzers	90101			Standard gas/ SICT-CP-90101
Dry process		(0.000 ~ 0.080) %BAC (0.080 ~ 0.190) %BAC	3.3×10^{-2} 2.1×10^{-2}	
Wet process		(0.000 ~ 0.080) %BAC (0.080 ~ 0.150) %BAC (0.150 ~ 0.400) %BAC	2.9×10^{-2} 1.6×10^{-2} 1.3×10^{-2}	
Environmental air monitoring instruments	90102			Standard gas/ SICT-CP-90102
Oxygen		(0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	
Carbon monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Carbon dioxide		(0 ~ 0.50) cmol/mol (0.50 ~ 5.00) cmol/mol (5.00 ~ 19.00) cmol/mol	2.0×10^{-2} 1.5×10^{-2} 2.1×10^{-2}	
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}	
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}	
Ozone		0.0 nmol/mol (0.0 ~ 1 000.0) nmol/mol	2.2 nmol/mol 2.5×10^{-2}	
Gas analyzers	90103			Standard gas/ SICT-CP-90103
Oxygen		(0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	
Carbon monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Carbon dioxide		(0 ~ 0.50) cmol/mol (0.50 ~ 5.00) cmol/mol (5.00 ~ 19.00) cmol/mol	2.0×10^{-2} 1.5×10^{-2} 2.1×10^{-2}	
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}	
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}	
Ozone		0.0 nmol/mol (0.0 ~ 1 000.0) nmol/mol	2.2 nmol/mol 2.5×10^{-2}	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Exhaust gas test instruments	90104			Standard gas/ SICT-CP-90103
Oxygen		(0 ~ 1.5) $\mu\text{mol/mol}$ (1.5 ~ 20) cmol/mol	2.0×10^{-2} 1.1×10^{-2}	
Carbon monoxide		(0 ~ 5.0) cmol/mol	2.1×10^{-2}	
Carbon dioxide		(0 ~ 19) cmol/mol	2.0×10^{-2}	
Nitrogen monoxide		(0 ~ 2 000) $\mu\text{mol/mol}$	2.0×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Others: pH meter, Electrical conductivity meter	90104			CRM/ SICT-CP-90199
pH meter		(4 ~ 10) pH	0.013 pH	
Electrical conductivity meter		100 $\mu\text{S/cm}$ 1 413 $\mu\text{S/cm}$ 12.85 mS/cm 111.3 mS/cm	3.1 $\mu\text{S/cm}$ 9.7 $\mu\text{S/cm}$ 0.073 mS/cm 0.78 mS/cm	